

From: Penny Lanzisera
To: shabasol@pahosp.com
Date: Mon, Sep 10, 2007 9:57 AM
Subject: Additional Information to Support Clinical Use, Mail Control No. 140850

License No. 37-06864-06
Docket No. 030-15163
Mail Control No. 140850

Dr. Shabason,

Please respond to this email in writing indicating: (i) the duration of time over which timer accuracy is measured; and confirming that (i) the door interlock causes the source to retract when the door is pushed open; and (ii) the lights on the treatment console are confirmed as operational during the daily checks. Additionally, please indicate whether Mr. Etemadnia and Dr. Lustig should be approved for HDR use, and if so, submit confirmation of their vendor training.

You may reply to my attention via fax to 610-337-5269. Please refer to mail control number 140850.

Penny Lanzisera
Senior Health Physicist
US NRC, Region 1

Mail Envelope Properties (46E55B76.E8C : 17 : 55940)

Subject: Additional Information to Support Clinical Use, Mail Control No.
140850

Creation Date Mon, Sep 10, 2007 9:57 AM

From: Penny Lanzisera

Created By: PAN@nrc.gov

Recipients

pahosp.com

shabasol (shabasol@pahosp.com)

Post Office

Route

pahosp.com

Files

MESSAGE

Size

1312

Date & Time

Monday, September 10, 2007 9:57 AM

Options

Expiration Date:

None

Priority:

Standard

ReplyRequested:

No

Return Notification:

None

Concealed Subject:

No

Security:

Standard