

U.S. NUCLEAR REGULATORY COMMISSION

Date: 9/14/07

TELEPHONE CONVERSATION RECORD

Time:

Mail Control 140330 License No(s). 45-00034-30 Docket No(s). 03003067
or Report No(s).

Name of Licensee: University of Virginia (gamma knife license)

Name of Participant(s): Alan Aqualino, Ph.D., RSO and AMP

Telephone No. 484-982-0781

Subject: Followup to letter dated September 13, 2007

(NOTE: This will be used as the Documents Title in ADAMS)

Summary:

a) Dr. Aqualino said he would like to retain the Type C authorization for now, removing it after the additional AUs and AMPs receive Perfexion training and are authorized for Perfexion use. [Type C sources were re-loaded into the Perfexion, so don't need documentation of return of sources to vendor.]

b) I asked Dr. Aqualino if he intended to tie down to the license the final portion of the letter dated September 13, 2007, titled "Official Response" by Elekta's Per Kjall given as documentation for the Sector-Positioning QA.' Dr. Aqualino said this was provided for information only, and he would not like this to be considered a license commitment, as it is not consistent with information previously provided by the licensee.

Action Required: Issue amendment, retaining the Type C authorization and eliminating as a tie-down pages 23-38 from the letter dated September 13, 2007.

Document Availability: Publicly Available Non-Publicly Available

Non-Sensitive Non-Sensitive Copyright Sensitive Sensitive Copyright

Immediate Release Normal Release Delay Release Date

Prepared & SUNSI Review Completed By: S. Gabriel / RA / Date: 9/17/07