


NRC FORM 591M PART 1 (10-2003) 19 CFR 2.201		U.S. NUCLEAR REGULATORY COMMISSION	
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION			
1. LICENSEE/LOCATION INSPECTED: Community Medical Center 89 Highway 37 West Toms River, New Jersey 08755		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region I, 475 Allendale Road King of Prussia, Pennsylvania 19406-1415	
REPORT Nos 2007001			
3. DOCKET NUMBER(S) 030-12158	4. LICENSE NUMBER(S) 29-09806-03	5. DATE(S) OF INSPECTION September 11, 2007	
LICENSEE:			
<p>The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records. Interviews with personnel, and observations by the inspector. The inspection findings are as follows:</p>			
<input type="checkbox"/> 1. Based on the inspection findings, no violations were identified.			
<input type="checkbox"/> 2. Previous violation(s) closed.			
<input type="checkbox"/> 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.			
<input type="checkbox"/> Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):			
<input checked="" type="checkbox"/> 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.			
<p>10 CFR 35.633(b) requires, in part, that full calibration measurements must include determination of timer accuracy and linearity over the typical range of use.</p> <p>Contrary to the above, the licensee did not determine timer linearity over the typical range of use. Specifically, as of September 11, 2007, the licensee's typical treatment range of use for Mammosite is 10-12 minutes, and timer linearity checks were only performed over 300 seconds.</p>			
Licensee's Statement of Corrective Actions for Item 4, above.			
<p>I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.</p>			
Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	FRANK L. GELORMINI		9/20/07
NRC INSPECTOR	Shirley Xu		9/17/2007

NRC FORM 591M PART 1 (Rev. by RI 07/06)

G:\Doc\Curry\Insp\Records\29-09806-03.07-001.591P1.wpd

SUNSI Review Completed By: IRA / S Xu

Public

Non-Sensitive