R		<b>SOUTHWEST RESEARCH INSTITUTE</b> 6220 Culebra Road, P.O. Drawer 28510 Institute Quality Systems Institute Calibration Laboratory Phone: 210-522-5215 Fax 210-522-4834				
	Certificate of Calibration					
Submittee	<b>i By:</b> DIV20	<u> </u>	Work Order:	303075861		
Add	Address: B57 Date Issue			I: Aug 15, 2007		
Contact: DON BANNON			Calibration Date: Aug 15, 2007			
Manufacturer / Model: VAISALA / HMP235			*Calibration Due: Aug 15, 2008			
<b>Description:</b> HUMIDITY/TEMPERATURE TRANSMITTER			Calibration Location: Bldg. 64			
Seria	l No: W1840037		Environment: 7	Temp. 73.0°F Hum. 40 %RH		
Asse	t No: 008768		**Data Type: FOUND-LEFT			
Procee	lure: HUMIDITY-DEW POI	NT - 19 MAY 06	DivID/Location:	N/A		
quality system conforms be reproduced, except ir product endorsement by of this calibration relate	to ISO/IEC 17025, 2005, ANSI/ full, without the written approva Southwest Research Institute, An only to the instrument described a comer, does not imply the instrum	NCSL Z540-1-1994 and relevant for Laboratory above at the time of calibration and ent will remain within tolerance as	equirements of the ISO 9000-2000 sta te Calibration Laboratory. This certif Accreditation (A2LA) or any agency I does not imply any long term stabilit any number of factors may cause an of the approximation of the short	ndard. This certificate shall not icate shall not be used to claim of the U. S. Government. Results y of the instrument.		
determination of in-/out	-of-tolerance or compliance/nonco	ompliance. See Remarks or attach	ed Measurement Report with the same	Work Order number for data.		
Reported uncertainty ca uncertainty with a cover	lculated in accordance with the IS age factor of k=2 to approximate	O "Guide to the Expression of Uno a 95% confidence level.	certainty in Measurement" (GUM) and	l represents an expanded		
<b>Remarks: -20</b> to 180°C	(4-20 mA) 0 to 100 %RH (4-20 m	mA)				
Standards Used						
Asset No. Serial No.	Manufacturer	Model	Description	Cal Due		
009917 604420	HART SCIENTIFIC	5612	RTD	Sep 28, 07		

-- -- -- --

009917	604420	HART SCIENTIFIC	5612	RTD	Sep 28, 07
010375	8426006	FLUKE	45	MULTIMETER	Jul 17, 08
012098	9096026	FLUKE	45	MULTIMETER	Mar 02, 08
006404	9806123	THUNDER SCIENTIFIC	2500	HUMIDITY GENERATOR	May 30, 08

srk () jrg () blt () swc Reviewed by: () wgh ( Metrology Technician

m:\a2la1.rpt Rev date August 15, 2005

L Measurements by: Bob Trollinger

Metrology Technician

Page 1 of 1

## Southwest Research Institute Calibration Laboratory Measurement Report

Work Order:	303075861	Mfr.	VAISALA	·····	Technician	:	blt	
Asset No.	008768	Model	HMP235 ·	20 to 180 C				
Serial No.	W1840037	Туре.	Temp/RH	Transmitter	Cal Date:		15-Aug-07	
Remarks:	Temperature accuracy is based on worst case error.							
	Manufacturer %RH Limits are +/-1% 0 to 90%RH.							
	CALIBRATED 25 to 65 °C							
Function/Range	Test Point	TI Re	ading	Difference	+/-Limit	+/-Uncertainty	Found/Left	
% RH	% RH	mAmp	% RH	% RH	% RH	% RH	Result	
0-100	20.00	7.074	19.21	-0.79	1.0	0.58	Pass	
	40.48	10.466	40.41	-0.07	1.0	0.58	Pass	
	60.06	13.552	59.70	-0.36	1.0	0.58	Pass	
	70.08	15.1	69.38	-0.70	1.0	0.58	Pass	
Degrees	°C	mAmp	°C	°C	°C	°C		
-	25.67	7.640	25.50	-0.17	0.60	0.021	Pass	
	44.24	9.141	44.26	0.03	0.60	0.038	Pass	
	63.38	10.681	63.51	0.14	0.60	0.055	Pass	
		EN	D OF REPO	RT				

**\*** 1

\*\*