

NRC FORM 314

U.S. NUCLEAR REGULATORY COMMISSION

APPROVED BY OMB: NO. 3150-0028

EXPIRES: 06/30/2007

(6-2004)
10 CFR 30.38(j)(1); 40.42(j)(1);
70.38(j)(1); and 72.54(j)(1)

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Records and FOIA/Privacy Services Branch (T-5 F52), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to infocollects@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

CERTIFICATE OF DISPOSITION OF MATERIALS

LICENSEE NAME AND ADDRESS

Roy N. Ford COMPANY INC.
P.O. BOX 9
BLAIRS, VA 24527

LICENSE NUMBER

45-24927-01

DOCKET NUMBER

030-30076

LICENSE EXPIRATION DATE

November 30, 2012

This license has expired.

A. LICENSE STATUS (Check the appropriate box)
This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.

a. Transfer of radioactive materials to the licensee listed below:

ADAM CONSTRUCTION CO. License 45-25423-01

b. Disposal of radioactive materials:

PURCHASE DATE MARCH 30, 2007

1. Directly by the licensee:

2. By licensed disposal site:

3. By waste contractor:

c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

RECEIVED
REGION 1
2007 OCT 10 AM 11:03

C. SURVEYS PERFORMED AND REPORTED

- 1. A radiation survey was conducted by the licensee. The survey confirms:
 - a. the absence of licensed radioactive materials
 - b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
- 2. A copy of the radiation survey results:
 - a. is attached; or b. is not attached (Provide explanation); or c. was forwarded to NRC on: _____ Date
- 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
 - a. The results of the latest leak test are attached; and/or
 - b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME Fred C. Ford	TITLE Vice President	TELEPHONE (Include Area Code) 434-836-4561 434 728-8979	E-MAIL ADDRESS
Mail all future correspondence regarding this license to: P.O. BOX 9, BLAIRS, VA 24527			

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE

Fred C. Ford V.P.

SIGNATURE

Fred C. Ford

DATE

9-18-07

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.


Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
 Research Triangle Park, NC 27709
 Tel: (877) 876-9537 Fax: (919) 485-2250

License: NC 032-0182-1

LEAK TEST CERTIFICATE
DEVICE:

Model: 4640 Serial No: 2595

SEALED SOURCES:

SERIAL NO.	MEASURE DATE	NUCLIDE	ACTIVITY	
			(GBq)	(mCi)
77-2958	9/15/2004	CS-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 04/02/2007

Sample analyzed on: 04/03/2007 at 9:57:00 AM

Analyzed by: HARRY P. PENDLETON, JR.

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.29E+01	2.02E+01
Background measurement (cpm)	1	25
Sample measurement (cpm)	0	22
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	5.1E-01	1.3E+00

This certifies that the above leak test results are:

Less than 185 Bq (0.005 uCi)

Greater than 185 Bq (0.005 uCi)

If greater than 185 Bq (0.005 uCi):


Person Notified _____ Date _____

Phone _____ and/or Fax _____

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p>Licensee</p> <p>1. Adams Construction Co.</p> <p>2. 523 Rutherford Avenue, NE Roanoke, Virginia 24016</p>	<p>In accordance with the letter dated January 18, 2007,</p> <p>3. License number 45-25423-01 is amended in its entirety to read as follows:</p> <p>4. Expiration date February 28, 2008</p> <p>5. Docket No. 03034638 Reference No.</p>
--	--



<p>6. Byproduct, source, and/or special nuclear material</p> <p>A. Cesium 137</p> <p>B. Americium 241</p>	<p>Chemical and/or physical form</p> <p>A. Sealed Sources (Troxler Dwg. A-102112)</p> <p>B. Sealed Sources (Troxler Dwg. A-102451)</p>	<p>8. Maximum amount that licensee may possess at any one time under this license</p> <p>A. No single source to exceed the maximum activity specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission or an Agreement State</p> <p>B. No single source to exceed the maximum activity specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission or an Agreement State</p>
---	--	---

9. Authorized use:

A. and B. In Troxler Electronic Laboratories, Inc., Model Nos. 3400 Series and Model 4640 portable gauging devices for measuring physical properties of materials.

CONDITIONS

10. Licensed material may be used or stored at the licensee's facilities located at 213 Gypsum Road, Danville, Virginia and may be used at temporary job sites of the licensee anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material, including areas of exclusive Federal jurisdiction within Agreement States.

Don Johnson Adams Co.
Remove Gadge From Roy A. Ford site 4-27-07 *Act*



141163

Roy M. Ford Co
P.O. Box 9
Blairs VA 24527

CERTIFIED MAIL™



7007 0220 0004 3438 5050




SEP 28 2007
SEP 28 2007

Region II, Division of Nuclear Material Safety
61 Forstn Street Suite 23T85
Atlanta, Georgia 30303-8931

CERTIFICATE OF Dispositio of materia
TERMINATE Licenses

3030389305 0034



This is to acknowledge the receipt of your letter/application dated

9/18/2007 ^(RECEIVED) (10/10/2007) to inform you that the initial processing which includes an administrative review has been performed.

TEAM. 45-24927-01
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 141163.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.