

RECEIVED
REGION 1

NM562

2007 OCT 15 PM 1:35

October 9, 2007

U.S.N.R.C. Regional 1
Nuclear Materials Branch
475 Allendale Road
King of Prussia, PA 19406-1415

LL 31282
03037563
03620

Office of the President
6001 University Boulevard
Moon Township, PA
15108-1189

412-262-8322 phone
412-397-5958 fax
dellomo@rmu.edu
www.rmu.edu

(37-31282-01)


To Whom It May Concern:

RE: Robert Morris University
New License Application

Enclosed please find an application for the initial license for the above referenced facility. We have followed NUREG – 1556, Vol. 9 in the preparation of this renewal.

If you have any questions, please feel free to contact Dr. Angela Macci Bires, the Director of Nuclear Medicine Technology at Robert Morris University. Dr. Bires can be reached at 412-397-5410 or bires@rmu.edu.

Sincerely,



Gregory G. Dell'Omo, Ph.D.
President
Robert Morris University

Enclosure

141130

NMSS/RGN1 MATERIALS-002

NRC FORM 313
(10-2005)
10 CFR 30, 32, 33,
34, 35, 36, 39, and 40

U.S. NUCLEAR REGULATORY COMMISSION

APPROVED BY OMB: NO. 3150-0120

EXPIRES: 10/31/2008

Estimated burden per response to comply with this mandatory collection request: 4.4 hours. Submittal of the application is necessary to determine that the applicant is qualified and that adequate procedures exist to protect the public health and safety. Send comments regarding burden estimate to the Records and FOIA/Privacy Services Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to infocollects@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0120), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

APPLICATION FOR MATERIAL LICENSE

INSTRUCTIONS: SEE THE APPROPRIATE LICENSE APPLICATION GUIDE FOR DETAILED INSTRUCTIONS FOR COMPLETING APPLICATION. SEND TWO COPIES OF THE ENTIRE COMPLETED APPLICATION TO THE NRC OFFICE SPECIFIED BELOW.

APPLICATION FOR DISTRIBUTION OF EXEMPT PRODUCTS FILE APPLICATIONS WITH:

DIVISION OF INDUSTRIAL AND MEDICAL NUCLEAR SAFETY
OFFICE OF NUCLEAR MATERIALS SAFETY AND SAFEGUARDS
U.S. NUCLEAR REGULATORY COMMISSION
WASHINGTON, DC 20555-0001

IF YOU ARE LOCATED IN:

ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI, OHIO, OR WISCONSIN, SEND APPLICATIONS TO:

MATERIALS LICENSING BRANCH
U.S. NUCLEAR REGULATORY COMMISSION, REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, IL 60532-4352

ALL OTHER PERSONS FILE APPLICATIONS AS FOLLOWS:

IF YOU ARE LOCATED IN:

ALABAMA, CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, FLORIDA, GEORGIA, KENTUCKY, MAINE, MARYLAND, MASSACHUSETTS, MISSISSIPPI, NEW HAMPSHIRE, NEW JERSEY, NEW YORK, NORTH CAROLINA, PENNSYLVANIA, PUERTO RICO, RHODE ISLAND, SOUTH CAROLINA, TENNESSEE, VERMONT, VIRGINIA, VIRGIN ISLANDS, OR WEST VIRGINIA, SEND APPLICATIONS TO:

ALASKA, ARIZONA, ARKANSAS, CALIFORNIA, COLORADO, HAWAII, IDAHO, KANSAS, LOUISIANA, MONTANA, NEBRASKA, NEVADA, NEW MEXICO, NORTH DAKOTA, OKLAHOMA, OREGON, PACIFIC TRUST TERRITORIES, SOUTH DAKOTA, TEXAS, UTAH, WASHINGTON, OR WYOMING, SEND APPLICATIONS TO:

LICENSING ASSISTANCE TEAM
DIVISION OF NUCLEAR MATERIALS SAFETY
U.S. NUCLEAR REGULATORY COMMISSION, REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PA 19406-1415

NUCLEAR MATERIALS LICENSING BRANCH
U.S. NUCLEAR REGULATORY COMMISSION, REGION IV
611 RYAN PLAZA DRIVE, SUITE 400
ARLINGTON, TX 76011-4005

LL 31282
030 37563
03620

(37-31282-01)

PERSONS LOCATED IN AGREEMENT STATES SEND APPLICATIONS TO THE U.S. NUCLEAR REGULATORY COMMISSION ONLY IF THEY WISH TO POSSESS AND USE LICENSED MATERIAL IN STATES SUBJECT TO U.S. NUCLEAR REGULATORY COMMISSION JURISDICTIONS.

1. THIS IS AN APPLICATION FOR (Check appropriate item)

- A. NEW LICENSE
- B. AMENDMENT TO LICENSE NUMBER _____
- C. RENEWAL OF LICENSE NUMBER _____

2. NAME AND MAILING ADDRESS OF APPLICANT (Include ZIP code)

Robert Morris University
6001 University Boulevard
Moon Township, PA 15108-1189

3. ADDRESS WHERE LICENSED MATERIAL WILL BE USED OR POSSESSED

6001 University Boulevard
Moon Township, PA 15108-1189

4. NAME OF PERSON TO BE CONTACTED ABOUT THIS APPLICATION

Vincent A. Gargaro, DABR, MS

TELEPHONE NUMBER
440-327-9432

SUBMIT ITEMS 5 THROUGH 11 ON 8-1/2 X 11" PAPER. THE TYPE AND SCOPE OF INFORMATION TO BE PROVIDED IS DESCRIBED IN THE LICENSE APPLICATION GUIDE.

5. RADIOACTIVE MATERIAL
a. Element and mass number, b. chemical and/or physical form; and c. maximum amount which will be possessed at any one time.

6. PURPOSE(S) FOR WHICH LICENSED MATERIAL WILL BE USED.

7. INDIVIDUAL(S) RESPONSIBLE FOR RADIATION SAFETY PROGRAM AND THEIR TRAINING EXPERIENCE

8. TRAINING FOR INDIVIDUALS WORKING IN OR FREQUENTING RESTRICTED AREAS.

9. FACILITIES AND EQUIPMENT. See attached

10. RADIATION SAFETY PROGRAM.

11. WASTE MANAGEMENT.

12. LICENSE FEES (See 10 CFR 170 and Section 170.31)

FEE CATEGORY 3 - P AMOUNT ENCLOSED \$ NA

13. CERTIFICATION. (Must be completed by applicant) THE APPLICANT UNDERSTANDS THAT ALL STATEMENTS AND REPRESENTATIONS MADE IN THIS APPLICATION ARE BINDING UPON THE APPLICANT.

THE APPLICANT AND ANY OFFICIAL EXECUTING THIS CERTIFICATION ON BEHALF OF THE APPLICANT, NAMED IN ITEM 2, CERTIFY THAT THIS APPLICATION IS PREPARED IN CONFORMITY WITH TITLE 10, CODE OF FEDERAL REGULATIONS, PARTS 30, 32, 33, 34, 35, 36, 39, AND 40, AND THAT ALL INFORMATION CONTAINED HEREIN IS TRUE AND CORRECT TO THE BEST OF THEIR KNOWLEDGE AND BELIEF.

WARNING: 18 U.S.C. SECTION 1001 ACT OF JUNE 25, 1948 62 STAT. 749 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

CERTIFYING OFFICER - TYPED/PRINTED NAME AND TITLE

Gregory G. Dell'omo

SIGNATURE

DATE

10-10-07

FOR NRC USE ONLY

TYPE OF FEE	FEE LOG	FEE CATEGORY	AMOUNT RECEIVED	CHECK NUMBER	COMMENTS
		S			
APPROVED BY				DATE	

**Robert Morris University
6001 University Boulevard
Moon Township, PA 15108-1189**

NRC LICENSE APPLICATION

NRC FORM 313

Item 5 Radioactive Material

<u>Byproduct Material</u>	<u>Amount</u>	<u>Purpose</u>
A. Technetium 99m	B. 50 mCi	C. Equipment calibration
B. Cesium 137	B. 700 uCi	C. Equipment calibration

Item 6 Purpose for Which Licensed Material Will Be Used

A through C. Teaching and training of students

Item 7 Individual Responsible for Radiation Safety Program and Their Training and Experience.

The RSO for this facility is Angela R. Macci Bires, Ed.D., MPM, CNMT, RT(N). For training and experience refer to prior NRC License #47-14258-01 for Wheeling Jesuit University

Authorized Users:

Angela R. Macci Bires, Ed.D.
Vincent A. Gargaro, DABR

For training and experience refer to prior NRC License #47-14258-01 for Wheeling Jesuit University

Item 8 **Safety Training for Individuals Working in or Frequenting Restricted Areas.**

Safety instruction for individuals working in or frequenting restricted areas is an ongoing program in accordance with Appendix J of NUREG-1556, Vol.9, Rev. 2

Occupational Dose

Either we will perform a prospective evaluation demonstrating that unmonitored individuals are not likely to receive, in one year, a radiation dose in excess of 10% of the allowable limits in 10 CFR Part 20 or we will provide dosimetry that meets the requirements listed under "Criteria" in NUREG – 1556, Vol.9, Rev. 2, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licensees," dated October, 2002. R. S. Landauer, Inc. is the vendor utilized for film badge exchanges on a monthly basis.

**Robert Morris University
6001 University Boulevard
Moon Township, PA 15108-1189**

Item 9 Facility and Equipment

- Refer to the attached room diagram.

Equipment

<u>Type</u>	<u>Manufacturer / Model</u>
Well System	Ludlum 2200 or equivalent
Survey Meter	Bicron Surveyor 2000 or equivalent
Survey Meter	Eberline E-520

Radiation Monitoring Instruments

Radiation monitoring instruments will be calibrated by a person qualified to perform survey meter calibrations. We reserve the right to upgrade our survey instruments as necessary as long as they are adequate to measure the type and level of radiation for which they are used. Equipment used to measure dosages will be calibrated in accordance with nationally recognized standards or the manufacturer's instructions.

Survey Meters will be calibrated in accordance with Ohio License No. 03223480006 for Universal Consultants, Inc., N. Ridgeville and Appendix K of NUREG-1556, Vol.9, Rev. 2. Enclosed you will find a copy of their license for calibration of meters and wipe testing analysis.

Item 10 Radiation Protection Program

- **Leak Testing**

Leak testing of sealed sources will be done in accordance with Appendix Q of NUREG-1556 Vol. 9, Rev. 2 and with Ohio License #03223480006 for Universal Consultants, Inc., North Ridgeville, Ohio or other licensed service.

- **Safe Use of Unsealed License Material**

We have developed and will implement and maintain written procedures for safe use of unsealed byproduct material that meet the requirements of Appendix T of NUREG-1556, Vol.9, Rev. 2.

- **Spill, Emergency Surgery and Autopsy Procedures**

We have developed and will implement and maintain written procedures for safe response to spills of licensed material in accordance with Appendix N of NUREG-1556, Vol.9, Rev. 2.

**Robert Morris University
6001 University Boulevard
Moon Township, PA 15108-1189**

Item 10 Radiation Protection Program (continued)

- **Area Survey Records**

We have established, document and will implement the model procedures for area surveys as published in Appendix R of NUREG-1556, Vol.9, Rev. 2.

Item 11 Waste Management

We have developed and will implement and maintain written waste disposal procedures for licensed material in accordance as published in Appendix W of NUREG-1556, Vol.9, Rev. 2.

Duplicate

MATERIALS LICENSE

Duplicate

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p style="text-align: center;">Licensee</p> <p>1. Wheeling Jesuit University</p> <p>2. 316 Washington Avenue Wheeling, West Virginia 26003-6295</p>	<p>In accordance with the application dated February 15, 2005,</p> <p>3. License number 47-14258-01 is amended in its entirety to read as follows:</p> <hr/> <p>4. Expiration date July 30, 2015</p> <hr/> <p>5. Docket No. 03014938 Reference No.</p>
---------------------------------------------------------------------------------------------------------------------------------------------------------	------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

<p>6. Byproduct, source, and/or special nuclear material</p> <p>A. Cobalt 60</p> <p>B. Technetium 99m</p> <p>C. Cesium 137</p>	<p>7. Chemical and/or physical form</p> <p>A. Calibration or Reference Sources</p> <p>B. Any</p> <p>C. Sealed Sources (NEN Model NES-356, Isotope Products Laboratories Model RV-XXX series)</p>	<p>8. Maximum amount that licensee may possess at any one time under this license</p> <p>A. 60 microcuries</p> <p>B. 50 millicuries</p> <p>C. 700 microcuries</p>
--------------------------------------------------------------------------------------------------------------------------------	--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	-------------------------------------------------------------------------------------------------------------------------------------------------------------------

9. Authorized use:

A. Through C. Teaching and training of students.

CONDITIONS

10. Licensed material may be used or stored only at the licensee's facilities located at 316 Washington Avenue, Wheeling, West Virginia.
11. Licensed material shall be used by, or under the supervision of Angela R. Macci Bires or Vincent A Gargaro.
12. The Radiation Safety Officer for this license is Angela R. Macci Bires.
13. The licensee shall not use licensed material in or on human beings.

Duplicate

Duplicate

Duplicate

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number
47-14258-01

Docket or Reference Number
03014938

Amendment No. 08

14. A. Sealed sources shall be tested for leakage and/or contamination at intervals not to exceed six months or at the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State.
- B. In the absence of a certificate from a transferor indicating that a leak test has been made within the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State, prior to the transfer, a sealed source received from another person shall not be put into use until tested and the test results received.
- C. Sealed sources need not be tested if they are in storage and are not being used; however, when they are removed from storage for use or transferred to another person and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.
- D. The leak test shall be capable of detecting the presence of 0.005 microcurie (185 becquerels) of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie (185 becquerels) or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations.
- E. Tests for leakage and/or contamination, including leak test sample collection and analysis, shall be performed by the licensee or by other persons specifically licensed by the U.S. Nuclear Regulatory Commission or an Agreement State to perform such services.
- F. Records of leak test results shall be kept in units of microcuries and shall be maintained for 5 years.
15. The licensee shall conduct a physical inventory every six months, or at other intervals approved by the U.S. Nuclear Regulatory Commission, to account for all sources and/or devices received and possessed under the license. Records of inventories shall be maintained for 5 years from the date of each inventory and shall include the radionuclides, quantities, manufacturer's name and model numbers, and the date of the inventory.
16. The licensee is authorized to hold byproduct material with a physical half-life of less than or equal to 120 days for decay-in-storage before disposal without regard to its radioactivity if the licensee:
- A. Monitors byproduct material at the surface before disposal and determines that its radioactivity cannot be distinguished from the background radiation level with an appropriate radiation detection survey meter set on its most sensitive scale and with no interposed shielding; and

Duplicate

Duplicate

Duplicate

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number

47-14258-01

Docket or Reference Number

03014938

Amendment No. 08

- B. Removes or obliterates all radiation labels, except for radiation labels on materials that are within containers and that will be managed as biomedical waste after they have been released from the licensee; and
- C. Maintains records of the disposal of licensed materials for 3 years. The record must include the date of disposal, the survey instrument used, the background radiation level, the radiation level measured at the surface of each waste container, and the name of the individual who performed the disposal.
17. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
18. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations:
- A. Application dated February 15, 2005 (ML050820488)
- B. Letter dated May 24, 2005 (ML051600589)
- C. Letter dated June 28, 2005 (ML051860246)

For the U.S. Nuclear Regulatory Commission

Original signed by Todd J. Jackson, CHP

Date July 21, 2005

By

Todd J. Jackson, CHP
Commercial and R&D Branch
Division of Nuclear Materials Safety
Region I
King of Prussia, Pennsylvania 19406

Duplicate

Duplicate

Duplicate

OHIO DEPARTMENT OF HEALTH LICENSE FOR RADIOACTIVE MATERIAL

Pursuant to Chapter 3748 of the Ohio Revised Code, and in reliance on statements and representations made by the licensee, a license is hereby issued authorizing the licensee named herein to receive, acquire, possess, and transfer radioactive material as designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the applications of Chapter 3748 of the Ohio Revised Code and all applicable rules promulgated thereunder. This license is subject to all applicable rules, regulations and orders of the Ohio Department of Health now or hereinafter in effect and to any conditions specified below.

LICENSEE	LICENSE NUMBER
1. Universal Consultants, Inc. 2. 35888 Center Ridge Road Suite # 7 North Ridgeville, OH 44039	3. 03223480006
	EXPIRATION DATE
	4. September 1, 2009
	BUREAU DOCKET NUMBER
	5. KK27-10-00

6. RADIOACTIVE MATERIAL	7. CHEMICAL AND/OR PHYSICAL FORM	8. MAXIMUM QUANTITY THAT LICENSEE MAY POSSESS AT ANY ONE TIME UNDER THIS LICENSE
A. Cesium-137	A. Sealed source (New England Model NES-356 or any models approved by NRC or an Agreement State)	A. No single source to exceed 9.25 MBq (250 μ Ci). Total possession limit not to exceed 74 MBq (2 mCi)
B. Cobalt-60	B. Sealed source (New England Model NES-354 or any models approved by NRC or an Agreement State)	B. No single source to exceed 3.7 MBq (100 μ Ci). Total possession limit not to exceed 37 MBq (1 mCi)
C. Barium-133	C. Sealed source (New England Model NES-358 or any models approved by NRC or an Agreement State)	C. No single source to exceed 18.5 MBq (500 μ Ci). Total possession limit not to exceed 74 MBq (2 mCi)
D. Cobalt-60	D. Sealed source	D. 2 sources not to exceed 370 MBq (10 mCi) each
E. Any radioactive material	E. See Item 9.E. below	E. See Item 9.E. below
F. Technetium-99m	F. Any	F. 9.25 GBq (250 mCi)
G. Cesium-137	G. Sealed source (Amersham Model No. 77302)	G. 6.105 GBq (165 mCi)

9. Authorized Use

- A. through D, and F. To be used for instrument calibration and testing.
- E. For the possession incident to the performance of tests for leakage/contamination on sources and devices as specified in renewal application dated June 7, 2004.
- G. To be used in an Amersham Corp. (Formerly Tech/Ops) Model no. 773 device for instrument calibration.

OHIO DEPARTMENT OF HEALTH LICENSE FOR RADIOACTIVE MATERIALS SUPPLEMENTARY SHEET	Page <u>2</u> of <u>3</u>
	License Number: 03223480006
	Bureau Docket Number: KK27-10-00
	Amendment No. 3

CONDITIONS

10.
 - A. Licensed material may only be used at the licensee's facility located at 35888 Center Ridge Road, # 7, North Ridgeville, Ohio
 - B. Licensed material in Items 6.A., 6.B., and 6.C. may also be used at temporary job sites of the licensee anywhere in the State of Ohio. Tests for Leakage or contamination shall be performed only at temporary job sites of the licensee anywhere in the State of Ohio.
11. The Radiation Safety Officer for this license is:

 Vincent Gargaro
12. Licensed material shall only be used by Vincent Gargaro, Charles Leone, Ronald J. Scala, or Dale Robert Keith.
13. Sealed sources or detector cells shall not be opened or sources removed from source holders by the licensee.
14.
 - A. Sealed sources and detector cells shall be tested for leakage and/or contamination at intervals not to exceed 6 months or at such other intervals as specified by the certificate of registration referred to in *OAC 3701:1-46-09(C)*.
 - B. Notwithstanding Paragraph A of this condition, sealed sources designed to emit alpha particles shall be tested for leakage and/or contamination at intervals not to exceed three (3) months.
 - C. In the absence of a certificate from a transferor indicating that a leak test has been made within the required leak test interval, a sealed source shall not be put into service until tested.
 - D. Sealed sources need not be leak tested if:
 - (i). they contain only a radioactive gas; or
 - (ii). the half-life of the isotope is 30 days or less; or
 - (iii). they contain 3.7 MBq (100 μ Ci) or less of beta or gamma emitting material or 0.37 MBq (10 μ Ci) or less of alpha emitting material; or
 - (iv). they are in storage and are not being used. However, when they are removed from storage for use or transfer to another person, and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of 10 years without being tested for leakage and/or contamination.
 - E. The leak test shall be capable of detecting the presence of 185 Bq (0.005 μ Ci) of radioactive material on the test sample. If the test reveals the presence of 185 Bq (0.005 μ Ci) or more of removable contamination, a report shall be filed with the Ohio Department of Health, and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Ohio Department of Health Regulations. The report shall be filed within 5 days of the date the leak test result is known with the Ohio Department of Health, 246 N. High St., Bureau of Radiation Protection / 7th Floor 35 Bldg., P.O. Box 118, Columbus, Ohio 43216-0118. The report shall specify the source involved, the test results, and the corrective action taken.
 - F. Tests for leakage and/or contamination shall be performed by the licensee or by other persons specifically licensed by the Director, the USNRC, or an Agreement State to perform such services.
15. Licensed material shall not be used on human beings.
16. The licensee shall conduct a physical inventory every 6 months to account for all sources and/or devices received and possessed under the license. Records of inventories shall be maintained for 5 years from the date of each inventory.
17. The licensee is authorized to transport licensed material only in accordance with the provisions of Chapter 3701:1-50 of the Administrative Code.
18. The license does not authorize commercial distribution of licensed material.

OHIO DEPARTMENT OF HEALTH LICENSE FOR RADIOACTIVE MATERIALS SUPPLEMENTARY SHEET	Page <u>3</u> of <u>3</u>
	License Number: 03223480006
	Bureau Docket Number: KK27-10-00
	Amendment No. 3

19. In addition to the possession limits in Item 8, the licensee shall further restrict the possession of licensed material to quantities below the minimum limit specified in OAC 3701:1-40-17(C) for establishing decommissioning financial assurance.
20. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The Ohio Department of Health's statutes, rules, and orders shall govern unless statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- A. *Application dated June 7, 2004.*
- B. *Amendment No. 3 renews License No. 03223480006 in its entirety.*

For the Ohio Department of Health

DATE: 7/17/04 BY: 
 Director, Ohio Department of Health

This is to acknowledge the receipt of your letter/application dated

10/10/2007, and to inform you that the initial processing which includes an administrative review has been performed.

NEW LICENSE APPLICATION (03037563)
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 141130.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.

: (FOR LFMS USE)
 : INFORMATION FROM LTS
 : -----
 :
 :
 : Program Code: 03620
 : Status Code: 3
 : Fee Category: _____
 : Exp. Date: 0
 : Fee Comments: _____
 : Decom Fin Assur Req'd: _
 : ::::::::::::::::::::::::::::::::::::::::::::::::::::

BETWEEN:
 License Fee Management Branch, ARM
 and
 Regional Licensing Sections

LICENSE FEE TRANSMITTAL

A. REGION **I**

1. APPLICATION ATTACHED
 Applicant/Licensee: ROBERT MORRIS UNIVERSITY
 Received Date: 20071015
 Docket No: 3037563
 Control No.: 141180
 License No.: **37-31262-01**
 Action Type: New Licensee

2. FEE ATTACHED
 Amount: _____
 Check No.: /

3. COMMENTS

Signed *M. A. Perkins*
 Date 10/15/2007

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is entered /__/)

- 1. Fee Category and Amount: _____
- 2. Correct Fee Paid. Application may be processed for:
 Amendment _____
 Renewal _____
 License _____
- 3. OTHER _____

Signed _____
 Date _____