



Superior Well Services, Inc.
1380 RT 286 East Suite #121
Indiana, PA 15701
Telephone 724-465-8904

October 25, 2007

Nuclear Regulatory Commission – Region I
475 Allendale Road
King of Prussia, PA 19406-1415

Attn: Judith Joustra

Dear Judith Joustra:

Superior Well Services, Ltd. wishes to inform you and your department that the nuclear density gauge model 5192 serial number B7000, which was lost in August of 2006 near Vernal, UT, has been recovered.

The gauge was recovered south of Duchesne, UT along SR-191 by a member of the general public on October 19th, 2007.

I have included the leak test results taken from the gauge immediately after the recovery.

If you have any questions or if you require any additional information, please feel free to contact me.

Thank you for your assistance in this matter.

Respectfully,

Lew Cessna
HSE Coordinator
Superior Well Services, Ltd.
724-465-8904 Office
724-465-8325 Fax
lcessna@swwsi.com

2007 OCT 29 AM 10:30

RECEIVED
REGION 1

Enclosure: Leak test results for device serial number B7000

Applied Health Physics, Inc.

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: ahp.inc@comcast.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics, Inc.

Device Information

Device Serial Number:	B7000	Source Serial Number:	B7000
Device Manufacturer:	Thermo Electron	Source Manufacturer:	Thermo Electron
Device Model:	5192	Source Model:	5192
Device Type:	Make:	Location:	0298
Radioisotope: Cs-137	Activity: 200 mCi		

Leak Test Information

Leak Test Performed by:	Lee Atkinson	Leak Test Performed on:	10/19/2007
Leak Test Specimen Number:	A-027569	Leak Test Analysis Date:	10/25/2007
Leak Test Analysis Performed by:	K. Mobley		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

The analysis of this leak test sample was provided by Applied Health Physics, Inc. U.S. Nuclear Regulatory Commission License Number 37-09135 and State License Number PA-0228. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics, Inc. will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics, Inc. is required by our license to report leakage and/or contamination in excess of 0.005 microcuries.

Client Information

Superior Well Services, LTD (UT)	Contact:	John Brown / Lee Atkinson
1453 East 335 South	Phone:	(435) 781-0266
P.O. Box 1049	Contact:	Fax #
Attn: Lee Atkinson	Phone:	(435) 781-0270
Vernal	UT	84078

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 4/19/2008