

**PRIMECARE MEDICAL GROUP, P.A.**

November 5, 2007

*Bhudev Sharma, M.D., F.A.C.C.*  
*Raj S. Ballal, M.D., F.A.C.C.*  
*Mohammad K. Memon, M.D., F.A.C.C.*

INTERNAL MEDICINE  
NON-INVASIVE & INTERVENTIONAL CARDIOLOGY

James P. Dwyer, Chief  
Commercial and R&D Branch  
Division of Nuclear Materials Safety  
Nuclear Regulatory Commission  
Region I  
475 Allendale Road  
King of Prussia, PA 19400-1415

IR 2007001

Dear Mr. Dwyer,

**Reply to a Notice of Violation**  
**Docket No. 03034893 License No. 29-30488-01**

In response to your Notice of Violation dated September 28, 2007, as a result of the inspection of our facilities made by Mr. Todd Jackson on September 18, 2007, the following are the corrective steps we have taken so far:

- Violation A - 1. I have instructed Dave Muretic, our Nuclear Medicine technologist, to make sure that all personnel that are exposed to radiation, including the physicians, are regularly wearing the radiation badges assigned to them. And these badges will be regularly sent to Landauer for processing in a timely manner.
2. We will maintain records of occupational exposure of personnel from other employer to make sure that the dose is reduced by the amount received while employed by that other employer.

As of November 1, 2007, we are in full compliance of 10 CFR 20.1502(a)(1) and 10 CFR 20.1201(f).

- Violation B - 1. We already had purchased a 500 nci Cs-137 rod source to be used for calibrating our well counter.
2. On November 2, 2007, our consulting physicist Jessie Z, Trivino came and tried to calibrate the well counter. Unfortunately, he encountered some technical problem which will require the well counter to be sent to the manufacturer for repair and calibration. The manufacturer will lend us a loaner well counter to use while they have our unit for service.

REC'D IN LAT NOV - 9 2007


NMSS/RGNI MATERIALS-004

3. Once the unit is in full calibration and in proper working order, we will calibrate the unit annually and following any repair that affect the calibrations.

We expect to be in full compliance of 10CFR 35.61 by November 30, 2007.

Thank you very much for all the assistance you have extended us.

Sincerely,



Bhudev Sharma, M.D.  
President