

TRANSMISSION VERIFICATION REPORT

TIME : 11/15/2007 07:31  
NAME : USNRC RIII  
FAX : 6308299782  
TEL :  
SER. # : 000A7J925774

DATE, TIME	11/15 07:30
FAX NO./NAME	87344538851
DURATION	00:00:29
PAGE(S)	03
RESULT	OK
MODE	STANDARD ECM



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION III  
2443 WARRENVILLE ROAD, SUITE 210  
LISLE, ILLINOIS 60832-4352

TELEFAX TRANSMITTAL

DATE: November 15, 2007      NUMBER OF PAGES: 3  
*(Including this page)*

SEND TO: Ray Carlson

LOCATION: Radiological Physics Service

FAX NUMBER: (734) 453-8851       **VERIFY BY CALLING  
SENDER**

FROM:  
(SENDER) **Bill Reichhold**

TELEPHONE NUMBER (630) 829-9839      FAX NUMBER (630) 515-1078  
If you do not receive the complete fax transmittal, please contact the sender as soon as possible at the telephone number provided above.

MESSAGE



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION III  
2443 WARRENVILLE ROAD, SUITE 210  
LISLE, ILLINOIS 60532-4352

### TELEFAX TRANSMITTAL

DATE: November 15, 2007 NUMBER OF PAGES: 3  
(Including this page)

SEND TO: Ray Carlson

LOCATION: Radiological Physics Service

FAX NUMBER: (734) 453-8851  **VERIFY BY CALLING  
SENDER**

FROM:  
(SENDER) **Bill Reichhold**

TELEPHONE NUMBER (630) 829-9839 FAX NUMBER (630) 515-1078

If you do not receive the complete fax transmittal, please contact the sender as soon as possible at the telephone number provided above.

#### MESSAGE

This is regarding the response for the amendment to NRC License 21-24380-01 (Michael Lala, M.D.), Control 316534. Please have Ray Carlson confirm and sign that the corrected copy of Northway's sealed source inventory dated March 23, 2007 is accurate.

Please send a facsimile (630- 515-1078) of your response to the above by Monday, 19 November 2007, and refer to control 316534 . Please call me at 630-829-9839 if you have any questions.

Bill Reichhold

#### NOTICE

This message is intended only for the use of the individual or entity to which it is addressed and may contain information that is privileged, confidential, or exempt from disclosure under applicable law. If the reader of this message is not the intended recipient or the employee responsible for delivering the message to the intended recipient, you are hereby notified that any dissemination, distribution or copying of this communication is strictly prohibited. If you have received this communication in error, please notify the sender immediately by telephone and return the original to the above address, by U.S. Mail. Thank you.

**C O V E R**

**FAX**

**S H E E T**

**To:** Bill Reichhold  
**Fax #:** 630-515-1078  
**Subject:** Corrected Northway Inventory Page  
**Date:** November 13, 2007  
**Pages:** 2, including this cover sheet.

**COMMENTS:**

Ray A. Carlson, M.S.  
 Cell: (734) 395-7361

Don L. Marx, M.S.  
 Cell: (810) 730-6004

Laura Smith, M.S.  
 Cell: (586) 215-5947

Vince McCormick, M.S.  
 Cell: (734) 395-9323

Radiological Physics Service, Inc.  
 3830 Napier Road  
 Plymouth, MI 48170

(734) 455-4730  
 FAX: (734) 453-8851

M1073100107

**Location:** Northway Medical Center **Date of Inspection:** March 23, 2007

**SEALED SOURCE INVENTORY**

Nuclide	Type	Location	Calibration Activity	Calibration Date	Mfgr.	M/N	S/N
Cs-137	vial	Hot Lab	253.5 uCi	05/01/81	Amersham	CDC.VI	3148MA
Ba-133	vial	Hot Lab	270.3 uCi	06/01/81	Amersham	BCD.VI	2099MA
Cs-137	vial	Hot Lab	201 uCi	09/19/88	DuPont	NES-256	S256009-52
Cs-137	spot	Hot Lab	1 uCi		Nucleus		

RSO: \_\_\_\_\_

**SEALED SOURCE LEAK TEST**

Nuclide	Type	Location	Calibration Activity	Calibration Date	Mfgr.	M/N	S/N
Cs-137	vial	Hot Lab	253.5 uCi	05/01/81	Amersham	CDC.VI	3148MA
Ba-133	vial	Hot Lab	270.3 uCi	06/01/81	Amersham	BCD.VI	2099MA
Cs-137	vial	Hot Lab	201 uCi	09/19/88	DuPont	NES-256	S256009-52

**COMMENT:** The sources listed above were leak tested using a wipe technique and were found to have less than 0.005 uCi removable activity.

**TESTED BY:** Ray A. Carlson, M.S.

**DATE:** March 23, 2007

RSO: \_\_\_\_\_

*KAY,  
PLEASE CONFIRM AND  
SIGN THAT THE CORRECTED  
INVENTORY IS ACCURATE.*