



3058 Research Drive
 State College, Pennsylvania 16801 USA
 Telephone: 814.272.1039
 Fax: 814.272.1019

MPI Research, Inc.
 3058 Research Drive
 State College, PA 16801

15 Nov 07

NMSB2

Licensing Assistant Section
 Nuclear Materials Safety Branch
 U.S. Nuclear Regulatory Commission, Region I
 475 Allendale Road
 King of Prussia, PA 19406-1415

Re: License no. 37-30095-01
 Docket no. 030033359

Dear Sir or Madam:

As indicated to you in our letter of 12 Jul 07, MPI Research was in the process of decommissioning the Radioisotope Laboratory and disposing of all radioactive material in preparation for termination of license no. 37-30095-01. The decommissioning is now complete and all licensed material has been disposed. A copy of the Final Status Survey is enclosed. The survey documents that the facilities at MPI Research meet the requirements for unrestricted release. A Certificate of Disposition of Materials (NRC Form 314) is enclosed to verify the disposal of all licensed material.

MPI Research requests that license no. 37-30095-01 be terminated and thanks you for your prompt attention to this matter.

Sincerely,

Kevin Lloyd
 General Manager, Analytical Services

Enclosure: NRC Form 314
 Final Status Survey

RECEIVED
 REGION I
 2007 NOV 29 AM 10: 29

141361

NMSS/RGN1 MATERIALS-002

NRC FORM 314 (9-2007) 10 CFR 30.36(j)(1); 40.42(j)(1); 70.38(j)(1); and 72.54(k)(5)(1)(1)	U.S. NUCLEAR REGULATORY COMMISSION	APPROVED BY OMB: NO. 3150-0028 Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Records and FOIA/Privacy Services Branch (T-5 F52), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to infocollects@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.	EXPIRES: 08/31/2010
CERTIFICATE OF DISPOSITION OF MATERIALS			

LICENSEE NAME AND ADDRESS MPI Research, Inc., 3058 Research Drive State College, PA 16801	LICENSE NUMBER 37-30095-01	DOCKET NUMBER 030-33359
	LICENSE EXPIRATION DATE 05/31/2014	

A. LICENSE STATUS (Check the appropriate box)

This license has expired.
 This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

1. No radioactive materials have ever been procured or possessed by the licensee under this license.

2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.

a. Transfer of radioactive materials to the licensee listed below:

b. Disposal of radioactive materials:

1. Directly by the licensee:

2. By licensed disposal site:

3. By waste contractor:

Duratek Services, Inc., 1560 Bear Creek Road, Oak Ridge, TN 37831 865-481-0222

c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

C. SURVEYS PERFORMED AND REPORTED

1. A radiation survey was conducted by the licensee. The survey confirms:

a. the absence of licensed radioactive materials

b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.

2. A copy of the radiation survey results:

a. is attached; or b. is not attached (Provide explanation); or c. was forwarded to NRC on: 11/15/2007

Date

3. A radiation survey is not required as only sealed sources were ever possessed under this license, and

a. The results of the latest leak test are attached; and/or
 b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME Ralene Kreiser	TITLE Senior Safety Associate	TELEPHONE (Include Area Code) (800) 281-3219	E-MAIL ADDRESS ralene.kreiser@mpi-research.com
-------------------------------	---	--	--

Mail all future correspondence regarding this license to:
MPI Research, Inc., 3058 Research Drive, State College, 16801

C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE Kevin Lloyd, General Manager	SIGNATURE 	DATE 11/15/2007
---	---	---------------------------

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

141361

Final Status Survey
Nuclear Regulatory Commission License 37-30095-01

MPI Research
3058 Research Drive
State College, PA 16801

15 November 2007

Introduction

This report is the Final Status Survey that is required for the termination of Nuclear Regulatory Commission (NRC) byproduct material license number 37-0095-01. It is also intended to serve as a historical reference for MPI Research for information on facilities and the activities conducted under the NRC license.

Historical Information

Nuclear Regulatory Commission byproduct material license 37-30095-01 was issued to Centre Analytical Laboratories, Inc. at 3048 Research Drive, State College, PA 16801 on 28 Feb 94. The radiochemistry laboratory was located in room 6 of the basement of what is now Building 1. The activity limits were:

Carbon 14	95 millicuries
Phosphorus 32	6 millicuries
Sulfur 35	10 millicuries

In October 1996 the radioisotope laboratory was moved to room 209 of the new Building 2 next door to Building 1 at 3058 Research Drive. In March of 2002 the license was amended to also allow the use of licensed material in a new Building 3 located at 3117 Research Drive. The environmental fate studies using Building 3 were completed and it was removed from the license in September 2003 and released for unrestricted use. Building 3 sold shortly thereafter.

In February 2004 the radiochemistry laboratory was moved from Building 2 to the present location in room 101 of Building 1.

The following amendments have been made to the original license.

Amendment #1 was issued 4 Aug 94. It changed the Radiation Safety Officer to Rodger W. Granlund.

On 8 Apr 96 the NRC notified Centre Analytical Laboratories that the expiration date of the license had been extended to 28 Feb 2004.

Amendment #2 was issued 10 Oct 96. This amendment changed the location of the radioisotope lab to room 209 of Building 2.

Amendment #3 was issued 16 Apr 01. This amendment added 20 millicuries of ^3H to the license and acknowledged the change of the person responsible for the license from Michael Arjmand to Richard Grazzini.

Amendment #4 was issued on 12 Sep 01. This amendment acknowledged the name change to Exygen Research, Inc.

Amendment #5 was issued on 12 Mar 02. This amendment covered the use of licensed material in Building 3, 3117 Research Drive.

Amendment #6 was issued on 11 Sep 2003. This amendment removed Building 3 from the license and released the building for unrestricted use.

Amendment #7 was issued on 18 Feb 04. This amendment covered the relocation of the Radioisotope Laboratory from room 209 in Building 2 to room 101 of Building 1.

Amendment #8 was issued on 13 May 04. This amendment renewed the license until 31 May 2014.

Amendment #9 was issued on 12 Oct 06 and acknowledged the transfer of the ownership of Exygen Research to MPI Research.

Amendment #10 was issued 10 Aug 07 for the replacement of Richard Grazzini as the management person named on the license with Kevin Lloyd, General Manager. The letter for this amendment also notified the NRC of the intent of MPI Research to terminate the NRC license.

The current radioisotope and activity limits (any chemical or physical form) are:

Hydrogen 3	20 mCi
Carbon 14	95 mCi
Phosphorus 32	6 mCi
Sulfur 35	10 mCi

No ^{35}S was possessed or used under this license. The last ^{32}P acquired under the license was in April 2001, so any ^{32}P activity has been removed by radioactive decay. No ^3H was used in any projects. A NIST standard of tritiated water with an activity of 10.8 microcuries was received on 9 Oct 98. This was divided into several vial and flame sealed for storage. Some of the ^3H standard was used for calibration of the liquid scintillation counter; the remainder was disposed in the final waste shipment. Two items with 250 microcuries of ^3H each were received on 22 Nov 04. This material was stored frozen and never used. It was disposed in the final waste shipment. Therefore, only ^{14}C was of concern for the final status survey, although, those items that were used for storage of tritium-labeled compounds, such as the freezer(s), were checked for ^3H contamination. Also, the smears used to check for residual contamination were counted in a liquid scintillation counter in a window that would include tritium counts with those for ^{14}C .

Surveys

The previous locations of the radiochemistry laboratory in room 6 of Building 1 and room 209 of Building 2 were surveyed at the time the laboratories were moved. A closeout survey was also performed for Building 3 after radioisotope use was terminated there and before the building was sold. The guidelines published by the NRC in Regulatory Guide 1.86 in 1974 (NRC 1974) and in 1993 (NRC 1993) were used as the contamination limits. For ^{14}C the values were 5,000 dpm/100 cm^2 averaged over up to 1 m^2 , 15,000 dpm/100 cm^2 maximum for up to 100 cm^2 , and 1,000 dpm/100 cm^2 for

removable contamination. The current limits for building surface contamination published by the NRC in NUREG 1757 (NRC 2006) are based on specific radionuclides and the value for ^{14}C is $3.7\text{E}+6$ dpm/100 cm^2 . This value assumes 10% of the activity is removable and is based on a dose of 25 mrem/year for building occupants.

The closeout surveys for each of the laboratories are discussed below and copies of the surveys are included as attachments to this report. Building surfaces were all well within the limits for residual contamination and in most instances were below detection limits. The only surfaces with residual contamination were in the interiors of hood ductwork. These surfaces were also within the limits for unrestricted release.

Radioisotope Laboratory, Room 6, Building 1

This laboratory was used for ^{14}C work from February 1994 until October 1996. Routine surveys during this period did not detect any significant contamination. The cabinets, bench tops and floors were surveyed on 27 Feb 97. All smears for removable contamination were less than 100 dpm/100 cm^2 . No contamination was detectable with a thin-window GM detector and ratemeter (MDA=4500 dpm/100 cm^2) or with a scaler (MDA=100 dpm/100 cm^2).

The hood ductwork was the only location that any contamination was found. The hood exhaust duct passes through the wall to the outside then up along the outside wall to the roof. The exhaust fan and a short stack with a venturi to draw in extra dilution air are located on the roof. A filter box for a HEPA filter or an activated carbon filter is located just above the point where the duct emerges through the basement wall, but no filters were ever used.

A survey of the interior of the hood and the ductwork accessible at the top of the hood and at the filter box was conducted on 30 Dec 06. The baffles were removed from inside the hood to make all the interior surfaces and the inside of the ductwork available. Smears of the interior of the hood were all less than 100 dpm/100 cm^2 . One smear from the inside of the duct at the top of the hood showed 195 dpm/100 cm^2 of removable contamination. No contamination was detectable by direct survey of the surfaces with a thin-window pancake GM detector (MDA=4500 dpm/100 cm^2 with a ratemeter). The inside of the filter box was surveyed before cleaning to estimate the contamination level for the inaccessible surfaces of the ductwork. Smears from inside the filter box had 130 to 640 dpm/100 cm^2 . Direct counts inside the filter box with a thin-window pancake GM detector indicated up to 1900 dpm/100 cm^2 (MDA=1200 dpm/100 cm^2 for a 1 minute count with a scaler).

On 2 Nov 02 the fan and stack on the roof were opened for repairs, making some interior surfaces that were not normally accessible available to survey. The filter box and the interior of the hood were also rechecked at that time. The interior of the fan housing had a maximum of 170 dpm/100 cm^2 . The interior the exhaust stack had 430 dpm/100 cm^2 at the bottom and 40 dpm/100 cm^2 at the top. Smears from the interior of the filter box were all less than 100 dpm/100 cm^2 . Smears from inside the ductwork above and below the filter box were 250 and 160 dpm/100 cm^2 respectively. All smears from the hood and

a small attached canopy were less than 100 dpm/100 cm². There was no detectable contamination on any of the surfaces with a thin-window pancake GM detector and ratemeter.

All the smears for removable contamination and the GM measurements for total contamination in the Building 1 Radioisotope Laboratory were within the limits listed above (NRC 1993). No further surveys or decontamination are required. Copies of the laboratory and hood surveys are included as Attachment A.

Radioisotope Laboratory, Room 209, Building 2

The Radioisotope Laboratory was moved from the basement of Building 1 to room 209 Building 2 during Jan-Feb 97. The laboratory remained there until February 2004, when it was moved to the present location in Room 101 of Building 1. During that period ¹⁴C was used in the laboratory for a number of environmental fate studies. During the period December 1999 through May 2001 ³²P was also used in genetic studies. The closeout survey of Room 209 was performed in Jan-Feb 2004. A draft of the closeout survey report was prepared in March 2004. Through an oversight the final review was not completed until the draft was examined again in July 2007.

The only residual activity after cleaning was in the hood ductwork. Smears of the accessible ductwork were less than 100 dpm/100 cm². One side of the ductwork for hood 2-209-A had spots with 100-200 gross cpm with a pancake GM detector. This is equivalent to 3600-8700 dpm/100 cm² and is within the release limits for spot contamination. A copy of the closeout survey for Room 209 is enclosed as Attachment B

Building 3

The license was amended in March 2002 to allow the use of ¹⁴C in environmental fate studies in Building 3 at 3117 Research Drive. This building contained several environmental chambers that were used for long-term environmental fate studies until November 2002. It also contained a large walk-in freezer that was used for sample storage until July 2003. A closeout survey of Building 3 was conducted on 17 Jul 03. The survey was submitted to the NRC with a request to remove Building 3 from the license in August 2003. Amendment 6 to the license in September 2003 removed Building 3 from the license and released it for unrestricted use. Building 3 was sold after it was released for unrestricted use. A copy of the closeout survey that was submitted to the NRC for Building 3 is included as Attachment C.

Radioisotope Laboratory, Room 101, Building 1

In February 2004 the radioisotope laboratory was moved from room 209 of Building 2 to room 101 of Building 1. In June 2007 work to decommission the laboratory and terminate the license was started. All equipment in the laboratory was surveyed and decontaminated, as necessary, for unrestricted release. The work surfaces, floors, drawers, cabinets, walls and ceiling were all surveyed. With the exception of general license or exempt instrument calibration sources, all radioactive material was shipped for disposal on 3 Oct 07.

A hood on the first floor of Building 2 was used for a ^{14}C compound synthesis in May-June 2005 because the hood in room 1 of Building 1 was not large enough to accommodate the apparatus. There was no detectable contamination on accessible surfaces of the hood after cleaning. The hood was not disassembled to check the ductwork, but the contamination level at the exhaust baffles in the hood indicated that any ductwork contamination would be well below the limits for building surfaces. A hood on the third floor of Building 3 was also used for several weeks in 2006 for extractions of samples containing less than 1 microcurie of ^{14}C . No volatile ^{14}C compounds were involved and there was no detectable contamination at the conclusion of the work. The surveys for these hoods are included with the surveys for room 101 of Building 1. After decommissioning there was no detectable activity with survey meters on any accessible room surfaces or equipment and removable contamination was less than 10 dpm/100 cm². A copy of the closeout survey for room 101 Building 1 is included as Attachment D.

Conclusion

Centre Analytical Laboratories, Inc. was issued NRC byproduct material license number 37-30095-01 on 28 Feb 94. The license, as amended, allowed the possession and use of ^3H , ^{14}C , ^{32}P and ^{35}S . Only ^{14}C and ^{32}P were used in significant quantities and there has been no ^{32}P uses since 2001, so only ^{14}C is of concern for this survey. Licensed material was used in the Radioisotope Laboratories in room 6 of Building 1, room 209 of Building 2, room 101 of Building 1, and the environmental chambers of Building 3. Separate closeout surveys were conducted of these laboratories as work was completed and moved to a new location. The surveys for each of the four areas are included as attachments to this report.

All the laboratories were surveyed with a portable survey meter and smears on a weekly basis when licensed material was in use. These surveys show that contamination of the laboratory was limited to a few experiments and was removed at the completion of the work. Only minor contamination was detected during the Final Status Surveys and it was removed. At the conclusion of the closeout surveys for each area there was no contamination of accessible surfaces detectable with portable survey meters and removable contamination was less than 100 dpm/100 cm² (less than 10 dpm/100 cm² for the last closeout survey of room 1-101). The inaccessible hood ductwork of the radioisotope laboratories in room 6 of Building 1 and room 209 of Building 2 are known to contain some ^{14}C contamination, which is well below the limit for building surfaces and poses no hazard for persons using or maintaining the hoods.


In summary the Final Status Survey shows that the facilities previously used for work with licensed material are within the contamination limits for unrestricted release.

References

NRC 1974. Regulatory Guide 1.86, Termination of Operating Licenses for Nuclear Reactors, NRC June 1974.

NRC 1993. Guidelines for Decontamination of Facilities and Equipment Prior to Release for Unrestricted Use or Termination of Licenses for Byproduct, Source, or Special Nuclear Material, NRC, April 1993.

NRC 2006. Table B.1, NUREG-1757, v.1, rev. 2, Consolidated Decommissioning Guidance, Decommissioning Process for Materials Licensees, NRC Sep 2006

Prepared by:  15 Nov 07
Rodger W. Granlund, CHP
Radiation Safety Officer

Attachments: A. Closeout Survey, Radioisotope Laboratory, Room 6, Building 1
B. Closeout Survey, Radioisotope Laboratory, Room 209, Building 2
C. Closeout Survey, Environmental Fate Facilities, Building 3
D. Closeout Survey, Radioisotope Laboratory, Room 101, Building 1

Attachment A

**Closeout Survey
Radioisotope Laboratory
Room 6 Building 1**



Centre Analytical Laboratories, Inc.

3048 Research Drive
Phone: (814) 231-8032

State College, PA 16801
Facsimile: (814) 231-253

Page 5

RADIATION CONTAMINATION SURVEY


LOCATION: E. Fate Lab. Bldg. 1

DATE: 3 Dec 96


SURVEYOR: R.W. Grandlund, Hoeschung Ye

CONDITIONS: Decommissioning check of hood. Top + bottom baffle plates removed for cleaning

INSTRUMENT(S): Bicron Frisk Tech., Bicron surveyor-M, SGI-Inspector for fixed counts



- a. INCUBATOR
- b. REFRIGERATOR
- c. BENCH
- d. AMBIS
- e. FUME HOOD
- f. BENCH
- g. OXIDIZER
- h. BENCH
- i. HPLC
- j. BENCH
- k. SINK
- l. BENCH
- m. LSC



No.	DPM	LOCATION	No.	DPM	LOCATION
4	²⁵⁰ < 100	inside hood duct	16	< 100	top, behind baffle, rt
5	"	" " "	17	"	light, inside hood
6	"	" " "	18	"	rear behind baffle
7	"	Top baffle inside rt	19	"	" " "
8	"	" " " left	20	"	" " "
9	"	" " back rt	21	"	" " "
10	"	" " " left	22	"	side, left top
11	"	bottom baffle inside ^{ret}	23	"	" " bottom
12	"	" " ductside ³⁰	24	"	" rt top
13	"	Sash inside	25	"	" " bottom
14	"	Top front panel inside	26	"	base, left
15	"	Top, behind baffle, left	27	"	" , right

go to page 57

COMMENTS: GMSH check for fixed contamination α , < 100 cpm on contact w/ pancake GM.

NRC Guidelines - surface contamination

March 1994/0

Max. 15,000 dpm/100 cm² α Avg. 5,000 dpm/100 cm² α Removable 1,000 dpm/100 cm² α

RW



Centre Analytical Laboratories, Inc.

3048 Research Drive
Phone: (814) 231-8032

State College, PA 16801
Facsimile: (814) 231-1253

Page 57

RADIATION CONTAMINATION SURVEY


LOCATION: _____

DATE: 30 Dec 96


SURVEYOR: _____

CONDITIONS: continuation of page 56

INSTRUMENT(S): _____



- a. INCUBATOR
- b. REFRIGERATOR
- c. BENCH
- d. AMBIS
- e. FUME HOOD
- f. BENCH
- g. OXIDIZER
- h. BENCH
- i. HPLC
- j. BENCH
- k. SINK
- l. BENCH
- m. LSC



No.	DPM	LOCATION	No.	DPM	LOCATION
28	< 100	base, front	41	< 100	bench
29	"	sink	42	"	"
30	"	apron	43	"	floor
31	"	sash, outside	44	"	"
32	"	top, front panel, outside	45	"	bench
33	" 276	Filter box, cover inside	33	GMSM 1min count on surface	
34	" 486	" " base	34	" " " " " "	
35	" 690	" " "	35	" " " " " "	
36	" 127	" " rear	36	" " " " " "	
37	"	duct, inlet, inside			
38	"	smear lost up duct			
39	< 100 200	duct, inside			

$\frac{dpm}{5cm^2}$
 47
 78 " " 290
 67 " " 220
 38 " " 27
 1930 dpm
 100 cm²

Attire eff. = 0.15 for $\frac{1}{6}$
 Rec'd
 bkg = 37cpm
 30 Dec 96

COMMENTS: no filter was used in filter box

RCY



Centre Analytical Laboratories, Inc.

3048 Research Drive
Phone: (814) 231-8032

State College, PA 16801
Facsimile: (814) 231-1253

Page _____

RADIATION CONTAMINATION SURVEY

LOCATION: E. Fete Lab Bldg 1


DATE: 27 Feb 87

SURVEYOR: R. Gerxlund

Page 1 of 2

CONDITIONS: Decommissioning Survey

INSTRUMENT(S): _____



No.	DPM	LOCATION	No.	DPM	LOCATION
1	< 100	cabinet inside	13	< 100	cabinet inside
2		" "	14		drawer "
3		" "	15		cabinet "
4		" "	16		drawer "
5		bench top	17		drawer "
6		" "	18		cabinet "
7		" "	19		drawer "
8		" "	20		" "
9		" "	21		cabinets outside
10		cabinet inside	22		" "
11		drawers "	23		" "
12	↓	" "	24	↓	wall

COMMENTS: Used LSC vials in cabinet @ 10 & 13 (counted on 3/4/97 my)
GMSM @ hood base = 177 c/5m, GMSM @ bench top (B) = 190/5, floor @ 26 = 196 c/5m

March 1994/0

"THIS IS AN EXACT COPY OF
THE ORIGINAL DOCUMENT."

BY RJG DATE 30 Dec 97

RJG



Centre Analytical Laboratories, Inc.

3048 Research Drive
Phone: (814) 231-8032

State College, PA 16801
Facsimile: (814) 231-1253


Page _____

RADIATION CONTAMINATION SURVEY

LOCATION: _____
SURVEYOR: _____
CONDITIONS: _____

DATE: 27 Feb 97
Page 2 of 2

INSTRUMENT(S): _____



39

No.	DPM	LOCATION	No.	DPM	LOCATION
25	< 100	floor	37	< 100	cabinet inside
26		"	38		cabinet outside
27		wall	39		floor
28		File cabinet	40	↓	drum w/65C vials
29		" "			
30		Table top.			
31		End of cabinet			
32		drainboard			
33		"			
34		sink sides			
35		" bottom			
36	↓	cabinet inside			

RCC
27 Feb 97

COMMENTS: GMSM/wipe check @ - tubing and condensers under sink.
GMSM @ sink bottom 197 c/s min

March 1994/0

"THIS IS AN EXACT COPY OF
THE ORIGINAL DOCUMENT."


BY RCC DATE 30 Dec 97

RCC



RADIATION CONTAMINATION SURVEY

LOCATION: E-Fate Lab Bldg 1 DATE: 7 Mar 97
SURVEYOR: R.W. Granlund
CONDITIONS: Decommissioning survey
INSTRUMENT(S): Beckman LSC, SEI-I GMSM



No.	DPM	LOCATION	No.	DPM	LOCATION
1	< 100	bench top	13	< 100	drawer, inside
2		"	14		" "
3		"	15		wall
4		"	16		"
5		drawer inside	17		hood, left
6		" "	18		" right
7		cabinet "	19	330	cabinet inside
8		" "	20	< 100	" "
9		" "	21	50	cabinet inside after cleaning (19)
10		drawer "	22	< 100	shelves
11		" "	23		bench top
12		" "	24		"

COMMENTS: Location #21 GMSM indicates ~20-50 cpm above background,
cabinet cleaned and rechecked during survey. Other GMSM & Wipes OK
March 1994/0

"THIS IS AN EXACT COPY OF
THE ORIGINAL DOCUMENT."

BY RW DATE 30 Dec 97



Centre Analytical Laboratories, Inc.

3048 Research Drive
Phone: (814) 231-8032

State College, PA 16801
Facsimile: (814) 231-1253

Page 2 of 3

RADIATION CONTAMINATION SURVEY

LOCATION: E-Fate Lab Bldg-1
SURVEYOR: _____
CONDITIONS: _____

DATE: 7 Mar 97

INSTRUMENT(S): _____

No.	DPM	LOCATION	No.	DPM	LOCATION
25	< 100	drawers inside	37	< 100	drawers, inside
26		" "	38		cabinets, outside
27		cabinet "	39		" "
28		drawers "	40		" "
29		" "	41		floor
30		cabinets "	42		"
31		wall	43		"
32		benchtop	44	✓	"
33		"			
34		cabinet, inside			
35		drawer "			
36	✓	" "			

Reef
7 Mar 97

COMMENTS: _____

"THIS IS AN EXACT COPY OF THE ORIGINAL DOCUMENT."
BY Reef DATE 30 Dec 97



Centre Analytical Laboratories, Inc.

3048 Research Drive
Phone: (814) 231-8032

State College, PA 16801
Facsimile: (814) 231-1253


Page _____
Page 3 of 3

RADIATION CONTAMINATION SURVEY

LOCATION: E-Fate Lab Bldg-1
SURVEYOR: _____
CONDITIONS: _____

DATE: 7 Mar 97

INSTRUMENT(S): _____



No.	DPM	LOCATION	No.	DPM	LOCATION
45	< 100	tabletop	57	< 100	cabinets inside
46		benchtop	58		" "
47		"	59		" "
48		"	60		" "
49		cabinet inside	61		" outside
50		drawer "	62		" "
51		cabinet "	63		" "
52		drawer "	64	↓	" "
53		" "			
54		cabinet "			
55		drawers "			
56	↓	" "			

Ruey
7 Mar 97

COMMENTS: Background radiation level = 0.01 mrem/hr - 0.02 mrem/hr.

March 1994/0

"THIS IS AN EXACT COPY OF
THE ORIGINAL DOCUMENT."


BY Ruey DATE 30 Dec 97

Exact Copy from original
RWG
11 Nov 02

RADIOACTIVE CONTAMINATION SURVEY

LOCATION Rm 6 Bldg 1 DATE 8 Nov 02
 SURVEYOR R. Granlund
 CONDITIONS Check of canopy hood in old radchem lab
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M LSC

Canopy hood
RWG
8 Nov 02




NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	< 10	hood inside side 1	 RWG 8 Nov 02 		
2	< 10	" " " 2			
3	< 10	" " " 3			
4	< 10	" " " 4			
5	< 10	duct inside			
6	< 10	" "			
7	< 10	" outside			
8	< 10	canopy outside			
25	< 10	Inside 'T'			
		RWG 8 Nov 02			

COMMENTS Paper smears, LSC 20 min count x 2, bkg 29-27 cpm
EFF = 52-93% (color quench from dust), LLD = 3-6 dpm/smears, smears
covered ~ 300 cm², smears screened w/ Frisk Tech before LSC
 Count. Frisk tech efficiency appears to be ~ 0.57 x 0.11 or 0.063 c/c
 beta or about 1/2 that for ⁹⁹Tc beta.
 CALSURVEY/RWG
 REV. 14 Jul 00

RWG

RADIOACTIVE CONTAMINATION SURVEY

LOCATION Rm 6 Bldg 1 DATE 8 Nov 02
 SURVEYOR RW Grantlund
 CONDITIONS Old radiochem lab hood
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
9	26	base left	19	< 10	Front bypass cover inside
10	24	" rt	20	< 10	insiduct
11	< 10	side rt	21	< 10	" "
12	< 10	side left	22	< 10	inside panel-ductside
13	< 10	inside panel	23	< 10	" " " "
14	< 10	back panel left	24	< 10	" " hood side
15	< 10	" " right			
16	< 10	Top - left			
17	< 10	" - right			
18	< 10	sash, inside			

COMMENTS paper smears, no detectable contamination
w/GMSM (bkg ~ 30 cpm).

RWG


RADIOACTIVE CONTAMINATION SURVEY

LOCATION Inside filter box for Rm 6 Bldg 1 hood DATE 8 Nov 02

SURVEYOR R. Granlund

CONDITIONS Check of old radiachem hood ductwork

INSTRUMENT(S) Bicron Frisk-Tech ✓ Bicron Surveyor M Ludlum 3 ✓ LSC _____




NO.	DPM	LOCATION	NO.	DPM	LOCATION
26	< 10	Filter box inside base	Blank 8 Nov 02		
27	< 10	back			
28	< 10	side			
29	< 10	"			
30	11	Top			
31	< 10	cover			
* 32	245	inside duct above box			
* 33	155	" " below box			
		Blank 8 Nov 02			

COMMENTS paper smears. No detectable contamination w/ GMSM survey (Ludlum 3/4-9). * cloth smear.

RGW

RADIOACTIVE CONTAMINATION SURVEY

LOCATION Roof of Bldg I. Fan for Room 6 hood DATE 8 Nov 02
 SURVEYOR R. Granlund
 CONDITIONS check of ductwork for old radioactive lab hood.
 INSTRUMENT(S) Bicron Frisk-Tech ✓ Bicron Surveyor M LSC ✓
hood num 3 ✓



NO.	DPM	LOCATION	NO.	DPM	LOCATION
34	10	Impeller upstream side	Plus 8 Nov 02		
35	93	" downstream "			
36	168	" side			
37	34	Fan housing, inside			
38	434	Exhaust duct, bottom, inside			
39	40	" " top "			
40	14	Fan base under housing drain			
		Plus 8 Nov 02			

COMMENTS cloth smears. No detectable contamination
w/ GM S/M (Bkg ~30 cpm)

Plus

ID: SMears

8 NOV 2002 17:03

USER: 5

COMMENT: C-14 SMears 1 MIN LSC2

PRESET TIME : 20.00

DATA CALC : SL DPM HH : YES SAMPLE REPEATS: 1 PRINTER : EDIT
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : STD
 TWO PHASE : NO ACC : NO CYCLE REPEATS : 2 DISK : EDIT
 SCINTILLATOR: LIQUID LUMEX: YES LOW SAMPLE REJ: 0 RWM LIST : OFF
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: none

ISOTOPE 1: 14C %ERROR: 0.00 FACTOR: 1.000000 BKG. SLB: 0
 WIDE OPEN WINDOW %ERROR: 0.00 FACTOR: 1.000000 BKG. SLB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low:-10.44 High:302.29

SAM NO	POS	TIME MIN	HH	14C		WIDE		14C	14C	LUMEX	ELAPSED TIME
				CPM	%ERROR	CPM	%ERROR	DFM	EFF-1	%	
B1	**1	20.00	94.6	25.50	9.65	43.25	7.16	27.45	92.90	10.18	21.26
B2	**2	20.00	92.9	25.05	9.29	41.65	7.09	26.95	92.96	4.89	42.40
B3	**3	20.00	93.2	22.40	9.73	37.45	7.44	24.10	92.95	3.85	63.52
Blank Average				DFM	for	14C :	26.16	COEF. OF VAR:	6.908		
1	**5	20.00	98.0	22.45	10.46	38.90	7.63	-1.96	92.76	11.92	84.84
2	**6	20.00	100.1	23.85	9.90	39.25	7.49	-0.43	92.68	9.65	106.10
3	**7	20.00	92.6	20.80	10.68	36.45	7.79	-3.79	92.98	9.77	127.34
4	**8	20.00	103.6	21.00	12.58	36.20	8.74	-3.47	92.53	28.05	148.87
5	**9	20.00	126.4	22.30	11.23	36.20	8.31	-1.76	91.36	20.30	170.31
6	**10	20.00	126.2	24.15	9.37	36.95	7.50	0.26	91.38	4.14	191.44
7	**11	20.00	128.2	23.00	9.89	38.25	7.50	-0.96	91.26	7.21	212.65
8	**12	20.00	122.4	22.30	11.22	36.95	8.20	-1.82	91.59	19.74	234.05
9	**1	20.00	211.5	39.70	10.60	50.65	8.81	24.38	79.55	49.17	256.30
10	**2	20.00	198.1	39.65	11.20	51.25	9.16	22.71	81.13	53.67	278.37
11	**3	20.00	100.6	23.30	13.74	37.80	9.30	-1.02	92.66	39.27	300.09
12	**4	20.00	101.8	27.55	12.65	43.15	9.05	3.59	92.60	43.49	321.95
13	**5	20.00	114.9	28.35	18.62	41.80	13.22	4.65	91.99	72.57	343.77
14	**6	20.00	107.4	24.55	9.45	39.95	7.78	0.42	90.71	5.88	365.77
15	**7	20.00	142.5	25.80	8.79	38.35	7.89	2.40	90.33	16.51	387.14
16	**8	20.00	123.3	27.10	14.85	40.35	10.76	3.44	91.54	57.20	409.27
17	**9	20.00	125.5	30.50	9.97	45.00	7.74	7.20	91.42	26.11	430.85
18	**10	20.00	104.4	30.05	8.60	45.30	6.88	6.33	92.99	7.03	452.08
19	**11	20.00	126.8	27.15	9.99	40.20	7.28	3.56	91.34	6.50	473.29
20	**12	20.00	119.0	30.15	9.36	43.60	7.48	6.69	91.78	18.29	494.72
21	**1	20.00	117.0	30.00	8.93	44.55	7.13	6.49	91.88	12.06	516.15
22	**2	20.00	126.0	23.90	10.14	37.55	7.81	-0.01	91.39	12.89	537.46
23	**3	20.00	111.7	26.05	9.34	39.25	7.45	2.10	92.15	8.49	558.71
24	**4	20.00	114.7	26.75	18.10	42.05	12.24	2.91	92.00	68.15	581.16
25	**5	20.00	146.0	23.90	9.66	37.25	7.59	0.37	90.08	7.22	602.38
26	**6	20.00	128.0	29.50	9.15	43.75	7.28	6.16	91.27	13.94	623.73
27	**7	20.00	130.7	32.50	8.27	46.65	6.80	9.51	91.11	7.39	644.92
28	**8	20.00	130.3	30.20	8.62	43.40	7.07	6.97	91.13	8.10	666.23
29	**9	20.00	129.6	30.50	8.56	42.95	7.10	7.29	91.12	7.95	687.48
30	**10	20.00	146.3	33.05	8.36	45.55	6.99	10.53	90.06	10.55	708.79
31	**11	20.00	133.5	28.70	8.93	44.15	7.04	5.40	90.94	8.10	730.05
32	**12	20.00	190.6	221.75	3.08	211.75	3.97	242.09	82.52	2.94	751.42
33	**1	20.00	132.0	142.45	3.78	152.90	3.65	151.03	90.39	1.77	772.74
34	**2	20.00	154.1	32.35	8.64	46.70	7.00	10.00	89.46	12.77	794.09
35	**3	20.00	325.1	65.20	5.71	71.10	5.45	98.73	52.20	5.56	815.35

WARNING: QUENCH VALUE IS OUTSIDE QUENCH LIMIT

SAM NO	POS	TIME MIN	HH	14C		WIDE		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR	CPM	%ERROR				
36	**4	20.00	373.7	78.05	5.18	84.30	4.98	167.07	40.39	4.31	836.73
WARNING: QUENCH VALUE IS OUTSIDE QUENCH LIMIT											
37	**5	20.00	212.8	47.05	6.73	59.85	5.93	33.94	78.28	5.06	887.96
38	**6	20.00	233.2	346.20	2.42	356.90	2.38	428.32	76.17	1.38	879.26
39	**7	20.00	162.0	56.20	6.14	69.65	5.48	37.12	88.81	4.70	900.48
40	**8	20.00	191.6	32.95	7.93	44.15	6.82	13.85	82.34	2.88	921.60
41	**9	20.00	98.4	23.60	9.97	40.60	7.37	-0.72	92.75	9.42	942.89
42	**10	20.00	94.6	23.50	10.07	38.65	7.60	-0.87	92.90	10.84	964.17

b/k
b/k

Dvicle 2 of 2

B1	**1	20.00	95.1	23.85	9.35	40.35	7.13	25.68	92.88	2.60	986.30
B2	**2	20.00	92.7	24.35	9.23	40.75	7.08	26.19	92.97	2.44	1007.40
B3	**3	20.00	94.2	20.80	10.04	35.80	7.58	22.39	92.91	3.00	1028.49

Blank Average DPM for 14C : 24.75 COEF. OF VAR: 8.340

1	**5	20.00	98.7	21.45	10.59	37.20	7.88	-1.62	92.74	13.83	1049.83
2	**6	20.00	100.9	22.55	10.35	37.85	7.71	-0.41	92.64	11.29	1071.09
3	**7	20.00	93.0	20.60	11.07	35.95	8.00	-2.59	92.96	13.46	1092.40
4	**8	20.00	104.4	20.00	12.38	35.00	8.63	-3.13	92.49	23.64	1113.85
5	**9	20.00	126.2	21.95	10.71	36.05	8.01	-0.73	91.38	13.81	1135.17
6	**10	20.00	125.7	21.95	10.39	37.50	7.69	-0.74	91.40	10.01	1156.43
7	**11	20.00	127.4	23.75	9.99	37.55	7.71	1.26	91.30	10.84	1177.70
8	**12	20.00	127.3	23.40	12.22	37.40	8.86	0.87	91.31	32.09	1199.30
9	**1	20.00	210.9	41.00	10.65	50.80	9.03	27.37	78.66	51.70	1221.55
10	**2	20.00	195.6	40.50	11.95	50.85	9.93	24.88	81.60	60.05	1243.89
11	**3	20.00	132.1	24.00	12.32	39.35	8.63	1.17	92.59	31.07	1265.21
12	**4	20.00	103.7	25.15	12.79	41.05	7.15	2.43	92.52	39.96	1287.28
13	**5	20.00	117.0	28.95	20.26	45.25	13.56	6.75	91.89	75.86	1310.14
14	**6	20.00	109.0	26.15	9.07	39.70	7.27	3.59	92.28	5.07	1331.30
15	**7	20.00	145.0	26.70	10.07	40.80	7.77	4.86	90.16	18.99	1352.73
16	**8	20.00	124.9	29.25	10.50	41.50	8.30	7.23	91.45	30.22	1374.35
17	**9	20.00	127.1	30.45	9.41	45.70	7.34	8.59	91.33	19.11	1395.81
18	**10	20.00	106.5	31.45	8.29	48.80	6.56	9.29	92.40	5.08	1417.02
19	**11	20.00	126.6	28.80	8.65	43.70	6.93	6.78	91.35	5.11	1438.20
20	**12	20.00	119.8	32.55	8.26	46.35	6.82	10.73	91.74	7.51	1459.45
21	**1	20.00	118.1	29.60	8.53	44.70	6.86	7.48	91.83	5.14	1480.72
22	**2	20.00	127.0	26.65	9.01	40.70	7.19	4.43	91.33	5.38	1501.90
23	**3	20.00	114.1	24.10	9.45	39.80	7.25	1.44	92.03	4.93	1523.07
24	**4	20.00	116.2	31.65	14.93	46.20	10.87	9.68	91.93	63.31	1545.43
25	**5	20.00	145.1	23.95	9.66	37.40	7.58	1.82	90.15	7.37	1566.65
26	**6	20.00	127.6	27.30	9.03	42.50	7.11	5.15	91.29	7.11	1587.92
27	**7	20.00	128.8	31.40	8.26	45.00	6.72	9.67	91.22	4.77	1609.05
28	**8	20.00	128.8	31.40	8.26	45.00	6.72	9.60	91.11	4.60	1630.30
29	**9	20.00	131.3	29.00	8.59	41.05	7.15	7.07	91.13	4.96	1651.39
30	**10	20.00	143.6	31.30	7.81	47.55	6.67	14.36	90.36	10.11	1672.66
31	**11	20.00	133.1	27.20	8.86	41.50	7.34	11.21	91.33	19.11	1693.93
32	**12	20.00	131.1	27.20	8.86	41.50	7.34	11.21	91.33	19.11	1715.20
33	**1	20.00	154.9	30.70	9.57	43.70	7.04	9.59	89.40	2.11	1736.47
34	**2	20.00	237.2	66.55	5.63	70.00	5.33	57.77	59.10	1.11	1757.74
35	**3	20.00	133.1	36.60	9.60	41.50	7.34	167.85	91.15	4.09	1800.07
WARNING: QUENCH VALUE IS OUTSIDE QUENCH LIMIT											
36	**4	20.00	154.9	46.85	6.67	58.40	5.95	34.49	79.09	3.35	1821.25
37	**5	20.00	221.1	355.60	2.38	364.60	2.35	439.51	76.59	0.85	1842.47
38	**6	20.00	159.9	59.60	5.88	72.30	5.32	42.23	88.96	2.62	1863.63
39	**7	20.00	191.6	32.55	7.92	45.65	6.67	14.78	81.35	1.11	1884.80

SAM NO	POS	TIME MIN	H#	14C		WIDE		14C	14C	LUNEX %	ELAPSED TIME
				CPM	%ERROR	CPM	%ERROR	DFM	EFF-1		
41	**9	20.00	98.2	22.20	9.94	36.85	7.37	-0.62	93.75	5.55	1877.91
42	**10	20.00	94.1	31.25	7.76	37.20	7.56	-0.75	93.75	5.55	1877.09

Attachment B

**Closeout Survey
Environmental Fate Laboratory
Room 209 Building 2**

**EXYGEN RESEARCH
RADIOISOTOPE LABORATORY CLOSEOUT SURVEY
Building 2, Room 209**

SUMMARY

The radioisotope laboratory in Room 209 of Building 2 has only been used with ^{14}C and ^{32}P . No ^{32}P has been received since May 2001. It has all decayed to undetectable levels and only ^{14}C could be present. All areas of the laboratory and all equipment used with radioactive material were checked for contamination using a pancake GM probe and smears assayed with a liquid scintillation counter. Historically there was only one incident that resulted in significant contamination. There was some contamination still present on the inside surfaces of the exhaust system for hood 209-B. The inside panels of the hood were removed and the contamination removed. There was also contamination on the inside of the ductwork for this hood, primarily on the outside curve of the first elbow above the hood. This contamination was also removed. Any remaining contamination on the inaccessible inside surface of the ductwork between the hood and the roof should be within the acceptable limits for unrestricted release.

The survey shows that there is no significant residual radioactive contamination in Room 209 of Building 2 in excess of the release criteria in NUREG 1556 v7 Table Q.2 for release of equipment (15,000 dpm/100 cm² maximum, 5,000 dpm/100 cm² average and 1,000 dpm/100 cm² removable) and in Table Q.3 for building surfaces (3.7E6 dpm/100 cm²). Room 209 may be released for unrestricted use.

This report was prepared 1 March 2004. Final review and minor editorial revisions were made 14 Jun 07.

INTRODUCTION

The radioisotope laboratory was moved from the basement of Building 1 to Room 209 of Building 2 after the issuance of amendment 3 to the NRC license in October of 1996. The only radioactive isotopes used in this laboratory were ^{14}C and ^{32}P . The last ^{32}P received was in May 2001, so it has decayed to undetectable levels and only ^{14}C is of concern. There are plated disc sources of $^{90}\text{Sr}/^{90}\text{Y}$ and ^{99}Tc for calibration of the pancake GM probes and ^{14}C and ^3H solutions for the calibration of the liquid scintillation counters. The calibration sources are license exempt quantities and their use has not resulted in any contamination. However, if any contamination from the calibration sources was present, it would have been detected in the wide ^{14}C channel of the liquid scintillation counter that was used to assay the samples in this survey.

Work areas are checked with a pancake GM survey meter each day that unsealed radioactive material is used. Weekly checks are made with smears counted with a pancake GM tube or a liquid scintillation counter. There has been only one incident involving significant contamination in the laboratory. In August and September 2002 some ^{14}C stock solution contamination was purified using thin layer chromatography. When the material was scraped from the glass TLC plate, some of the fine particles contaminated the work area and one hood.

Removable contamination levels were up to 1,300 dpm/100 cm² on the benchtop and up to 13,000 dpm/100 cm² in the rear air intake of the hood. The contamination was removed from the benchtop and the accessible sections of the hood, but it was assumed that some contamination remained in the upper, covered portions of the hood and inside the ductwork. Therefore, in the interior panels of both hoods were removed to make the air intake and ductwork connections of the hoods accessible for this survey.

In addition to the benchtops and hoods all the other parts of the lab were checked for contamination, including the drawers, cabinets, shelves, floor, walls, ceiling, sink and equipment.

SURVEY METHOD

Surfaces were checked with a portable survey meter with a pancake GM probe for any detectable contamination. The surfaces were also checked for removable contamination with paper smears counted in the wide ¹⁴C window of a liquid scintillation counter. A scaler with a GM pancake probe for direct counting of smears was also available, but the liquid scintillation counter is more sensitive and was used in this survey.

The equipment and materials used for the survey included.

- Beckman 6500 liquid scintillation counter.
- Bicron Surveyor M survey meter with PGM probe.
- National Diagnostics Ecoscint liquid scintillation fluid (10 ml/sample)
- Plastic scintillation vials, 20 ml.
- Acme paper smears, 1 5/8" diameter.
- RC5BCTWF 42-mm cloth smears (for sink drains and some hood areas).

SENSITIVITY

The liquid scintillation counter has a background range of 20-30 cpm in the wide ¹⁴C channel and an efficiency of 90+%. Smears were counted for 5 minutes each giving an MDA ($4.65 \sigma_{\text{bkg}}$) of about 15 dpm/smear. Results less than 100 dpm are simply reported as <100 dpm. Assuming that smears cover an area of 100 cm², this is equal to a sensitivity of 15 dpm/100 cm² for removable contamination.

The background for the portable GSM is about 30 cpm and the efficiency for ¹⁴C (as determined with ⁹⁹Tc) is 0.13 count/beta. A count rate increase of about 3 times background is usually detectable with the audible indicator in a survey for surface contamination. This is equivalent to 90 cpm/15 cm² or a sensitivity of about 4,600 dpm/100 cm² for total contamination.

Both of these survey methods provide adequate sensitivity to meet the criteria for the maximum ¹⁴C contamination in NUREG 1556 v7 Table Q.2 for release of equipment (15,000 dpm/100 cm² maximum, 5,000 dpm/100 cm² average and 1,000 dpm/100 cm² removable) and in Table Q.3 for building surfaces (3.7E6 dpm/100 cm²).

RESULTS

The results are presented below for each area of the laboratory. Copies of the individual survey forms are included with the report.

Benchtops

Three surveys with 29 smears included benchtops. There was no detectable contamination with the GMSM and all smears were <100 dpm/100 cm².

Cabinets, shelves and drawers.

Five surveys with 23 smears included cabinets, shelves and drawers. There was no detectable contamination with the GMSM and all smears were <100 dpm/100 cm².

Floor

There were 3 surveys of the floor with 22 smears. There was no detectable contamination with the GMSM and all smears were <100 dpm/100 cm².

Walls

Two surveys included 9 smears of wall surfaces. There was no detectable contamination with the GMSM and all smears were <100 dpm/100 cm².

Ceiling

One survey included a ceiling tile from each of the three areas of the lab. There was no detectable contamination with the GMSM and all smears were <100 dpm/100 cm² (p.26).

Hoods

The laboratory contains two large hoods, 209-B (north) and 209-A (south). The hoods are adjacent, but have separate exhaust fans and ductwork. Removable panels in the back and top of the hood with adjustable openings provide the exhaust passage for the hoods. Each hood has two exhaust ducts at the top. The ducts are joined in a Y to the duct that goes to the fan on the roof above the third floor. With the top panel removed it is possible to reach up in the duct to where the two exhaust ducts make the Y connection to the duct that goes to the roof. Hood 209-A (nearest the sink) was the hood primarily used with radioactive material, although the 209-B was occasionally used for such work. As indicated above in the Introduction, 209-A was known to contain some contamination in the inside of the back panels and possibly in the ductwork. The back and top panels were removed in both hoods to make these areas accessible.

The top and back panels of hood 209-B were removed to allow access to the interior of the exhaust passage and the ductwork. There was no detectable contamination with the GMSM and all smears were <100 dpm/100 cm² (p. 5).

The survey of hood 209-A before removal of the panels showed no detectable contamination with the GMSM. One smear on the ceiling of the hood had 100 dpm and all others were <100 dpm/100 cm² (p. 14). Before cleaning the left side of the ductwork Y connection showed no detectable contamination with the GMSM and <100 dpm/100cm² with smears. Smears from the

right side were 230 to 310 dpm/100 cm² and there was one area on the inside of the elbow that showed about 200 gross cpm with the GSM (about 8,700 dpm/100 cm²) (p. 15). Before cleaning the inside surfaces of the back of the hood the panels indicated 200-400 gross cpm with the pancake GM probe or 8,700-19,000 dpm/100 cm² (p. 16). After cleaning, smears from the panels and the back and top of the hood were all <100 dpm/100 cm² and there was no detectable activity with the GSM (p. 16, 17, 19). After wiping, the inside surface of the right side of the ductwork had a gross count rate with the GSM of 100-200 cpm (3600-8700 dpm/100 cm²) (p. 19) and all smears were <100 dpm/100 cm² (p. 20).

Sinks

The laboratory has three sinks. Only the sink on the south wall near hood 209A has been used with radioactive material for washing labware and for the discharge of liquid waste. There is also a sink on the north wall and a small cup sink in hood 209-A. The traps in all three sinks were disassembled and the interior wiped with cloth smears to check for contamination. The water in the traps was also assayed for ¹⁴C.

The smear from the inside of the inlet pipe to the trap of the sink on the south wall had 360 dpm. The smears from the trap and the outlet pipe from the trap were <100 dpm. The water in the trap had no detectable ¹⁴C.

The sink on the north wall had no detectable activity on the smears or the liquid in the trap.


The smears from the cup sink in the hood had <100 dpm. The water sample from the trap had 120 dpm/2 ml (2.7E-5 μCi/ml or about 2.7E-3 μCi in 100 ml). The sink has not been used so the activity was probably from some of the particles from TLC plates that caused the hood contamination. The water from the trap was absorbed and put in the solid ¹⁴C waste for disposal and the trap was reassembled.

Laboratory Equipment

The following equipment items were surveyed. There was no detectable contamination with the GSM and all smears were <100 dpm/100 cm².

- Sonicator #2 (p. 1)
- Nitrogen evaporator (p. 1)
- Pyro Magnestir #4 (p. 2)
- Mini-Vortex #2 (p. 2)
- Sonicator #5 (p. 2)
- Centrifuge #7 (p. 2)
- Incubator #5 (p. 3)
- Incubator #6 (p. 3)
- Computer PC00420 (p. 4)
- Biological material oxidizer #1 (p. 4)
- Biological material oxidizer #2 (p. 4)
- Balance #11 (p. 6)
- Computer PC898 (p. 6)


TLC Scanner #1 (p. 6)
TLC Scanner #2 (p. 6)
P-10 cylinder (p. 6)
Oxygen cylinder (p. 6)
Refrigerator/freezer #24 (p. 7)
Multi-Magnestir (p. 7)
Lamp (p. 8)
Porta-Trace lightbox (p. 8)
Recirculator (p. 8)
Soxhlet extractor #2 (p. 8)
Soxhlet extractor #3 (p. 8)
Stability chamber #4 (p. 9)
HPLC #3 (p. 10)
HPLC#7 (p. 11)
Refrigerator/freezer #13 (p. 25)

Prepared by: Rodger W. Granlund, CHP  Date: 14 Jun 07
Radiation Safety Officer

Attachments: Survey forms (26 pages) plus counter printouts (21 pages).

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B2-209 DATE 1/7/04
 SURVEYOR Brent McCracken
 CONDITIONS SONICATOR 2, NITROGEN EVAPORATOR 1
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC




NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	2100	OUTSIDE OF SONICATOR			
2	2100	OUTSIDE OF BOX			
3	2100	INSIDE OF BOX			
4	2100	TIPS			
5	2100	BASE INSIDE			
6	2100	SAMPLE HOLDER			
7	2100	NEEDLES			
8	2100	OUTSIDE OF BOTTOM			

Handwritten notes in table:
 - A diagonal line is drawn across the right half of the table.
 - 'BAM 1/7/04' is written in the right half.
 - 'BAM 1/7/04' is written in the bottom left corner of the table area.

COMMENTS GMSM SURVEY OKAY, 2100 cpm NCPM BAM 1/7/04

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B2-209 DATE 1/7/04
 SURVEYOR BRENT McCracken
 CONDITIONS PYRO MULTI-MAGNESTIR #4, SONICATOR #5, MINI VORTEX #2, CENTRIFUGE #7
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC





NO.	DPM	LOCATION	NO.	DPM	LOCATION
9	2100	OUTSIDE OF MAGNESTIR	EXAM 1/7/04		
10	2100	TOP OF MAGNESTIR			
11	2100	OUTSIDE OF VORTEX			
12	2100	OUTSIDE OF SONICATOR			
13	2100	INSIDE OF SONICATOR			
14	2100	OUTSIDE OF CENTRIFUGE			
15	2100	INSIDE BOTTOM			
16	2100	BUCKETS			
17	2100	ARMS			
18	2100	LIGHT GREEN HOLDERS			

COMMENTS GAUSS SURVEY OKAY. BAMA 1/7/04
2100 NCPM

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B2-209 DATE 1/7/04
 SURVEYOR BRENT M. CRACKEN
 CONDITIONS INCUBATOR 5 & 6
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
19	2100	OUTSIDE DOOR #5			
20	2100	INSIDE DOOR #5			
21	2100	INSIDE INCUBATOR #5			
22	2100	RACKS IN #5			
23	2100	OUTSIDE DOOR #6			
24	2100	INSIDE DOOR #6			
25	2100	INSIDE INCUBATOR #6			
26	2100	RACKS IN #6			
		BAM 1/7/04			
		BAM 1/7/04			

COMMENTS GMSM SURVEY OKAY. BAM 1/7/04
2100 NCPM

ID: C-14 CHECK

7 JAN 2004 15:4

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC1
 PRESET TIME : 5.00
 DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : ST
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : OF
 TWO PHASE : NO AGC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: none

ISOTOPE 1: 14C %ERROR: 0.50 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low: 0.000 High: 331.13

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	** -1	5.00	79.0	25.80	17.61	26.89	95.95	5.80	5.51
		Blank Average		DPM for	14C :	26.89	COEF. OF VAR:	0.000	
1	** -3	5.00	105.1	41.20	13.93	16.47	95.01	29.93	11.21-1
2	** -4	5.00	107.6	63.60	11.22	40.12	94.91	24.43	16.88-2
3	** -5	5.00	99.3	43.60	13.55	18.89	95.24	17.15	22.52-3
4	** -6	5.00	96.0	32.80	15.62	7.51	95.36	24.79	28.20-4
5	** -7	5.00	97.5	84.60	9.72	61.88	95.31	43.93	33.94-5
6	** -8	5.00	118.1	73.80	10.41	51.25	94.44	12.06	39.59-6
7	** -9	5.00	103.8	67.20	10.91	43.80	95.06	9.35	45.23-7
8	** -10	5.00	118.0	38.80	14.36	14.19	94.45	14.42	50.88-8
9	** -11	5.00	100.2	41.00	13.97	16.18	95.21	13.47	56.51-9
10	** -12	5.00	109.1	32.20	15.76	7.06	94.85	15.49	62.16-10
11	** -1	5.00	100.3	46.80	13.07	22.27	95.20	22.45	67.94-11
12	** -2	5.00	103.9	34.80	15.16	9.72	95.06	12.43	73.59-12
13	** -3	5.00	100.7	37.80	14.55	12.82	95.18	9.56	79.23-13
14	** -4	5.00	99.1	40.80	14.00	15.95	95.25	12.29	84.88-14
15	** -5	5.00	114.3	35.00	15.12	10.10	94.62	8.40	90.51-15
16	** -6	5.00	125.3	93.40	9.25	72.37	94.09	4.04	96.14-16
17	** -7	5.00	117.9	35.00	15.12	10.17	94.45	8.60	101.78-17
18	** -8	5.00	105.3	46.80	13.07	22.37	95.00	6.00	107.40-18
19	** -9	5.00	109.5	33.40	15.48	8.33	94.83	10.92	113.05-19
20	** -10	5.00	106.8	39.40	14.25	14.61	94.94	7.16	118.67-20
21	** -11	5.00	108.3	46.40	13.13	22.01	94.88	5.12	124.30-21
22	** -12	5.00	103.0	25.60	17.68	0.03	95.10	8.48	129.92-22
23	** -1	5.00	104.8	45.00	13.33	20.47	95.02	7.34	135.64-23
24	** -2	5.00	110.1	33.00	15.57	7.92	94.80	7.93	141.27-24
25	** -3	5.00	109.8	38.00	14.51	13.19	94.82	7.40	146.89-25
26	** -4	5.00	110.2	28.80	16.67	3.49	94.80	7.94	152.50-26

RADIO ACTIVE CONTAMINATION SURVEY


WIPES LSC COUNTED DIRECT

ECOSCIANT COCKTAIL

BAM
1/7/04

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B2-209 DATE 1/9/04
 SURVEYOR BRENT M^C CRACKEN
 CONDITIONS PC00420, BMO#1, BMO#2
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
27	<100	MONITOR	37	<100	ARM AREA OF BMO#2
28	<100	CASE	38	<100	REAR OUTLETS BMO#2
29	<100	KEYBOARD			
30	<100	MOUSE			
31	<100	OUTSIDE OF BMO#1			
32	<100	TIPS OF BMO#1			
33	<100	ARM AREA OF BMO#1			
34	<100	REAR OUTLETS BMO#1			
35	<100	OUTSIDE OF BMO#2			
36	<100	TIPS OF BMO#2			

COMMENTS GMSM SURVEY OKAY. BKM 1/9/04
<100 NCPM

ID# C-14 CHECK

9 JAN 2004 10:0

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC1

PRESET TIME : 5.00
 DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : ST
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : OF
 TWO PHASE : NO ADC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: none

ISOTOPE 1: 14C %ERROR: 0.50 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low: 0.000 High: 331.13

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	** -1	5.00	78.8	23.20	18.57	24.18	95.95	2.21	5.49
		Blank Average		DPM	for 14C :	24.18	24.18	COEF. OF VAR:	0.000
1	** -3	5.00	102.2	49.00	12.78	27.33	95.13	30.62	11.20-27
2	** -4	5.00	102.9	62.40	11.32	41.44	95.10	13.22	16.84-28
3	** -5	5.00	101.3	44.20	13.45	22.27	95.16	21.73	22.50-29
4	** -6	5.00	96.6	68.00	10.85	47.14	95.34	42.08	28.24-30
5	** -7	5.00	104.0	67.60	10.88	46.94	95.05	28.65	33.94-31
6	** -8	5.00	98.6	34.00	15.34	11.51	95.27	17.88	39.58-32
7	** -9	5.00	99.8	30.00	16.33	7.33	95.22	18.53	45.23-33
8	** -10	5.00	108.4	53.20	12.26	31.90	94.87	30.52	50.92-34
9	** -11	5.00	100.4	24.20	18.18	1.24	95.20	16.71	56.55-35
10	** -12	5.00	105.2	33.40	15.48	10.98	95.01	8.66	62.17-36
11	** -1	5.00	101.7	51.60	12.45	30.05	95.15	6.16	67.89-37
12	** -2	5.00	101.0	38.60	14.40	16.38	95.17	21.27	73.56-38

RADIOACTIVE CONTAMINATION SURVEY


WIPES LSC COUNTED DIRECT

ECOSCINT COCKTAIL

BAM
1/9/04

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B2-209 B2-209-B DATE 1/9/04
 SURVEYOR BRENT McCracken
 CONDITIONS EXHAUST FAN 209-1 AND 209-2
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC
2-209-B=209-1 209-2=2-209-A



NO.	DPM	LOCATION	NO.	DPM	LOCATION
39	4100	GRID 2-209-B	49	4100	BACK RIGHT 2-209-B
40	4100	BACK/LEFT/LOWER PANEL 2-209-B	50	4100	TOP LEFT 2-209-B
41	4100	BACK/RIGHT/LOWER PANEL 2-209-B	51	4100	TOP RIGHT 2-209-B
42	4100	BACK/LEFT/MIDDLE PANEL 2-209-B	52	4100	INSIDE LEFT EXHAUST PIPE 2-209-B
43	4100	BACK/RIGHT/MIDDLE PANEL 2-209-B	53	4100	INSIDE RIGHT EXHAUST PIPE 2-209-B
44	4100	BACK/LEFT/UPPER PANEL 2-209-B	54	4100	LEFT COUNTER 2-209-B
45	4100	BACK/RIGHT/UPPER PANEL 2-209-B	55	4100	RIGHT COUNTER 2-209-B
46	4100	LEFT SIDE 2-209-B			
47	4100	RIGHT SIDE 2-209-B			
48	4100	BACK LEFT 2-209-B			

COMMENTS GMISM SURVEY OKAY. BAM 1/9/04
4100 NCPM

HOOD WAS DISASSEMBLED FOR CLEANING AND WIPE TEST.

ID: C-14 CHECK

9 JAN 2004 15:5

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC1
 PRESET TIME : 5.00
 DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : ST
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : OF
 TWO PHASE : NO AGC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: none

ISOTOPE 1: 14C %ERROR: 0.50 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low: 0.000 High: 331.13

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	** -1	5.00	79.4	19.60	20.20	20.43	95.93	4.32	5.50
		Blank	Average	DPM	for 14C :	20.43	COEF. OF VAR:	0.000	
1	** -3	5.00	115.1	87.40	9.57	71.98	94.58	22.85	11.23-39
2	** -4	5.00	109.9	111.60	8.47	97.28	94.81	12.51	16.90-40
3	** -5	5.00	102.8	72.20	10.53	55.49	95.10	19.70	22.57-41
4	** -6	5.00	113.4	66.80	10.94	50.14	94.66	24.19	28.26-42
5	** -7	5.00	121.2	71.40	10.59	55.29	94.30	28.08	33.97-43
6	** -8	5.00	109.9	47.40	12.99	29.57	94.81	19.83	39.63-44
7	** -9	5.00	107.2	47.00	13.05	29.08	94.93	30.89	45.31-45
8	** -10	5.00	106.7	34.60	15.21	16.01	94.94	17.88	50.97-46
9	** -11	5.00	106.7	27.60	17.03	8.64	94.95	14.54	56.60-47
10	** -12	5.00	121.3	36.80	14.74	18.60	94.29	11.12	62.24-48
11	** -1	5.00	109.1	53.20	12.26	35.66	94.84	18.32	67.98-49
12	** -2	5.00	111.3	38.60	14.40	20.31	94.75	11.04	73.64-50
13	** -3	5.00	112.9	36.60	14.78	18.23	94.68	13.46	79.27-51
14	** -4	5.00	244.5	61.00	11.45	60.30	75.56	9.97	84.91-52
15	** -5	5.00	152.3	34.40	15.25	16.76	92.50	12.67	90.55-53
16	** -6	5.00	119.9	36.60	14.78	18.36	94.36	12.31	96.19-54
17	** -7	5.00	106.6	32.40	15.71	13.69	94.95	16.48	101.84-55

RADIOACTIVE CONTAMINATION SURVEY


WIPES LSC COUNTED DIRECT

ECOSCINT COCKTAIL

BAM
1/9/04

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B2-209 DATE 1/13/04
 SURVEYOR BRENT MCLACKEN
 CONDITIONS BALANCE #11, TLC SCANNER #1 & #2, PC898, GAS CYLINDERS
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
56	2100	FRONT CONTROLS	66	2100	FRONT CONTROLS SCANNER
57	2100	PAN AREA	67	2100	INSIDE PLATE CHANGER SCANNER #2
58	2100	GLASS TOP	68	2100	INSIDE UNIT HEAD SCANNER #2
59	2100	TRAY	69	2100	TOP OF SCANNER #1
60	2100	KEYBOARD	70	2100	FRONT CONTROLS SCANNER #1
61	2100	MOUSE	71	2100	INSIDE PLATE CHANGER SCANNER #1
62	2100	MONITOR	72	2100	INSIDE UNIT HEAD SCANNER #1
63	2100	COMPUTER	73	2100	ULTRA P-10 TANK
64	2100	PRINTER	74	2100	OXYGEN TANK
65	2100	TOP OF SCANNER #2	—	—	—

COMMENTS GMSM SURVEY OKAY BMM 1/13/04
2100 NCPM.

ID: C-14 CHECK

13 JAN 2004 10:2

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC1
 PRESET TIME : 5.00
 DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : ST
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : OF
 TWO PHASE : NO AGC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: none

ISOTOPE 1: 14C %ERROR: 0.50 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low: 0.000 High: 331.13

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	** -1	5.00	79.5	26.40	17.41	27.52	95.93	2.43	5.48
		Blank	Average	DPM	for 14C :	27.52	COEF. OF	VAR:	0.000
1	** -3	5.00	108.1	31.00	16.06	5.15	94.89	6.72	11.12-56
2	** -4	5.00	97.6	30.00	16.33	3.96	95.30	6.93	16.71-59
3	** -5	5.00	107.1	30.20	16.28	4.29	94.93	15.26	22.33-58
4	** -6	5.00	103.2	29.20	16.55	3.19	95.09	7.57	27.95-59
5	** -7	5.00	101.7	48.00	12.91	22.93	95.15	13.93	33.60-60
6	** -8	5.00	98.1	41.80	13.83	16.35	95.29	9.87	39.23-61
7	** -9	5.00	103.4	52.20	12.38	27.38	95.08	12.22	44.89-62
8	** -10	5.00	103.9	42.80	13.67	17.51	95.06	9.10	50.52-63
9	** -11	5.00	106.0	41.40	13.90	16.07	94.97	11.21	56.16-64
10	** -12	5.00	98.8	36.40	14.82	10.69	95.26	6.74	61.77-65
11	** -1	5.00	105.5	29.40	16.50	3.43	94.99	7.19	67.47-66
12	** -2	5.00	101.2	32.40	15.71	6.53	95.17	9.98	73.11-67
13	** -3	5.00	107.5	34.60	15.21	8.94	94.91	5.65	78.73-68
14	** -4	5.00	110.0	29.20	16.55	3.28	94.80	7.31	84.34-69
15	** -5	5.00	98.1	31.00	16.06	5.01	95.29	5.60	89.96-70
16	** -6	5.00	110.7	30.40	16.22	4.56	94.78	8.32	95.59-71
17	** -7	5.00	99.4	26.00	17.54	-0.22	95.24	7.31	101.20-72
18	** -8	5.00	123.3	27.00	17.21	1.15	94.19	7.96	106.82-73
19	** -9	5.00	126.2	28.60	16.72	2.89	94.05	5.79	112.44-74

RADIOACTIVE CONTAMINATION SURVEY

WIPES LSC COUNTED DIRECT

ECOSCIINT COCKTAIL

BAM
1/13/04



RADIOACTIVE CONTAMINATION SURVEY

LOCATION B2-209 DATE 1/13/04

SURVEYOR BRENT MCCrackEN

CONDITIONS REF./FREEZER #24, MULTI-MAGNESTIR

INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
75	2100	TOP SHELF REF.	85	2100	LEFT INSIDE FREEZER
76	2100	MIDDLE SHELF REF.	86	2100	RIGHT INSIDE FREEZER
77	2100	BOTTOM SHELF REF.	87	2100	BACK INSIDE FREEZER
78	2100	FLOOR IN REF.	88	2100	OUTSIDE HANDLE FREEZER
79	2100	LEFT INSIDE REF.	89	2100	OUTSIDE TOP OF UNIT
80	2100	RIGHT INSIDE REF.	90	2100	TOP OF MAGNESTIR
81	2100	SHELF ON DOOR REF.	91	2100	FRONT PANEL MAGNESTIR
82	2100	OUTSIDE HANDLE REF.	92	2100	SIDE PANEL MAGNESTIR
83	2100	TOP SHELF FREEZER	—	—	—
84	2100	BOTTOM SHELF FREEZER	—	—	—

COMMENTS GMSM SURVEY OKAY 6AM 1/13/04
2100 NCPM,

ID: C-14 CHECK

13 JAN 2004 12:11

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC1
 PRESET TIME : 5.00
 DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : ST
 COUNT BLANK : YES IC# : NO REPLICATES : 1 R9232 : OF
 TWO PHASE : NO AGC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: none

ISOTOPE 1: 14C %ERROR: 0.50 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low: 0.000 High: 331.13

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	** -1	5.00	80.3	23.60	18.41	24.61	95.91	2.30	5.49
		Blank	Average	DPM	for 14C :	24.61	COEF. OF VAR:	0.000	
1	** -3	5.00	100.5	28.20	16.84	5.02	95.19	11.34	11.13-75
2	** -4	5.00	105.4	31.40	15.96	8.45	95.00	9.35	16.74-76
3	** -5	5.00	107.1	30.20	16.28	7.21	94.93	6.23	22.35-77
4	** -6	5.00	101.5	21.80	19.16	-1.70	95.16	7.13	27.96-78
5	** -7	5.00	97.2	27.40	17.09	4.14	95.32	5.44	33.56-79
6	** -8	5.00	102.9	33.00	15.57	10.09	95.10	7.84	39.18-80
7	** -9	5.00	96.6	22.40	18.90	-1.11	95.34	6.06	44.79-81
8	** -10	5.00	101.4	31.40	15.96	8.39	95.16	5.26	50.40-82
9	** -11	5.00	101.4	27.60	17.03	4.40	95.16	5.10	56.01-83
10	** -12	5.00	105.2	39.40	14.25	16.86	95.01	4.21	61.60-84
11	** -1	5.00	97.1	35.00	15.12	12.11	95.32	4.39	67.29-85
12	** -2	5.00	104.1	20.00	20.00	-3.57	95.05	5.57	72.91-86
13	** -3	5.00	113.3	23.60	18.41	0.32	94.66	5.87	78.52-87
14	** -4	5.00	102.6	20.60	19.71	-2.95	95.11	5.41	84.12-88
15	** -5	5.00	101.0	26.80	17.28	3.55	95.17	7.33	89.75-89
16	** -6	5.00	95.9	25.00	17.89	1.61	95.37	6.09	95.36-90
17	** -7	5.00	99.9	24.20	18.18	0.81	95.22	4.55	100.97-91
18	** -8	5.00	103.2	22.20	18.98	-1.26	95.09	4.17	106.59-92

RADIOACTIVE CONTAMINATION SURVEY


WIPES LSC COUNTED DIRECT

ECOSCINT COCKTAIL

BAM
1/13/04

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B2-209 DATE 1/16/04
 SURVEYOR BRENT McCracken
 CONDITIONS SHELVES & SOXHALET EXTRACTOR, PORTA-TRACE #1, MINERAL TRAY Lamp #2
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
93	2100	OUTSIDE LAMP	3	4100	SHELF B3
94	4100	TOP OF PORTA-TRACE	4	4100	SHELF B4
95	4100	TOP OF RECIRCULATOR	5	4100	TOP OF CABINET
96	4100	RECIRCULATOR CONTROLS	6	2100	SHELF C1
97	4100	HEATER #3 BASE	7	4100	SHELF C2
98	4100	TIPS OF GLASS TUBES	8	4100	SHELF C3
99	4100	HEATER #2 BASE	9	4100	SHELF C4
100	4100	TIPS OF GLASS TUBES	10	4100	SHELF C5
1	4100	SHELF B1	11	4100	SHELF C6
2	4100	SHELF B2	12	4100	SHELF C7

COMMENTS GMSM SURVEY OKAY BAM 1/16/04
4100 NCPM.

ID: C-14 CHECK

16 JAN 2004 15:

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC2
 PRESET TIME : 5.00
 DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : S
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : OI
 TWO PHASE : NO ADC : NO CYCLE REPEATS : 1 DISK : OI
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0 RUM LIST : OI
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: 4 JUN 1998 00:00

ISOTOPE 1: 14C ZERROR: 0.00 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low:-10.44 High:302.29

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	ZERROR				
31	**-1	5.00	76.6	23.40	18.49	25.04	93.52	3.22	5.50
		Blank Average		DFM for	14C :	25.04	COEF. OF VAR:	0.000	
1	**-3	5.00	95.1	37.20	14.66	15.04	92.88	27.82	11.18-93
2	**-4	5.00	95.2	50.40	12.60	29.26	92.88	18.48	16.84-94
3	**-5	5.00	101.5	41.60	13.87	19.91	92.62	20.48	22.50-95
4	**-6	5.00	94.9	37.80	14.55	15.69	92.89	15.99	28.14-96
5	**-7	5.00	99.8	30.60	16.17	8.00	92.69	13.33	33.78-97
6	**-8	5.00	95.5	27.40	17.09	4.49	92.86	14.97	39.41-98
7	**-9	5.00	99.3	32.20	15.76	9.72	92.71	12.33	45.04-99
8	**-10	5.00	93.9	31.00	16.06	8.35	92.93	19.06	50.70-100
9	**-11	5.00	102.1	37.40	14.63	15.38	92.59	22.91	56.35-1
10	**-12	5.00	95.6	31.60	15.91	9.01	92.86	15.25	62.00-2
11	**-1	5.00	93.5	42.60	13.70	20.83	92.94	8.28	67.72-3
12	**-2	5.00	95.6	42.00	13.80	20.22	92.86	8.61	73.35-4
13	**-3	5.00	95.2	45.40	13.27	23.88	92.88	13.87	79.00-5
14	**-4	5.00	99.8	26.80	17.28	3.90	92.69	13.63	84.64-6
15	**-5	5.00	93.6	25.60	17.68	2.53	92.94	8.25	90.26-7
16	**-6	5.00	97.8	27.80	16.96	4.95	92.77	11.12	95.90-8
17	**-7	5.00	95.1	22.60	18.81	-0.69	92.88	9.11	101.52-9
18	**-8	5.00	101.0	28.20	16.84	5.42	92.64	13.52	107.15-10
19	**-9	5.00	97.4	29.60	16.44	6.88	92.79	9.10	112.78-11
20	**-10	5.00	99.3	26.60	17.34	3.67	92.71	7.73	118.40-12

RADIOACTIVE CONTAMINATION SURVEY

WIPES LSC COUNTED DIRECT


ECOSCIINT COCKTAIL

BAM
1/16/04

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B2-209 DATE 1/16/04 1/19/04
 SURVEYOR BRENT McCracken
 CONDITIONS STABILITY CHAMBER #4
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC

WD
BAM
1/19



NO.	DPM	LOCATION	NO.	DPM	LOCATION
13	2100	SHELF #1	23	2100	BACK INSIDE
14	2100	SHELF #2			
15	4100	SHELF #3			
16	2100	SHELF #4			
17	2100	FLOOR			
18	2100	CEILING			
19	2100	LEFT INSIDE			
20	2100	RIGHT INSIDE			
21	2100	DOOR INSIDE			
22	2100	DOOR HANDLE			

BAM
1/19/04

COMMENTS GMSM SURVEY OKAY BAM 1/19/04
2100 NORM.

ID: C-14 CHECK

19 JAN 2004 11:4

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC2
 PRESET TIME : 5.00
 DATA CALC : BL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : ST
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : OF
 TWO PHASE : NO AGC : NO CYCLE REPEATS : 1 DISK : OF
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0 RWM LIST : OF
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: 4 JUN 1998 00:00

ISOTOPE 1: 14C %ERROR: 0.00 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low:-10.44 High:302.29

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	**-1	5.00	75.4	20.40	19.80	21.82	93.56	2.34	5.48
		Blank Average		DPM for	14C :	21.82	COEF. OF VAR:	0.000	
1	**-3	5.00	97.9	36.40	14.82	17.45	92.77	25.48	11.15
2	**-4	5.00	98.1	33.80	15.38	14.64	92.76	16.49	16.79
3	**-5	5.00	94.2	25.00	17.89	5.11	92.91	10.25	22.39
4	**-6	5.00	95.0	28.80	16.67	9.21	92.88	12.06	28.02
5	**-7	5.00	115.7	21.60	19.25	1.69	91.95	12.80	33.64
6	**-8	5.00	110.5	25.80	17.61	6.18	92.21	10.72	39.28
7	**-9	5.00	116.3	23.80	18.33	4.09	91.92	20.96	44.91
8	**-10	5.00	124.1	29.80	16.38	10.77	91.50	14.13	50.56
9	**-11	5.00	102.4	27.00	17.21	7.37	92.58	5.51	56.16
10	**-12	5.00	98.4	22.80	18.73	2.78	92.75	5.28	61.76
11	**-1	5.00	146.7	31.20	16.01	12.86	90.03	17.85	67.50

RADIOACTIVE CONTAMINATION SURVEY


WIPES LSC COUNTED DIRECT

ECOSCINT COCKTAIL

*BMM
1/19/04*

RADIOACTIVE CONTAMINATION SURVEY

LOCATION BZ-209 DATE 1/20/04
 SURVEYOR BRENT McCracken
 CONDITIONS HPLC #3
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC




NO.	DPM	LOCATION	NO.	DPM	LOCATION
24	<100	KEYBOARD	34	<100	PUMP A
25	<100	MOUSE	35	<100	PUMP B
26	<100	MONITOR	36	<100	COLUMN OVEN
27	<100	CASE	37	<100	INSIDE COLUMN OVEN
28	<100	SOLUTION BOTTLE	38	<100	B-RAM DETECTOR
29	<100	SYRINGE PUMP			
30	<100	UV-VIS DETECTOR			
31	<100	SYSTEM CONTROLLER			
32	<100	AUTO INJECTOR			
33	<100	INSIDE AUTO INJECTOR			

COMMENTS GM/SM SURVEY OKAY BAM 1/20/04
<100 NCPM

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B2-209 DATE 1/20/04
 SURVEYOR BRENT MCCrackEN
 CONDITIONS HPLC #7
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC





NO.	DPM	LOCATION	NO.	DPM	LOCATION
39	2100	KEYBOARD	49	2100	INSIDE COLUMN OVEN
40	2100	MOUSE	50	2100	Pump A
41	2100	MONITOR	51	2100	Pump B
42	2100	CASE	52	2100	UV-VIS DETECTOR
43	2100	SOLUTION BOTTLE	53	2100	β -RAM DETECTOR
44	2100	SYRINGE PUMP			
45	2100	SYSTEM CONTROLLER			
46	2100	AUTO INJECTOR			
47	2100	INSIDE AUTO INJECTOR			
48	2100	COLUMN OVEN			

COMMENTS GMSM SURVEY OKAY BAM 1/20/04
2:00 PM

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B2-209 DATE 1/21/04
 SURVEYOR BRENT McCracken
 CONDITIONS REMOVING DESK & WALL
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
54	2100	FLOOR	64	2100	FLOOR
55	2100	FLOOR	65	2100	FLOOR OPPOSITE SIDE
56	2100	FLOOR	66	2100	WALL
57	2100	INSIDE CABINET	67	2100	WALL
58	2100	INSIDE DRAWER	68	2100	WALL OPPOSITE SIDE
59	2100	INSIDE DRAWER	69	2100	WALL OPPOSITE SIDE
60	2100	INSIDE DRAWER			
61	2100	INSIDE DRAWER			
62	2100	TOP LEFT			
63	2100	TOP RIGHT			

COMMENTS GMSM SURVEY OKAY BMM 1/21/04
2100 NCPM

ID: C-14 CHECK 21 JAN 2004 16:00
 USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC2
 PRESET TIME : 5.00
 DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : 57
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : OF
 TWO PHASE : NO ADC : NO CYCLE REPEATS : 1 DISK : OF
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0 RWM LIST : OF
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: 4 JUN 1998 00:00

ISOTOPE 1: 14C %ERROR: 0.00 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low:-10.44 High:302.29

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	** -1	5.00	74.9	26.20	17.47	28.02	93.57	1.59	5.49
Blank Average				DFM for	14C :	28.02	COEF. OF VAR:		0.000
1	** -3	5.00	113.8	30.60	16.17	5.25	92.05	19.94	11.16-54
2	** -4	5.00	118.8	40.40	14.07	16.03	91.79	27.23	16.81-55
3	** -5	5.00	120.6	38.00	14.51	13.45	91.69	17.10	22.46-56
4	** -6	5.00	100.5	26.20	17.47	0.28	92.66	15.44	28.09-57
5	** -7	5.00	91.8	28.00	16.90	2.11	93.01	10.05	33.71-58
6	** -8	5.00	101.1	23.80	18.33	-2.31	92.64	10.00	39.33-59
7	** -9	5.00	105.6	27.60	17.03	1.86	92.44	9.69	44.94-60
8	** -10	5.00	99.6	33.80	15.38	8.47	92.70	6.85	50.57-61
9	** -11	5.00	104.1	28.60	16.72	2.92	92.50	11.68	56.19-62
10	** -12	5.00	109.0	23.00	18.65	-3.08	92.28	11.59	61.81-63
11	** -1	5.00	113.8	29.60	16.44	4.16	92.05	13.65	67.56-64
12	** -2	5.00	120.0	30.00	16.33	4.71	91.73	10.07	73.18-65
13	** -3	5.00	94.7	23.20	18.57	-3.03	92.90	9.82	78.81-66
14	** -4	5.00	97.3	25.20	17.82	-0.84	92.79	8.36	84.44-67
15	** -5	5.00	101.2	27.60	17.03	1.80	92.63	7.87	90.05-68
16	** -6	5.00	94.9	25.80	17.61	-0.22	92.89	8.77	95.69-69

RADIO ACTIVE CONTAMINATION SURVEY


WIPES LSC COUNTED DIRECT

ECOSCINT COCKTAIL

*BAM
1/21/04*

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B2-209 DATE 1/22/04
 SURVEYOR BRENT McCracken
 CONDITIONS BENCHTOP AND DRAWERS
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
70	2100	BENCHTOP	80	2100	INSIDE DRAWER
71	2100	BENCHTOP	81	2100	INSIDE SHELF
72	2100	INSIDE SHELF	82	2100	INSIDE DRAWER
73	2100	FLOOR	83	2100	INSIDE DRAWER
74	2100	FLOOR	84	2100	INSIDE DRAWER
75	2100	BENCHTOP	85	2100	INSIDE DRAWER
76	2100	BENCHTOP	86	2100	INSIDE SHELF
77	2100	INSIDE DRAWER	87	2100	FLOOR
78	2100	INSIDE DRAWER	88	2100	FLOOR
79	2100	INSIDE DRAWER	89	2100	FLOOR

COMMENTS GMSM SURVEY OKAY FIRM 1/22/04
2100 NCPM.

ID# C-14 CHECK

22 JAN 2004 10:

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC2
 PRESET TIME : 5.00
 DATA CALC : BL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : S
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : 0
 TWO PHASE : NO ABC : NO CYCLE REPEATS : 1 DISK : 0
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0 RNM LIST : 0
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: 4 JUN 1998 00:00

ISOTOPE 1: 14C %ERROR: 0.00 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low:-10.44 High:302.29

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	** -1	5.00	75.5	26.20	17.47	28.02	93.56	1.26	5.49
Blank Average				DPM for	14C :	28.02	COEF. OF	VAR:	0.000
1	** -3	5.00	109.5	36.00	14.91	11.02	92.26	18.87	11.14 -70
2	** -4	5.00	105.5	33.60	15.43	8.35	92.44	15.56	16.78 -71
3	** -5	5.00	110.0	42.00	13.80	17.55	92.23	12.90	22.41 -72
4	** -6	5.00	108.4	31.60	15.91	6.23	92.31	8.22	28.04 -73
5	** -7	5.00	105.2	27.40	17.09	1.63	92.46	7.48	33.65 -74
6	** -8	5.00	106.9	25.00	17.89	-0.94	92.38	8.31	39.27 -75
7	** -9	5.00	107.2	35.40	15.03	10.33	92.36	9.36	44.89 -76
8	** -10	5.00	105.6	29.60	16.44	4.02	92.44	11.81	50.50 -77
9	** -11	5.00	95.5	23.60	18.41	-2.59	92.87	8.86	56.12 -78
10	** -12	5.00	99.8	27.60	17.03	1.77	92.69	10.20	61.74 -79
11	** -1	5.00	96.7	21.60	19.25	-4.74	92.82	6.99	67.45 -80
12	** -2	5.00	108.3	26.60	17.34	0.81	92.31	6.10	73.06 -81
13	** -3	5.00	100.9	22.60	18.81	-3.61	92.64	3.95	78.67 -82
14	** -4	5.00	100.0	23.80	18.33	-2.33	92.68	3.77	84.27 -83
15	** -5	5.00	99.0	22.00	19.07	-4.28	92.72	3.89	89.89 -84
16	** -6	5.00	104.9	24.20	18.18	-1.84	92.47	4.09	95.49 -85
17	** -7	5.00	117.3	37.60	14.59	12.93	91.87	4.55	101.10 -86
18	** -8	5.00	110.2	23.80	18.33	-2.20	92.23	4.88	106.71 -87
19	** -9	5.00	115.8	29.40	16.50	3.97	91.95	6.23	112.34 -88
20	** -10	5.00	114.4	24.00	18.26	-1.92	92.02	7.43	117.95 -89

RADIO ACTIVE CONTAMINATION SURVEY


WIPES LSC COUNTED DIRECT

ECSOINT COCKTAIL

BMM
1/22/04

RADIOACTIVE CONTAMINATION SURVEY

LOCATION 2-209 E. Hood 209-Fluor DATE 22 Jan 04
 SURVEYOR Granlund/McCracken Manot
 CONDITIONS Close-out survey. Hood before disassembly
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
56	4100	2 pron	66	4100	side
57	4100	under apron	67	<100	"
58	4100	base left	68	4100	"
59	4100	" right	69	4100	top
60	4100	rear intake	70	4100	light
61	4100	back-	71	1010	top
62	4100	"	72	112	top
63	4100	"	73	4100	light
64	4100	"			6 base
65	4100	side			

COMMENTS ^{*} #72 lost in hood 2/18/04 22 Jan 03 GSM background all locations.
 (DRE) BAM 1/22/04

ID: C-14 CHECK

22 JAN 2004 20:00

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC2
 PRESET TIME : 5.00
 DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : S
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : OI
 TWO PHASE : NO ACC : NO CYCLE REPEATS : 1 DISK : OI
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0 RWM LIST : OI
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: 4 JUN 1998 00:00

ISOTOPE 1: 14C %ERROR: 0.00 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low:-10.44 High:302.29

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	** -1	5.00	77.1	18.40	20.85	19.69	93.51	2.36	5.50
		Blank Average		DPM for	14C :	19.69	COEF. OF VAR:	0.000	
1	** -3	5.00	98.8	27.00	17.21	9.45	92.73	3.56	11.10-56
2	** -4	5.00	110.3	39.00	14.32	22.63	92.22	3.31	16.69-57
3	** -5	5.00	97.6	29.40	16.50	12.02	92.78	2.63	22.28-58
4	** -6	5.00	105.1	51.60	12.45	36.15	92.46	2.38	27.87-59
5	** -7	5.00	102.5	51.40	12.48	35.87	92.57	2.09	33.46-60
6	** -8	5.00	100.2	36.60	14.78	19.83	92.67	23.51	39.12-61
7	** -9	5.00	102.3	25.20	17.82	7.55	92.58	6.61	44.73-62
8	** -10	5.00	101.2	39.80	14.18	23.31	92.63	22.78	50.39-63
9	** -11	5.00	97.0	34.40	15.25	17.40	92.80	2.82	55.99-64
10	** -12	5.00	102.2	32.20	15.76	15.11	92.59	20.51	61.64-65
11	** -1	5.00	101.9	42.80	13.67	26.56	92.60	30.73	67.42-66
12	** -2	5.00	105.2	100.00	8.94	88.54	92.46	63.17	73.25-67
13	** -3	5.00	98.9	29.40	16.50	12.04	92.73	11.14	78.88-68
14	** -4	5.00	101.6	54.20	12.15	38.87	92.61	44.81	84.61-69
15	** -5	5.00	112.1	57.20	11.83	42.43	92.13	23.83	90.30-70
16	** -6	5.00	97.6	112.60	8.43	101.76	92.78	65.84	96.13-71
17	** -7	5.00	115.0	54.80	12.08	37.92	91.99	29.62	101.84-73

RADIO ACTIVE CONTAMINATION SURVEY


WIPES LSC COUNTER DIRECT

ECOSCIANT COCKTAIL

BAM
1/22/04

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B2-209 DATE 1/22/04
 SURVEYOR BRENT McCracken / ROGER GRANLUND
 CONDITIONS HOOD 2-209-A EXHAUST DUCT
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
74	313* 301	RIGHT INNER SIDE EXHAUST	 / 		
75	228* 234	RIGHT OUTER SIDE EXHAUST			
76	255* 258	RIGHT INNER TOP SIDE EXHAUST			
77	<100	LEFT INNER SIDE EXHAUST			
78	<100	LEFT *INNER OUTER SIDE EXHAUST			
79	<100	LEFT INNER TOP SIDE EXHAUST			

COMMENTS * (RE) BANN 1/22/04 GMSM @ #74 200 cpm w/ pancake GM.

ID: C-14 CHECK 22 JAN 2004 16:17
 USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC2
 PRESET TIME : 5.00
 DATA CALC : BL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : S
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : OF
 TWO PHASE : NO AGC : NO CYCLE REPEATS : 1 DISK : OF
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0 RWM LIST : OF
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: 4 JUN 1998 00:00

ISOTOPE 1: 14C %ERROR: 0.00 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low:-10.44 High:302.29

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CFM	%ERROR				
R1	** -1	5.00	76.5	27.00	17.21	28.89	93.53	2.41	5.51
		Blank Average		DPM for	14C :	28.89	COEF. OF VAR:		0.000
1	** -3	5.00	171.7	301.20	5.15	313.88	87.93	1.14	11.14-74
2	** -4	5.00	127.8	234.60	5.84	228.29	91.28	1.33	16.76-75
3	** -5	5.00	132.6	258.20	5.57	255.06	90.99	1.05	22.35-76
4	** -6	5.00	129.3	59.80	11.57	36.73	91.20	3.28	27.96-77
5	** -7	5.00	137.8	59.60	11.59	36.90	90.65	3.39	33.56-78
6	** -8	5.00	138.2	88.40	9.51	68.72	90.63	2.46	39.18-79

RADIOACTIVE CONTAMINATION SURVEY


WIPES LSC COUNTED DIRECT

ECOSCINT COCKTAIL

AMM
1/22/04

RADIOACTIVE CONTAMINATION SURVEY

LOCATION Bldg 2 Hood 2-209A DATE 12 Jan 04
 SURVEYOR Granlund/McCracken
 CONDITIONS Inside of hood w/ back + top panels removed - after cleaning
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
80	4100	apron	90	2100	back, top
81	4100	base under apron	91	4100	" "
82	4100	base left	92	4100	" "
83	4100	" center	93	4100	top - light
84	4100	" right	94	4100	top around duct
85	154	back	95	4100	" " "
86	4100	"	96	4100	top, light
87	475	"	97	4100	side left
88	149	"	98	4100	" "
89	4100	back-top	99, 1	4100	side right

COMMENTS Before cleaning pancake GM counts @ 87, 88, 91, 92
200-400 cpm. After cleaning GMSM @ background all areas.

RADIOACTIVE CONTAMINATION SURVEY

LOCATION 2-209-A hood DATE 22 Jan 04
 SURVEYOR Gramlund/McCracken
 CONDITIONS Panels (top + back) from hood after cleaning
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC

NO.	DPM	LOCATION	NO.	DPM	LOCATION
2	4100	Back panel - right - front	12	4100	top - left - front
3	4100	" " " rear	13	4100	" " - rear
4	4100	Back Panel - left - front			
5	4100	" " " - rear			
6	4100	Back-bottom - right - front			
7	4100	" " " - rear			
8	4100	back-bottom-left - front			
9	4100	" " " - rear			
10	4100	top - right - front			
11	4100	" " - rear			

*pieces
to fix bot*

COMMENTS Before cleaning back of panels showed 200-400 cpm w/panels ke probe in spots.

ID: C-14 CHECK

22 JAN 2004 17:

USER: 1

COMMENT: 14C PERFORMANCE CHECK LSC2

PRESET TIME : 5.00
 DATA CALC : 5L DPM H# : YES SAMPLE REPEATS : 1 PRINTER : S
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : OI
 TWO PHASE : NO AGC : NO CYCLE REPEATS : 1 DISK : OI
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0 RWM LIST : OI
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: 4 JUN 1998 00:00

ISOTOPE 1: 14C %ERROR: 0.00 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low:-10.44 High:302.29

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	** -1	5.00	76.4	21.60	19.25	23.11	93.53	2.92	5.50
Blank Average				DPM	for	14C :	23.11	COEF. OF VAR:	0.000

1	** -3	5.00	97.5	36.80	14.74	16.58	92.78	10.56	11.14-80
2	** -4	5.00	110.7	40.20	14.11	20.52	92.20	5.10	16.73-81
3	** -5	5.00	96.0	30.20	16.28	9.44	92.84	4.25	22.33-82
4	** -6	5.00	93.6	27.80	16.96	6.82	92.94	3.96	27.92-83
5	** -7	5.00	96.6	29.80	16.38	9.02	92.82	4.78	33.52-84
6	** -8	5.00	111.7	163.40	7.00	154.32	92.15	52.07	39.36-85
7	** -9	5.00	101.0	89.20	9.47	73.24	92.64	18.06	45.06-86
8	** -10	5.00	103.9	460.80	4.17	475.31	92.51	2.32	50.72-87
9	** -11	5.00	108.5	159.60	7.08	149.91	92.31	12.21	56.41-88
10	** -12	5.00	96.5	39.20	14.29	19.15	92.82	8.13	62.03-89
11	** -1	5.00	100.7	59.80	11.57	41.48	92.65	19.41	67.81-90
12	** -2	5.00	96.4	33.60	15.43	13.11	92.83	4.09	73.41-91
13	** -3	5.00	92.6	30.00	16.33	9.18	92.98	3.32	79.02-92
14	** -4	5.00	97.6	33.80	15.38	13.34	92.78	9.74	84.65-93
15	** -5	5.00	97.4	56.20	11.93	37.50	92.79	8.43	90.29-94
16	** -6	5.00	100.1	112.60	8.43	98.47	92.68	5.30	95.94-95
17	** -7	5.00	100.0	62.20	11.34	44.05	92.68	10.18	101.58-96
18	** -8	5.00	100.3	39.20	14.29	19.22	92.67	17.94	107.24-97
19	** -9	5.00	101.9	30.60	16.17	9.96	92.60	18.04	112.90-98
20	** -10	5.00	101.9	68.40	10.81	50.81	92.60	43.69	118.64-99
21	** -11	5.00	106.4	61.40	11.41	43.39	92.40	14.78	124.30-1

MISSING SAMPLE

23	** -1	5.00	99.1	50.40	12.60	31.29	92.72	32.87	130.12-2
24	** -2	5.00	92.0	32.00	15.81	11.32	93.00	3.44	135.72-3
25	** -3	5.00	96.2	40.40	14.07	20.44	92.84	30.81	141.42-4
26	** -4	5.00	95.2	48.60	12.83	29.25	92.87	24.83	147.10-5
27	** -5	5.00	94.2	27.80	16.96	6.83	92.92	3.72	152.71-6
28	** -6	5.00	93.9	24.40	18.11	3.17	92.93	4.28	158.32-7
29	** -7	5.00	95.5	26.00	17.54	4.91	92.87	4.10	163.93-8
30	** -8	5.00	96.9	28.00	16.90	7.08	92.81	8.58	169.56-9
31	** -9	5.00	97.0	33.00	15.57	12.47	92.81	7.72	175.17-10
32	** -10	5.00	96.6	36.80	14.74	16.56	92.82	2.72	180.77-11
33	** -11	5.00	94.6	31.00	16.06	10.28	92.90	5.17	186.40-12
34	** -12	5.00	100.7	31.00	16.06	10.37	92.65	3.52	192.01-13

RADIO ACTIVE CONTAMINATION SURVEY


WAPES LSC COUNTED DIRECT

FRONT COCKTAIL

BAM
1/22/04

RADIOACTIVE CONTAMINATION SURVEY

LOCATION Radiochem lab DATE 23 Jan 04
 SURVEYOR R. Granlund
 CONDITIONS check of carboy & trays for release
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC




NO.	DPM	LOCATION	NO.	DPM	LOCATION
14	< 100	Top of carboy B after cleaning	Aug 27 Jan 04		
15	< 100	evap. pan Teflon			
16	< 100	" " poly			
17	< 100	" " steel			

COMMENTS GMSM check all < 50 cpm above bkg.

RADIOACTIVE CONTAMINATION SURVEY

LOCATION Hood 2-209-A DATE 23 Jan 09
 SURVEYOR R. Granlund
 CONDITIONS Hood 2-209-A after further cleaning back panel
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
18	< 100	back panel	Blank	27 Jan 09	
19	< 100	" "			
20	< 100	" "			
21	< 100	" "			

COMMENTS QMSM check of back panel. all < 50 cpm above bkg.

ID: C-14 CHECK

23 JAN 2004 18:

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC2
 PRESET TIME : 5.00
 DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : 5
 COUNT BLANK : YES ID# : NO REPLICATES : 1 R9232 : 0
 TWO PHASE : NO AQC : NO CYCLE REPEATS : 2 DISK : 0
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0 RWM LIST : 0
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: 4 JUN 1998 00:00

ISOTOPE 1: 14C %ERROR: 0.00 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low:-10.44 High:302.29

SAM NO	POS	TIME MIN	H#	14C		14C DFM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	** -1	5.00	75.2	20.20	19.90	21.60	93.57	1.52	5.49
		Blank Average		DFM for	14C :	21.60	COEF. OF VAR:	0.000	
14	1 ** -3	5.00	93.6	41.80	13.83	23.40	92.94	35.91	11.19
15	2 ** -4	5.00	103.9	38.00	14.51	19.50	92.51	14.42	16.81
16	3 ** -5	5.00	93.7	31.20	16.01	11.99	92.93	11.56	22.45
17	4 ** -6	5.00	98.0	33.20	15.52	14.21	92.76	10.62	28.07
18	5 ** -7	5.00	93.6	32.80	15.62	13.71	92.94	17.17	33.72
19	6 ** -8	5.00	94.8	27.80	16.96	8.34	92.89	10.95	39.33
20	7 ** -9	5.00	95.0	27.60	17.03	8.11	92.96	8.11	44.95
21	8 ** -10	5.00	95.5	25.20	17.82	5.55	92.87	6.39	50.55

Cycle 2 of 2

B1	** -1	5.00	75.5	19.60	20.20	20.96	93.56	1.40	56.49
		Blank Average		DFM for	14C :	20.96	COEF. OF VAR:	0.000	
1	** -3	5.00	93.6	33.60	15.43	15.21	92.94	14.62	62.05
2	** -4	5.00	103.3	37.00	14.70	19.05	92.54	7.93	67.75
3	** -5	5.00	94.5	34.90	17.36	5.75	92.90	5.51	73.37
4	** -6	5.00	96.1	37.80	15.35	15.50	92.75	6.56	77.55
5	** -7	5.00	94.1	37.90	15.84	8.08	92.97	11.77	80.57
6	** -8	5.00	94.0	34.40	16.00	5.54	92.89	8.06	90.24
7	** -9	5.00	93.2	27.00	17.15	8.32	92.95	5.51	95.85
8	** -10	5.00	95.0	24.30	18.19	5.12	92.85	4.30	101.44

RADIO ACTIVE CONTAMINATION SURVEY


WIPES LSC COUNTED CORRECT

ECOSCINT COCKTAIL

BAM
1/29/04

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B2-209 DATE 1/27/04
 SURVEYOR BRENT McCracken / ROGER GRANLUND
 CONDITIONS Hood 2-209-A / 2-209-B
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



0/0	0/0		0/0	0/0
1	2		3	4


NO.	DPM	LOCATION	NO.	DPM	LOCATION
90	4100	RIGHT OUTER SIDE EXHAUST			
91	4100	RIGHT INNER SIDE EXHAUST			
92	4100	RIGHT INNER TOP SIDE EXHAUST			
93	4100	LEFT OUTER SIDE EXHAUST			
94	4100	LEFT INNER SIDE EXHAUST			
95	4100	LEFT INNER TOP SIDE EXHAUST			
1	4100	2-209-B INSIDE CABINET			
2	4100	2-209-B INSIDE CABINET			
3	4100	2-209-A INSIDE CABINET			
4	4100	2-209-A INSIDE CABINET			

BAMA
1/28/04

COMMENTS GMSM SURVEY OKAY BAMA 1/28/04
< 100 NCPM

RADIOACTIVE CONTAMINATION SURVEY

LOCATION 2-209 DATE 29 Jan 04
 SURVEYOR R. Granlund / G. McCracken
 CONDITIONS Survey of sink traps.
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC
 Window Sink by hood Hood Sink Sink by stairwell wall



NO.	DPM	LOCATION	NO.	DPM	LOCATION
96	4100	Window sink basin	1	4100	2ml from trap - window sink
97	364	drain inside	2	4100	" " " " " "
98	4100	trap inside	3	126	" " " " hood sink
99	4100	drain inside	4	120	" " " " " "
100	4100	hood sink basin	5	4100	" " " " stairwell sink
22	4100	STAIRWELL WALL SINK BASIN	6	4100	" " " " " "
23	4100	DRAIN INSIDE			
24	4100	TRAP INSIDE			
25	4100	DRAIN INSIDE			
—	—	—			

COMMENTS Smears from sink basin and inside traps. Water sample from traps. Hood sink trap inaccessible, not disassembled but water sample and basin smear obtained.

ICAL SURVEY/RWG

Hood sink trap emptied to red waste (100 ml) and trap & drain surveyed 29 Jan 04.

REV. 14 Mar 2003

GCW
29 Jan 04

ID: C-14 CHECK 27 JAN 2004 16:00
 USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC1
 PRESET TIME : 5.00
 DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : ST
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : OF
 TWO PHASE : NO AGC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: none

ISOTOPE 1: 14C %ERROR: 0.50 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low: 0.000 High: 331.13

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CFM	%ERROR				
B1	** -1	5.00	81.2	22.80	18.73	23.78	95.88	1.03	5.47
		Blank	Average	DPM	for 14C :	23.78	COEF. OF VAR:		0.000
1	** -3	5.00	128.3	111.40	8.47	94.81	93.93	6.85	11.13-90
2	** -4	5.00	116.1	59.20	11.62	38.84	94.53	9.23	16.76-91
3	** -5	5.00	124.5	88.40	9.51	70.13	94.13	4.42	22.38-92
4	** -6	5.00	121.5	32.60	15.62	11.01	94.28	9.93	28.01-93
5	** -7	5.00	120.1	36.60	14.78	15.01	94.35	8.09	33.62-94
6	** -8	5.00	123.4	49.20	12.75	28.46	94.19	6.28	39.24-95
MISSING SAMPLE									
11	** -1	5.00	105.8	53.00	12.29	32.02	94.98	18.95	45.07-96
12	** -2	5.00	158.4	357.40	4.73	364.44	92.06	0.82	50.68-97
13	** -3	5.00	149.9	86.00	9.64	69.04	92.65	5.10	56.31-98
14	** -4	5.00	141.5	89.60	9.45	72.36	93.19	4.28	61.92-99
15	** -5	5.00	106.3	39.20	14.29	17.50	94.96	12.10	67.57-100
16	** -6	5.00	116.9	28.00	16.90	5.85	94.50	9.79	73.19-72
17	** -7	5.00	210.3	38.00	14.51	22.37	82.34	8.52	78.82-23
18	** -8	5.00	261.4	27.00	17.21	13.76	71.92	7.79	84.43-24
19	** -9	5.00	221.7	32.00	15.81	16.13	80.18	11.82	90.06-25

RADIOACTIVE CONTAMINATION SURVEY

W/YES LSC COUNTED DIRECT

ECOSCIINT COCKTAIL

BAM
1/28/04

ID: C-14 CHECK

27 JAN 2004 17:3

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC1
 PRESET TIME : 5.00
 DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : ST
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : OF
 TWO PHASE : NO ADC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: none

ISOTOPE 1: 14C %ERROR: 0.50 FACTOR: 1.0000000 SKB. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: Off

Quench Limits Low: 0.000 High: 351.13

SAM NO	POS	TIME MIN	H#	14C		14C	14C	LOSER %	D REJECT TIME
				CPM	%ERROR	CPM	EFF-1		
B1	**1	3.00	117.0	24.55	17.97	24.55	94.48	2.47	5.49
B2	**2	3.00	117.8	24.80	17.96	26.26	94.45	2.01	11.08
Blank Average				DPM for 14C :		25.41	COEF. OF VAR: 4.736		
1	**4	5.00	106.9	28.60	16.72	4.72	94.94	1.98	16.68
2	**5	5.00	119.0	28.20	16.84	4.47	94.40	1.75	22.26
3	**6	5.00	135.5	142.00	7.51	126.40	93.54	0.86	27.87
4	**7	5.00	134.3	137.00	7.64	120.94	93.61	0.98	33.47
5	**8	5.00	118.0	19.60	20.20	-4.65	94.45	2.61	39.06
6	**9	5.00	117.4	21.20	19.43	-2.97	94.47	2.61	44.66

WENRO'S SINK
 HOD SINK
 STAIR WEL SINK

RADIO ACTIVE CONTAMINATION SURVEY

2x 2ml ALIQ. pip. 480

ECOSCINT COCKTAIL

*BAM
1/28/04*

ID: C-14 CHECK

28 JAN 2004 09:1

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC1
 PRESET TIME : 5.00
 DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : ST
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : DF
 TWO PHASE : NO AGC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: none

ISOTOPE 1: 14C %ERROR: 0.50 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low: 0.000 High: 331.13

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	** -1	5.00	B1.1	22.20	18.98	23.15	95.88	1.29	5.49
Blank Average				DPM for	14C :	23.15	COEF. OF VAR:		0.000
1	** -3	5.00	113.8	27.40	17.09	5.80	94.64	7.18	11.10-1
2	** -4	5.00	124.0	39.20	14.29	18.48	94.16	6.78	16.72-2
3	** -5	5.00	119.8	27.20	17.15	5.67	94.36	7.01	22.31-3
4	** -6	5.00	103.4	25.00	17.89	3.14	95.08	6.26	27.93-4

RADIOACTIVE CONTAMINATION SURVEY


WIPES LSC COUNTED DIRECT

ECSOFCINT COCKTAIL

BAM
 1/28/04

RADIOACTIVE CONTAMINATION SURVEY

LOCATION Radiochem Lab. Building 2 DATE 1/25/04
 SURVEYOR BRENT McCracken
 CONDITIONS B2-209
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
5	2100	WALL	15	2100	WALL
6	2100	WALL	16	2100	WALL
7	2100	WINDOW	17	2100	DESKTOP
8	2100	WINDOW	18	2100	DESKTOP
9	2100	WALL	19	2100	DESKTOP
10	2100	WALL	20	2100	DESKTOP
11	2100	WINDOW	21	2100	DESKTOP
12	2100	WALL	22	2100	DESKTOP
13	2100	WALL	23	2100	DESKTOP
14	2100	WALL	24	2100	DESKTOP

COMMENTS GM/SM SURVEY OKAY BMM 1/28/04
<2100 NCPM

* (cc) BMM
1/28/04

RADIOACTIVE CONTAMINATION SURVEY


LOCATION Radiochem Lab, Building 2

DATE 1/29/04

SURVEYOR BRENT McCracken

CONDITIONS B2-209

INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
25	2100	DESKTOP	35	2100	FLOOR
26	2100	DESKTOP	36	2100	FLOOR
27	2100	DESKTOP	37	2100	FLOOR
28	2100	DESKTOP	38	2100	FLOOR
29	2100	DESKTOP	39	2100	FLOOR
30	2100	DESKTOP	40	2100	FLOOR
31	2100	FLOOR	41	2100	FLOOR
32	2100	FLOOR	42	2100	FLOOR
33	2100	FLOOR	—	—	—
34	2100	FLOOR	—	—	—

COMMENTS GMSM SURVEY OKAY BMM 1/29/04
2100 NCPM.

ID: C-14 CHECK

28 JAN 2004 09:5

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC1
 PRESET TIME : 5.00
 DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : ST
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : OF
 TWO PHASE : NO ABC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: none

ISOTOPE 1: 14C %ERROR: 0.50 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On


Quench Limits Low: 0.000 High: 331.13

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	**1	5.00 Blank	80.7 Average	23.00 DPM	18.65 for 14C	23.98	95.89 23.98	0.94 COEF. OF	5.48 VAR: 0.000
1	**3	5.00	95.5	44.80	13.36	22.98	95.38	27.37	11.16-5
2	**4	5.00	98.8	30.40	16.22	7.93	95.26	7.89	16.78-6
3	**5	5.00	107.0	31.00	16.06	8.67	94.93	6.86	22.38-7
4	**6	5.00	109.0	31.20	16.01	8.91	94.85	7.10	28.00-8
5	**7	5.00	96.1	25.60	17.68	2.86	95.36	6.14	33.59-9
6	**8	5.00	99.6	23.40	18.49	0.59	95.23	6.38	39.19-10
7	**9	5.00	115.7	38.00	14.51	16.20	94.55	4.97	44.81-11
8	**10	5.00	103.0	29.40	16.50	6.93	95.10	4.76	50.41-12
9	**11	5.00	112.1	21.60	19.25	-1.18	94.72	5.89	56.01-13
10	**12	5.00	97.3	26.80	17.28	4.13	95.31	5.83	61.63-14
11	**1	5.00	104.4	23.00	18.65	0.22	95.04	4.95	67.32-15
12	**2	5.00	101.9	20.60	19.71	-2.33	95.14	5.15	72.92-16
13	**3	5.00	107.9	31.00	16.06	8.68	94.89	5.93	78.54-17
14	**4	5.00	120.4	27.60	17.03	5.27	94.33	5.87	84.16-18
15	**5	5.00	106.8	62.60	11.30	41.95	94.94	3.55	89.76-19
16	**6	5.00	114.8	28.00	16.90	5.62	94.59	5.21	95.38-20
17	**7	5.00	122.5	34.20	15.29	12.31	94.23	6.44	101.00-21
18	**8	5.00	117.0	24.80	17.96	2.26	94.49	10.28	106.62-22
19	**9	5.00	122.2	24.00	18.26	1.48	94.25	6.04	112.24-23
20	**10	5.00	113.6	28.40	16.78	6.02	94.65	5.10	117.86-24
21	**11	5.00	122.5	24.60	18.03	2.12	94.23	8.63	123.49-25
22	**12	5.00	152.3	55.60	12.00	36.13	92.49	7.75	129.13-26
23	**1	5.00	121.4	71.60	10.57	51.95	94.29	2.83	134.83-27
24	**2	5.00	111.1	31.00	16.06	8.73	94.76	5.31	140.44-28
25	**3	5.00	118.6	32.40	15.71	10.33	94.42	8.48	146.07-29
26	**4	5.00	110.1	33.40	15.48	11.25	94.80	5.41	151.68-30
27	**5	5.00	111.5	26.20	17.47	3.67	94.74	4.91	157.29-31
28	**6	5.00	120.7	29.00	16.61	6.76	94.32	11.43	162.93-32
29	**7	5.00	110.9	29.00	16.61	6.62	94.77	6.55	168.55-33
30	**8	5.00	106.2	27.00	17.21	4.45	94.97	5.86	174.15-34
31	**9	5.00	120.7	27.60	17.03	5.28	94.32	7.23	179.78-35
32	**10	5.00	115.1	36.60	14.78	14.71	94.58	6.88	185.40-36
33	**11	5.00	117.4	29.20	16.55	6.92	94.48	6.79	191.02-37
34	**12	5.00	120.5	24.60	18.03	2.09	94.33	6.20	196.64-38
35	**1	5.00	111.6	31.20	16.01	8.95	94.74	5.55	202.33-39
36	**2	5.00	111.8	21.80	19.16	-0.97	94.73	5.46	207.94-40
37	**3	5.00	122.8	21.60	19.25	-1.06	94.22	5.28	213.55-41
38	**4	5.00	117.6	26.20	17.47	3.75	94.46	7.65	219.16-42

RADIO ACTIVE CONTAMINATION SURVEY WIPES LSC COUNTED DIRECT ECOSYSTEM COCKETS

RADIOACTIVE CONTAMINATION SURVEY

LOCATION Hood sink 2-209-A DATE 29 Jan 04
 SURVEYOR R.W. Granlund
 CONDITIONS check of hood sink after emptying trap.
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
26	< 100	basin & top of drain	Rec'd 29 Jan 04		
27	< 100	drain outside			
28	< 100	" inside			
29	< 100	trap inside			
			Rec'd 29 Jan 04		

COMMENTS Towel wipes check w/ GM54 &
Trap contents (~100 ml) had assay of 120 dpm/2ml (see survey
of 27 Jan 04. Trap emptied and drained trap rechecked.

ID: C-14 CHECK

29 JAN 2004 15:4

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC1
 PRESET TIME : 5.00
 DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : ST
 COUNT BLANK : YES IC# : NO REPLICATES : 1 R5232 : OF
 TWO PHASE : NO ADC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: none

ISOTOPE 1: 14C %ERROR: 0.50 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On




Quench Limits Low: 0.000 High: 331.13

SAM NO	POS	TIME MIN	H#	<u>14C</u>		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	** -1	5.00	B1.2	24.20	18.18	25.24	95.88	2.86	5.50
			Blank Average	DPM for	14C :	25.24	COEF. OF VAR:	0.000	
1	** -3	5.00	102.6	37.00	14.70	13.66	95.11	12.08	11.15
2	** -4	5.00	100.6	22.40	18.90	-1.71	95.19	9.88	16.75
3	** -5	5.00	100.9	23.80	18.33	-0.23	95.18	6.28	22.36
4	** -6	5.00	100.9	24.00	18.26	-0.03	95.18	7.82	27.96

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B2-209 TO B1-101 DATE 2/20/04
 SURVEYOR BRENT MCCRACKEN
 CONDITIONS NINE REF/FREEZER #13 FROM B2-209 TO B1-101
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC

ALL STANDARDS AND SAMPLES REMOVED FROM UNIT AND PLACED INTO A BOX,
 BOX TRANSFERRED FROM B2-209 TO B1-101 AND STANDARDS AND SAMPLES PLACED
 BACK INTO UNIT.



NO.	DPM	LOCATION	NO.	DPM	LOCATION
45	2100	FREEZER SHELF	52 *	2100	DESKTOP
46	4100	FREEZER BOTTOM	53 *	2100	DESKTOP
47	2100	REF. SHELF			
48	2100	REF. SHELF			
49	2100	FREEZER INSIDE DOOR			
50	2100	REF. INSIDE DOOR			
51	2100	INSIDE BOX			

COMMENTS GMSM SURVEY OKAY BAM 2/20/04

* NO. 52/53 WERE RUN ON 2/23/04, BAM 2/23/04

ID: C-14 CHECK

20 FEB 2004 16:0

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC1
 PRESET TIME : 5.00
 DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : ST
 COUNT BLANK : YES IC# : NO REPLICATES : 1 RS232 : OF
 TWO PHASE : NO AGC : NO CYCLE REPEATS : 1
 SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0
 LOW LEVEL : NO HALF LIFE CORRECTION DATE: none

ISOTOPE 1: 14C %ERROR: 0.50 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low: 0.000 High: 331.13

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	** -1	5.00	78.2	37.80	14.55	39.39	95.97	0.82	5.48
Blank Average				DPM for	14C :	39.39	COEF. OF VAR:	0.000	
1	** -3	5.00	102.3	43.80	13.51	6.66	95.12	10.11	11.12-45
2	** -4	5.00	104.0	40.80	14.00	3.54	95.06	6.97	16.72-46
3	** -5	5.00	97.7	37.20	14.66	-0.35	95.30	7.92	22.32-49
4	** -6	5.00	103.9	25.60	17.68	-12.46	95.06	9.62	27.94-48
5	** -7	5.00	100.8	26.40	17.41	-11.65	95.18	7.87	33.54-49
6	** -8	5.00	104.7	26.00	17.54	-12.03	95.03	8.50	39.17-50
7	** -9	5.00	105.8	34.20	15.29	-3.38	94.98	10.32	44.79-51

RADIOACTIVE CONTAMINATION SURVEY

REP./FREEZER #13 MOVE FROM B2-209 TO B1-101

WIPES DIRECT LSC COUNTED

ECOSCINT COCKTAIL

*BMM
2/20/04*

ID# C-14 CHECK

23 FEB 2004 17:2

USER: 1 COMMENT: 14C PERFORMANCE CHECK LSC1

PRESET TIME : 5.00

DATA CALC : SL DPM H# : YES SAMPLE REPEATS: 1 PRINTER : BT

COUNT BLANK : YES IC# : NO REPLICATES : 1 R9232 : OF

TWO PHASE : NO AGC : NO CYCLE REPEATS : 1

SCINTILLATOR: LIQUID LUMEX: NO LOW SAMPLE REJ: 0

LOW LEVEL : NO HALF LIFE CORRECTION DATE: none

ISOTOPE 1: 14C %ERROR: 0.50 FACTOR: 1.000000 BKG. SUB: 0

BACKGROUND QUENCH CURVE: Off COLOR QUENCH CORRECTION: On

Quench Limits Low: 0.000 High: 331.13

SAM NO	POS	TIME MIN	H#	14C		14C DPM	14C EFF-1	LUMEX %	ELAPSED TIME
				CPM	%ERROR				
B1	**1	5.00	85.4	37.40	14.63	39.07	95.74	4.31	5.51
		Blank Average		DPM for	14C :	39.07	COEF. OF VAR:		0.000
1	**3	5.00	102.0	46.60	13.10	9.92	95.14	19.08	11.18 -52
2	**4	5.00	101.1	46.20	13.16	9.48	95.17	12.30	16.81 -53

RADIOACTIVE CONTAMINATION SURVEY

REF. FREEZER #13 MOVE FROM B2-209 TO B1-101


WIPES PERFORM LSC COUNTED

ECOSCINT COCKTAIL

BAM
2/23/04

RADIOACTIVE CONTAMINATION SURVEY

LOCATION Radiochem Lab, Building 2 DATE 2/25/04
 SURVEYOR BRENT McCracken
 CONDITIONS CEILING TILE
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	N-cpm DPM	LOCATION	NO.	DPM	LOCATION
1	<100	Tile #1			
2	<100	Tile #2			
3	<100	Tile #3			
BMM 2/25/04			BMM 2/25/04		
/					

COMMENTS BMSM SURVEY OKAY, BMM 2/25/04
<100 N-cpm

Attachment C

**Closeout Survey
Environmental Fate Laboratory
Building 3**

EXYGEN RESEARCH INC.
Closeout Survey
Building 3

Summary

Building 3 of Exygen research was used for the incubation and storage of samples containing ^{14}C . It is planned that the building will be sold and ownership transferred on or about 31 Aug 03. No contamination was detected in the routine surveys conducted while ^{14}C was used in the environmental chambers. All licensed material was removed from the building and a closeout survey was performed to determine whether there was any ^{14}C contamination left in the building. The interior of the complete building was included in the survey, with emphasis on the locations of the environmental chambers where ^{14}C was used and the freezer where ^{14}C samples were stored. No contamination was detected and the building can be released without any restrictions on use because of radioactive material.

Introduction

Exygen Research (formerly Centre Analytical Laboratories) was issued byproduct material license number 37-30095-01 on 28 Feb 94. The license allowed the use of ^3H , ^{14}C , ^{32}P and ^{35}S in the radioisotope laboratory located in the building (Building 1) at 3048 Research Drive, State College, PA. Amendment number 2, which was issued on 10 Oct 96 added Building 2 at 3058 Research Drive to the license and allowed the radioisotope laboratory to be moved to the second floor of Building 2. Amendment number 5 was issued on 12 Mar 02 and added Building 3 at 3117 Research Drive to the license. The use of radioactive material in building 3 was limited to 1 mCi of ^{14}C in samples maintained in environmental chambers. The large walk-in freezer in Building 3 was also used for the storage of samples that had been processed and were being held until the client(s) accepted the study report gave permission to dispose of the samples. The environmental chambers and the work areas around them were subject to the same requirements for surveys as the radiochemistry laboratory when radioactive material was in use. Building 3 also housed the receiving department for Exygen, including the receipt of radioisotope shipments. Radioisotope shipments were delivered to the radioisotope laboratory for opening. The receiving department was moved back to Building 1 in November 2002.

It is planned that Building 3 will be sold by Exygen Research on or about 31 Aug 03. This survey was made to document that all licensed radioactive material has been removed from the building and that there is no residual radioactive contamination.

Disposition of Licensed Material

The use of the environmental chambers in building 3 was terminated in November 2002. The chambers and samples were transferred to the radioisotope laboratory in Building 2, where they are still in use. The remaining licensed material in Building 3, the samples that were stored in the walk-in freezer, was transferred to a freezer in Building 1 on about 15 Jul 03.

Radiation Survey

The only licensed material used in Building 3 was ^{14}C , which does not produce an external radiation hazard. No specific measurements for external dose rates were made as part of this survey. However, the survey for surface contamination with a pancake GM probe indicated no radiation levels above the normal background level.

Contamination Survey

A survey for fixed and removable radioactive contamination was conducted in Building 3 on 17 Jul 03. All areas in the building were checked with a Geiger Mueller Survey Meter with a 2-inch pancake probe. Paper smears, assayed by liquid scintillation counting were used to check for removable contamination. Spot checks were made in all the rooms. The locations of the environmental chambers and the storage area for samples in the walk-in freezer were checked in more detail. There was no indication of contamination, either fixed or removable with the GMSM survey or the smears.

The surveys included the accessible portions of floor drains. The environmental chambers have drains for condensation that were routed to the floor drain with flexible plastic tubing. The environmental chambers were operated above room temperature and no condensation from volatile ^{14}C compounds was expected. The samples in the chambers were also connected to a gas collection system which continuously purged the gas from the samples and trapped $^{14}\text{CO}_2$ to measure the breakdown of the materials under test. No contamination was found in or near the floor drain in this survey or in the routine surveys when the environmental chambers were in use.

The walk-in freezer used for storage of samples in Building 3 is a closed unit with no ventilation. Air is blown over two sets of cooling coils inside the chamber by fans. The coils have a defrost cycle to remove frost. The catch pans for the defrost water drain to the outside of the building through 3/4" copper/plastic drains. Smears were taken from the inside of the exit ends of the drain pipes outside the building (the only accessible location without cutting open the pipes) to check for any contamination from condensation of volatile ^{14}C compounds. No ^{14}C was detectable in either of the drain pipe smears. One smear had an activity of about 70 dpm in the ^{14}C channel of the Liquid Scintillation Counter, but an 800-minute beta spectrum of the sample also indicated energies higher than ^{14}C and the spectrum was not distinguishable from background. The activity is attributed to natural activity in the buildup of dirt inside the pipe.


Instrumentation

The portable survey meter used in the survey was a Bicron Surveyor M, S/N B426K with a Bicron PGM probe S/N B107N calibrated on 27 Feb 03. The efficiency of the pancake probe on contact with a surface is 0.13 c/beta, as determined with a 45-mm plated ^{99}Tc source. Background is typically 25-30 cpm. The liquid scintillation counter used to count smears was a Beckman LS6500. This instrument has an internal quench correction program and is checked daily with a factory ^{14}C standard and is calibrated annually under the Exygen QA program with the same standard. Smears were counted on 17 Jul 03 for 5 minutes each with 2 vials with blank smears used for background. The sample set was recounted on 18 Jul 03 for 30 minutes each

(MDA estimate is 4.3 dpm). The activity on all paper smears was well below 100 dpm limit for removable contamination. The highest value was 11 dpm. The cloth smears were also counted using the Beckman LS6500.


Attachments

The contamination survey forms are attached.

Prepared by:  Date: 8 Aug 03
Rodger W. Granlund
Radiation Safety Officer

RADIOACTIVE CONTAMINATION SURVEY

LOCATION Exygen Bldg 3 (room outside freezer) DATE 17 Jul 03
 SURVEYOR R. W. Gerzylund
 CONDITIONS Closeout survey, no RAM in building
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M x LSC Ludlum-3




NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	< 10	table	54	< 10	floor drain - inside
2	< 10	cart top shelf	 5080803 8/20/03 		
3	< 10	" bottom "			
4	< 10	" trays			
5	< 10	cart			
6	< 10	freezer door			
7-17	< 10	walls			
18	< 10	doors			
19-32	< 10	floor			
53	< 10	floor drain top			

COMMENTS GMSM Survey - no detectable contamination

RADIOACTIVE CONTAMINATION SURVEY

LOCATION Exyggen Bldg. 3 (-20°C freezer room) DATE 17 Jul 03
 SURVEYOR R. W. Granlund
 CONDITIONS Cross out survey, No RAM in building Freezer #8
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor Mx LSCx Ludlum-3

Crosshatch = only shelves used for storage of RAM samples in boxes.




NO.	DPM	LOCATION	NO.	DPM	LOCATION
33	11	shelf 1, 2	 R. W. Granlund 8 Aug 03 		
34	< 10	" 3, 4			
35	< 10	" 5 (bottom)			
36-43	< 10	Floor			
44-52	< 10	walls			
101	70	drain - inside pipe			
102	< 10	" " "			

COMMENTS G/MSP survey - no detectable contamination
#101 & #102 cloth swabs from inside end of condensate drains for
freezer coils outside the building on 18 Jul 03

RADIOACTIVE CONTAMINATION SURVEY

LOCATION Exygen Bldg 3- West Lab #105 DATE 17 Jul 03
 SURVEYOR R. W. Granlund
 CONDITIONS Closeout survey - no RAM in building
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor MK LSC Ludlum-3



NO.	DPM	LOCATION	NO.	DPM	LOCATION
55-58	< 100*	bench top	Reed 8 Aug 03		
59, 60	< 100*	sinks			
61, 62 60, 62	< 100*	desks			
63-67	< 100*	floor			
68, 69	< 100*	drains			
70-73	< 100*	walls			

COMMENTS GMSM survey - no detectable contamination
Ceramic tile floor background elevated
* < 10 incorrect entry Reed 8 Aug 03


RADIOACTIVE CONTAMINATION SURVEY

LOCATION Frygen Bldg 3 - East lab - no RAM use = 106 DATE 17 Jul 03

SURVEYOR R. W. Grantlund

CONDITIONS closeout survey, no RAM in building

INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M LSC Ludlum-3




NO.	DPM	LOCATION	NO.	DPM	LOCATION
77-82	< 10	bench tops	Rec'd 8 Aug 03	Rec'd 8 Aug 03	Rec'd 8 Aug 03
83	< 10	sink			
84-87	< 10	floor			
88	< 10	drain			
89-92	< 10	walls			

COMMENTS GMSM survey - no detectable contamination
Ceramic tile floor background elevated

RADIOACTIVE CONTAMINATION SURVEY

LOCATION Exycen Bldg 3, Entrance + hallway DATE 17 Jul 03
 SURVEYOR R.W. Geranlund
 CONDITIONS closeout survey, no RAM in building
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M LSC Ludlum-3



NO.	DPM	LOCATION	NO.	DPM	LOCATION
93	< 10	Table	Filey & Aug 03		
94	< 10	countertop/sink			
95-97	< 10	floor-hallway			
98	< 10	" restroom			
99, 100	< 10	" garage			

COMMENTS GMSM survey - no detectable contamination

Attachment D

Closeout Survey Radioisotope Laboratory Room 101, Building 1

**Closeout Survey
Radioisotope Laboratory
Room 101, Building 1
MPI Research, State College, PA**

SUMMARY

The Radioisotope Laboratory in room 1-101 has been decommissioned. All radioactive material, except the general license sources in the liquid scintillation counters and exempt quantity calibration sources for the pancake GM counters, has been disposed as radioactive waste. The survey included the work surfaces, cabinets, drawers, floor, walls, ceiling, sink drains, hood, refrigerator/freezers, and all other equipment used or stored in the room.

Work surfaces and other flat surfaces were surveyed with a Ludlum 2360 ratemeter/scaler and a Ludlum 43-89 100 cm² detector. Smaller areas and equipment surfaces, not accessible to the Ludlum 43-89 probe, were checked with a Ludlum 3 ratemeter and Ludlum 44-9 thin end-window detector. Paper smears counted with a liquid scintillation counter were used to check for removable contamination. The MDA for smears from a 100 cm² area and counted for 30 minutes was 5 dpm. The MDA for the Ludlum 43-89 detector was 2320 dpm/100 cm² or 1370 dpm/100 cm² if the scaler function was used to integrate counts during the survey. The MDA for the pancake GM detectors was 8,000 dpm/100 cm² to 16,000 dpm/100 cm². These MDA's are all well below the limit of 3.7E+6 dpm/100 cm² for unrestricted release of building surfaces.

After decommissioning, accessible building and equipment surfaces had less than 10 dpm/100 cm² of removable contamination and no detectable contamination with survey meters. These values are well below the limits for unrestricted release and room 1-101 can be returned to normal use without restrictions.

INTRODUCTION

The Radioisotope Laboratory was moved from room 209 of Building 2 to room 101 of Building 1 in February 2004. Tritiated compounds were stored in the lab and some assays of samples containing tritium were performed, but the inventory did not exceed 0.51 mCi and no work with more than a few microcuries was performed. One project in 2005 involved a total of 15 mCi of ¹⁴C and resulted in minor contamination (<1000 dpm/100 cm²) of some work surfaces and equipment. All areas were cleaned to less than 100 dpm/100 cm² at the completion of the work. All other work was with microcurie amounts and there was essentially no contamination. In the routine weekly surveys of the laboratory in 2004-2006, only about 3 smears of the floor and uncovered bench tops exceeded 100 dpm/100 cm².

RELEASE LIMITS

The guidelines published by the NRC in Regulatory Guide 1.86 in 1974 (NRC 1974) and in 1993 (NRC 1993) were used as the contamination limits for previous closeout surveys. For ¹⁴C the values were 5,000 dpm/100 cm² averaged over up to 1 m², 15,000 dpm/100 cm² maximum for up to 100 cm², and 1,000 dpm/100 cm² for removable contamination. The current limits for building surface contamination published by the NRC in NUREG 1757 (NRC 2006) are based on specific radionuclides and the value for ¹⁴C is 3.7E+6

dpm/100 cm². This value assumes 10% of the activity is removable and is based on a dose limit of 25 mrem/year for building occupants.

SURVEY METHOD

Routine daily surveys when unsealed radioactive material was handled included a direct survey of the work surface with a thin-window pancake GM detector and wipes of the work surface, also checked with the thin-window pancake GM detector. The radioisotope user also checked hands, shoes and clothing with the GM detector. Routine weekly surveys included a direct survey with a thin-window pancake GM detector and smears of the floor and work surfaces. The 4 cm diameter paper smears were counted with either a thin-window GM detector and scaler or with a liquid scintillation counter. Styrofoam smears counted in the liquid scintillation counter were also used in some surveys.

For this survey paper smears were counted using a Beckman LS 6500 liquid scintillation counter to improve the MDA (minimum detectable activity). Surfaces were also surveyed with a Ludlum 43-89 alpha-beta probe. This probe has a thin ZnS alpha scintillator backed by a plastic beta scintillator. The probe was operated in the beta-only mode with a Ludlum 2360 ratemeter. More than 90% of the area of bench tops, cabinet exteriors and interiors and drawer exteriors and interiors were surveyed with this instrument. Equipment was surveyed with a thin-window pancake GM tube and smears before release.

CALIBRATION

The Beckman LS 6500 liquid scintillation counter was calibrated with factory ¹⁴C and ³H standards. The factory-installed quench curves and the external source quench factor were used to provide the quench correction for individual samples. The detection efficiency for ¹⁴C on smears with this counter is typically about 93%.

The portable Ludlum model 3 with Ludlum 44-9 pancake GM probe and the Bicon Surveyor with Bicon PGM pancake probe were routinely calibrated for beta detection efficiency with an Eberline electroplated ⁹⁹Tc disc source. The Bicon frisk-tech scaler with the Bicon LPGM lead-shielded pancake detector used for smear counting was also routinely calibrated with the Eberline electroplated ⁹⁹Tc disc source. The detection efficiency for ⁹⁹Tc beta radiation with these counters was 11% to 14% at the last calibration in July 2007.

For this survey the detectors were calibrated with ¹⁴C sources to achieve a more accurate efficiency for ¹⁴C surface contamination. Two ¹⁴C disc sources were prepared using ¹⁴C glucose solution that had been standardized using the liquid scintillation counter. The sources were prepared by depositing the same number of 10 ul aliquots of solution on several equal-area annular rings to simulate a uniform deposit. The 4.75 diameter discs had an active diameter of 4.4 cm. The discs were made of heavy paper with the annular rings marked on the paper and topped with a transparent adhesive plastic film. The solution was deposited on the film and evaporated at low temperature to produce a simulated uniform deposit of ¹⁴C without self-absorption. The detection efficiencies of the Ludlum 3/44-9, Bicon Surveyor M/PGM and Ludlum 2360/43-89 survey meters for ¹⁴C surface contamination determined with these sources were 5.3% to 6.5%.

DETECTION LIMITS

Most smears were counted for 30 minute periods in the liquid scintillation counter. A blank smear was used for the background count, which was typically 25-30 cpm. This equates to a standard deviation (σ) of 1 cpm or 1.1 dpm. The MDA ($4.66 \cdot \sigma_{\text{bkg}}$) for smears assayed with liquid scintillation counting is 5 dpm. The Bicon frisk-tech has an efficiency of 4.5% for ^{14}C on smears and a background of about 30 cpm. Routine smears counted for 1 minute have a standard deviation of 5.5 cpm or 120 dpm and a MDA of 570 dpm. Therefore, to achieve the lowest MDA for this survey the liquid scintillation counter was used with a 30 minute count time to assay smears.

For surveys with the portable survey meters, changes in the audible count rate are used to detect contamination on surfaces. A change in count rate to 2-3 times the average background is usually detectable. The background for the Ludlum and Bicon pancake GM meters was about 60 cpm, so an increase of 60-120 cpm is assumed to be detectable. For a detector area of 15 cm² and a ^{14}C detection efficiency of 5%, this corresponds to a MDA of 1200-2400 dpm or 8,000-16,000 dpm/100 cm².

The background for the Ludlum 2360/43-89 survey meter was 270 cpm and an increase to 400 cpm was detectable. For a detector area of 100 cm² and a ^{14}C efficiency of 5.6% this corresponds to a MDA of 2320 dpm or 2320 dpm/100 cm². In addition to an analog ratemeter, the Ludlum 2360 is equipped with a digital scaler that was used to integrate the counts for 1-minute periods during the scans. This provided the average count rate for the surface scanned during the 1-minute period. The MDA is the same as for a 1-minute count with a background of 270 cpm, which is 1,370 dpm. The average area scanned with the Ludlum 43-89 probe in one minute was 4,500 cm², giving a MDA of 1,370 dpm/100 cm² averaged over 4,500 cm². Using the scaler function during the scans reduced the detection limit for the Ludlum 2360/43-89 survey meter and that survey meter was used to survey all building surfaces. The smaller pancake GM detectors were used to survey the irregular surfaces of equipment.

RESULTS

The results of the surveys are described below. The 73 pages of survey records are attached to this report.

Bench top work surfaces

The results for the surveys of the bench top work surfaces are on pages 58-60 of the surveys. Only one smear (11 dpm) exceeded 10 dpm. No contamination was detected in the scan with the Ludlum 2360/43-89 survey meter.

Sinks

Room 101 contains two sinks. The small sink on the south wall is the one routinely used. The larger sink at the end of the counter island was rarely used. Samples of the water in the traps of both sinks were collected and 1-ml aliquots were assayed on 31 Aug 07 (page 37). Both samples were less than background with a MDC of 2E-6 microcuries/ml. The sink on the south wall was used after that date for washing equipment used in the preparation of liquid waste for disposal. Therefore, samples were collected from the surface and bottom of the

trap for the south wall sink on 4 Oct 07 and 2-ml aliquots of the samples were assayed (page 60). The results were 1.1E-6 microcuries/ml and 1.6E-6 microcuries/ml respectively, with a MDC of 2E-7 microcuries/ml. This is less than 1% of the sewer release concentration for ^{14}C and indicates only minor contamination in the sink trap, which would be expected to have the highest contamination in the plumbing. Therefore, no attempt was made to disassemble and survey the interior of any plumbing.

Floor

The floor of room 1-101 was surveyed on 10 and 11 Oct 07 after all other areas had been cleaned and surveyed (page 62-64). Two smears (35 dpm and 32 dpm) of 37 exceeded 10 dpm. After cleaning, smears from this area were also less than 10 dpm. Each of the floor smears covered a single 1 ft² tile, so the results were less than 10 dpm/930 cm². In addition to smears, 11 areas of about 10 ft² each were scanned with the Ludlum 2360/43-89 survey meter. There was no detectable contamination in these scans. Smears from the hallway between the building entrance and room 101 were also less than 10 dpm.

Walls

Thirty-five smears were taken of the wall surfaces (page 52-54), with more attention to the areas on the west wall behind the radioisotope lab bench and hood, the south wall behind the sink, and the east wall where an experiment using millicuries amounts of ^{14}C was conducted. The windows and shades behind the hood exhaust were also surveyed. About 160 ft² of those areas were also surveyed with the Ludlum 2360/43-89 survey meter. All smears were less than 10 dpm and there was no detectable contamination with the survey meter scan.

Ceiling

The suspended ceiling is made up of 2'x2' acoustic tile with 12 2'x4' fluorescent light diffuser panels. Smears taken from 5 of the diffuser panels and 4 tiles were all less than 10 dpm. There was no detectable contamination on the five diffuser panels and 9 tiles that were scanned with the Ludlum 2360/43-89 survey meter. However, the background of the tiles with the survey meter was about 130 cpm greater than for the rest of the room. This was determined by gamma spectroscopy to be from ^{232}Th in the tile. The tiles that were surveyed included those directly above the exhaust port for the portable hood.

Hoods

A portable hood was located on the radioisotope bench along the west wall of the lab. The hood was equipped with an activated carbon filter and a fiber prefilter. The carbon from the filter was assayed by combusting samples in the biological oxidizer and counting with the liquid scintillation counter. Two carbon filters were used in the hood during the time radioactive material was used in room 1-101. The filter removed on 8 Apr 05 contained 0.36 microcuries and the filter removed on 17 Jul 07 contained 38.6 microcuries. The higher activity in the second filter is attributed to the opening of a vial of ^{14}C -formic acid in the hood in June 2005. The activated carbon was included in the radioactive waste shipment of 3 Oct 07. The hood was disassembled and surveyed with smears and a pancake GM detector (page 17). No contamination was detected with the GM

survey meter but, three of the 20 smears had detectable contamination (12-23 dpm). After cleaning the contaminated area all smears were less than 10 dpm. The hood was released for unrestricted use.

In May and June 2005 the hood in room 111 of Building 2 was used for the synthesis of ^{14}C compounds, because the hood room 101 of Building 1 was not large enough to accommodate the apparatus. A total of about 3.2 mCi of ^{14}C was used in three syntheses. The glassware used for the syntheses was vented through a LN_2 trap to catch any volatile ^{14}C compounds. The hood, other work areas in the room and the equipment were cleaned and surveyed after each use. All accessible areas and equipment items had <100 dpm/100 cm^2 removable contamination after cleaning. The hood was not disassembled, but contamination levels at the air exhaust slots at the back of the hood were well below the limit for unrestricted release of building surfaces. Copies of the hood surveys are included as pages 67-72 of the attached surveys.

In August-October 2006 hoods 2-305-A and 2-305-B were used for some extractions of samples with microcuries quantities of ^{14}C . No volatile ^{14}C compounds were involved and the assay of the extraction waste indicated the total activity in the samples was less than 1 microcurie. Routine weekly smear and GSM surveys of the hoods did not detect any contamination. The final survey was done on 5 Oct 06 and is included as page 73 of the attached surveys.

Refrigerators and freezers

Room 1-101 contained two combination refrigerator/freezers that were used for the storage of radioactive standards and samples. The freezers were emptied, cleaned and checked with smears and a pancake GM survey meter (page 14-16). There was no detectable contamination on the smears or with the pancake GM survey for either unit. The smears were counted for both ^{14}C and ^3H . A smear used to absorb some condensate water on the floor under refrigerator/freezer #13 indicated 18 dpm ^3H . This was not considered significant because there was no indication of any contamination inside the refrigerator/freezer. The refrigerator/freezers were no longer needed and were released for disposal as municipal waste.

Walk-in freezer #10 had been used to store some samples containing ^{14}C and routine surveys did not indicate any contamination. Smear samples taken from the floor and shelves inside the freezer and the floor outside the freezer were all less than 10 dpm (page 64). The low-temperature freezer #35 had also been used for some storage of ^{14}C samples. Frost from this freezer was thawed and a 2 ml aliquot had 28 dpm ^3H (page 64). This was suspected to be cross contamination of the sample container so another sample of frost was removed from the freezer on 18 Oct 07. Duplicate 2 ml aliquots of this sample indicated no contamination (MDA=1.3 dpm ^3H , 4.0 dpm ^{14}C) and no further action was deemed necessary.

Equipment

Room 101 contained a great deal of equipment that had been used in radioisotope experiments including 2 liquid scintillation counters, 2 biological oxidizers, 2 TLC plate scanners and 2 HPLC units with Beta-Ram detectors. A large amount


of other TLC equipment was stored in the room. All the equipment was surveyed with smears and a pancake GM survey meter. Equipment that showed no detectable contamination with the pancake GM survey and had smears of less than 10 dpm was released for unrestricted use. Items that could not be readily decontaminated were disposed as radioactive waste. The surveys of equipment are contained in pages 1-13, 19-36, 50, 51, 61, 65, and 66. All equipment except the two liquid scintillation counters, the pancake GM meters and the office equipment used in preparing this report has been removed from room 1-101. The two Beckman LS 6500 liquid scintillation counters were surveyed (page 61) and will be transferred to the MPI laboratories at Mattawan, MI. The transfer of the ^{137}Cs sources in the liquid scintillation counters will be handled by a Beckman service representative.

Prepared by: Rodger W. Granlund
Rodger W. Granlund, CHP
Radiation Safety Officer

14 Nov 07

RADIOACTIVE CONTAMINATION SURVEY

LOCATION 1-101 DATE 25 Jun 07
 SURVEYOR R. Geram/uzd
 CONDITIONS Release survey for TAC Plate Scanners
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



AMBIS 4000 TAC Plate scanner

Scanner 1

Scanner 2

NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	80	Top	1	< 10	top
2	60	Front panels	2	< 10	front panels
3	70	" "	3	< 10	front panel
4	< 10	side panel	4	< 10	side panel
5	< 10	rear panel	5	< 10	rear panel
6	150	plate loader outside	6	< 10	plate loader outside
7	70	" " inside	7	< 10	" " inside
8	< 10	Detector chamber & window	8	< 10	Detector chamber & window
9	< 10	Plate holder	9	< 10	Plate holder
10	< 10	collimator plate	10	< 10	collimator plate


11 < 10 side cover (removed) 11 < 10 side cover (removed)
 12 < 10 top & front cover (removed) 12 < 10 top & front cover (removed)

COMMENTS GM5M survey - no contamination detected
Clean Scanner #1 and resurvey (RWR 27 Jun 07)
Scanner #2 approved for release

Rely

RADIOACTIVE CONTAMINATION SURVEY

LOCATION 1-101 DATE 27 Jun 07
 SURVEYOR R. Gerznlund
 CONDITIONS Equipment release survey
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



Scanner #1 see p. 1 #1-B-7-B

NO.	DPM	LOCATION	NO.	DPM	LOCATION
1-B	< 10	Top	11-B	< 10	sides light box
2-B	< 10	front panel	12-B	< 10	bottom " "
3-B	< 10	front panel	13-B	< 10	computer AMBLS 1 top
4-B	< 10	side panel	14-B	< 10	sides
5-B	< 10	rear panel	15-B	< 10	front
6-B	< 10	plate loader-outside	16-B	< 10	rear base
7-B	< 10	" " -inside	17-B	< 10	monitor-2439-front
8-B	< 10	Bottom # 1	18-B	< 10	Top
9-B	< 10	" # 2	19-B	< 10	sides
10-B	< 10	^{top} Porta-trace light box	20-B	< 10	rear base





COMMENTS GAMMA survey - no detectable contamination
(monitor face ~100 cpm, probably K-40 in glass).

Smears 1-B through 7-B from scanner 1 after cleaning

Reed

RADIOACTIVE CONTAMINATION SURVEY

LOCATION 1-101 DATE 27 Jun 07
 SURVEYOR R. Granlund
 CONDITIONS Equipment release survey
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
21-B	< 10	keyboard 21/B-top	7-C	< 10	paper cutter blade
22-B	< 10	5. " side base	 Aug 27 Jun 07 		
23-B	< 10	mouse 23/B			
24-B	< 10	printer 171 front			
1-C	< 10	" top			
2-C	< 10	" sides			
3-C	< 10	" right base			
4-C	< 10	" paper tray			
5-C	< 10	paper cutter top			
6-C	< 10	" " base			

COMMENTS GCMST survey - no detectable contamination
All items on pages 2 and 3 are released for unrestricted
use.

Key

RADIOACTIVE CONTAMINATION SURVEY





LOCATION 1-101

DATE 28 Jun 07

SURVEYOR R. Granlund

CONDITIONS Release surveys

INSTRUMENT(S) Bicon Frisk-Tech Bicon Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
D-1	< 10	Power Box A	D-11	< 10	HPLC Pump D709 Top+front
D-2	< 10	Power Box B	D-12	< 10	" " " side+base
D-3	< 10	Chart of Elements	D-13	< 10	" " D1256 top+front
D-4	< 10	D585 basin Ultrasonic cleaner	D-14	20	" " " side+base
D-5	< 10	" front	D-15	< 10	" " D1257 top+front
D-6	< 10	" side+base	D-16	< 10	" " " side+base
D-7	< 10	D561 control hotplate	D-17	< 10	" " D2308 top+front
D-8	13	top "	D-18	< 10	" " " side+base
D-9	< 10	side + base "	D-19	< 10	workbench
D-10	< 10	radio	D-20	< 10	"

COMMENTS GMSM survey - no contamination detected.
All items meet release limits. Clean and resurvey
hot plate (D561) and HPLC pump D1256.

RADIOACTIVE CONTAMINATION SURVEY

LOCATION 1-101


DATE 29 Jun 05

GT wd
RWG 2 Jul 07

SURVEYOR R. Granlund

CONDITIONS release surveys

INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
B-1	42	hotplate D561 top	B-11	< 10	Recirc/chiller D2112 front
B-2	14	" control	B-12	< 10	top
B-3	< 10	" sides + base	B-13	< 10	sides
B-4	< 10	HPHC Pump D1256 top	B-14	< 10	rear
B-5	< 10	" " sides + base	B-15	< 10	UV lamp D1852
B-6	< 10	Extraction heater 1 D1045 front panel	B-16	< 10	Extraction heater 2 D1044 same as B-6
B-7	< 10	sides + rear	B-17	< 10	" B-7
B-8	12	heaters	B-18	< 10	" B-8
B-9	< 10	condensers	B-19	< 10	" B-9
B-10	< 10	condensers inside*	B-20	< 10	" B-10*

COMMENTS * cotton swab. Pump D1256, UV lamp, chiller, Extraction heater D1044 released. More decon required for extraction heater D1045 heater plates and hot plates.

RADIOACTIVE CONTAMINATION SURVEY

LOCATION 1-101 DATE 2 Jul 07
 SURVEYOR R. Granlund
 CONDITIONS release surveys
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M x Ludlum 3 LSC x




Extraction Heater D1045

NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	< 10	Heater # 1			
2	< 10	" # 2			
3	< 10	" # 3			
4	< 10	" # 4			
5	< 10	" # 5			
6	< 10	" # 6			

COMMENTS Check to find contaminated heater plate
(see page 5) GMSM check - no detectable contamination
Extraction heater # D1045 is released. Ref 3/6/07

RADIOACTIVE CONTAMINATION SURVEY

LOCATION 1-101 DATE 9 Jul 07
 SURVEYOR R. Granlund
 CONDITIONS Release Survey
 INSTRUMENT(S) Bicron Frisk-Tech X Bicron Surveyor M X Ludlum 3 X LSC X



NO.	DPM	LOCATION	NO.	DPM	LOCATION
B-1	< 10	Lid 1	B-11	< 10	frame 2
B-2	< 10	Lid 2	B-12	< 10	frame 3
B-3	< 10	Lid 3	B-13	< 10	4V lamp # D 2204
-4	< 10	Lid 4			
-5	< 10	Lid 5			
-6	< 10	Tank 1			
-7	< 10	" 2			
-8	< 10	" 3			
-9	< 10	" 4			
-10	< 10	Frame 1			

COMMENTS GMSM check ok -
All items ok for release Aug
12 Jul 07

Reed


RADIOACTIVE CONTAMINATION SURVEY

LOCATION 1-101 DATE 12 Jul 07

SURVEYOR R. Granlund

CONDITIONS release survey

INSTRUMENT(S) Bicron Erisk-Tech Bicron Surveyor Mx Ludlum 3 LSCx



NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	< 10	Dispenser # 1	11	< 10	Top - Pump
2	< 10	" # 2	12	< 10	bottom "
3	< 10	" # 3	13	< 10	sides + rear "
4	< 10	" # 4	14	< 10	front "
5	< 10	" # 5	15	< 10	Front - frac. collector
6	< 10	" # 6	16	< 10	top " "
7	< 10	" # 7	17	< 10	sides " "
8	< 10	AC outlet center D2096	18	< 10	base " "
9	< 10	Vacuum Pump	19	< 10	tray " "
10	< 10	Silicon hot hands	20	< 10	holder base 1

COMMENTS GMSM survey - no detectable contamination
Release above items 13 Jul 07



PAGE 10

RADIOACTIVE CONTAMINATION SURVEY

LOCATION 1-101 DATE 12 Jul 07
 SURVEYOR R. Granlund
 CONDITIONS release surveys
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC

NO.	DPM	LOCATION	NO.	DPM	LOCATION
< 10	B-21	holder #1	D-7	< 10	vial holder 6
< 10	B-22	holder base #2	D-8	< 10	" " 7
< 10	B-23	holder #2			
< 10	B-24	Letter trays			
< 10	D-1	stapler			
< 10	D-2	vial holder 1			
< 10	D-3	" " 2			
< 10	D-4	" " 3			
< 10	D-5	" " 4			
< 10	D-6	" " 5			

Free
13 Jul 07

COMMENTS GSM survey - no detectable contamination
Release above items 13 Jul 07

Free

RADIOACTIVE CONTAMINATION SURVEY

LOCATION 1-101 DATE 23 Jul 02
 SURVEYOR R. Granlund
 CONDITIONS waste processing - equipment release.
 INSTRUMENT(S) Bicron Frisk-Tech x Bicron Surveyor M Ludlum 3 LSC x

Book B.

NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	< 10	5-gal pail 1 - outside	11	< 10	Chair #2 - top
2	< 10	" " 1 inside	12	< 10	" " base
3	< 10	" " 2 outside	13	10	chair #3 - top
4	< 10	" " 2 inside	14	< 10	" " base
5	< 10	" " 3 outside	15	< 10	chair #4 - top
6	< 10	" " 3 inside	16	< 10	" " -base
7	10	Fiber drum 1 - outside	17	33	fract. coll. tray 1
8	< 10	" " 4 inside	18	12	" " " 2
9	< 10	chair #1 - top	19	19	" " " 3
10	< 10	" " base	20	< 10	" " " 4

COMMENTS #2 fraction collector rack 500 cpm fixed contamination - rack discarded. All other GM/SM survey no detectable contamination.

ICALSURVEY/RWG

Fraction collector tray # 4 released.
 Chair # 3 released
 Chair 1, 2, 4 released
 Pails + drum released.
 Clean fraction collectors 1, 3

REV. 14 Mar 2003

RWG

RADIOACTIVE CONTAMINATION SURVEY

LOCATION 1-101 DATE 23 Jul 07
 SURVEYOR R. Granlund
 CONDITIONS equipment release
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor Mx Ludlum 3 LSCx

NO.	DPM	LOCATION	NO.	DPM	LOCATION
21	< 10	Syringe housing			
22	< 10	" cover			

COMMENTS GMSM check ok.

[Signature]

RADIOACTIVE CONTAMINATION SURVEY

LOCATION 1-101 DATE 24 Jul 07
 SURVEYOR R. Granlund
 CONDITIONS equipment release
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC


Book C

NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	< 100	Chair #3 top			
2	< 100	" " base			
3	< 100	frac. collect. tray 1			
4	< 100	" " " 3			
5	< 100	Cooler #1 top handles			
6	< 100	" " sidesd bottom			
7	< 100	" " lid inside			
8	< 100	" " bottom "			

COMMENTS All smears < 10 dpm. No detectable contamination with GEMSA. Release tags attached to 4 items listed above.

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 28 Jul 07
 SURVEYOR R. Granlund
 CONDITIONS Release survey Refrigerator - Freezer #24, D1998
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor MX Ludlum 3 LSCX



book D

NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	< 10	Top	11	< 10	back
2	< 10	"	12	< 10	freezer top
3	< 10	L. side	13	< 10	" L. side
4	< 10	" "	14	< 10	" R. side
5	< 10	front - freezer	15	< 10	" rear
6	< 10	" - ref.	16	< 10	" bottom
7	< 10	" "	17	< 10	" "
8	< 10	R. side	18	< 10	" door shelf
9	< 10	" "	19	< 10	" " "
10	< 10	back	20	< 10	ref. door shelf

COMMENTS GMSM survey - no detectable contamination
Poste w/ release form

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 28 Jul 07
 SURVEYOR RW Garza/ozd
 CONDITIONS equipment release
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC


NO.	DPM	LOCATION	NO.	DPM	LOCATION
21	< 10	ref. door shelf	31	< 10	1ml ice
22	< 10	" " "	 [Diagonal line through the right half of the table] RW 30 Jul 07 		
23	< 10	" " "			
24	< 10	ref. top			
25	< 10	" L. side			
26	< 10	" R. side			
27	< 10	" back			
28	< 10	" bottom			
29	< 10	" shelf glass			
30	< 10	shelves-grates			

COMMENTS _____

RW

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 2 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS Release survey
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 ✓ LSC ✗



Refrigerator/Freezer #13 D1126


NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	< 10	Top	11	< 10	Ref. L. Freezer shelf
2	< 10	L. side	12	< 10	" "
3	< 10	R. side	13	< 10	Ref-L. side
4	< 10	Back.	14	< 10	" R. "
5	< 10	Front top	15	< 10	" top
6	< 10	" bottom	16	< 10	" rear
7	< 10	Freezer L. side	17	< 10	" top shelf
8	< 10	" rear	18	< 10	" lower "
9	< 10	" R. side	19	< 10	" drawer L
10	< 10	" top	20	< 10	" " R
			21	< 10	door shelves
			22	< 10	" "

COMMENTS Count for ¹³C & ³H

No detectable contamination w/ GM5M. Sample of condensate from floor absorbed on paper smear - 17.8 dpm³/l

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 3 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS release survey
 INSTRUMENT(S) Bicron Frisk-Tech X Bicron Surveyor M X Ludlum 3 LSCX



Portable Hood D1186

NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	< 10	top	11	< 10	R. side inside
2	< 10	L. side	12	< 10	top inside
3	< 10	R. side	13	< 10	Filter holder
4	< 10	rear	14	< 10	sides & rear. Filter compartment
5	< 10	front-top	15	< 10	bottom
6	< 10	front panel	16	< 10	top
7	< 10	front shield	17	10	base, left
8	19	front shield back	18	< 10	" , rear
9	12	L. side inside	19	< 10	" right
10	23	10. rear "	20	< 10	L. side inside *

COMMENTS *after detergent wipe No detectable contamination with GM5M. Areas with ≥ 10 dpm (8-10, 12) were all < 10 dpm after cleaning w/ detergent (20-29).

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 3 Aug 07
 SURVEYOR R. Generalized
 CONDITIONS release survey
 INSTRUMENT(S) Bicon Frisk-Tech Bicon Surveyor M Ludlum 3 LSCX

NO.	DPM	LOCATION	NO.	DPM	LOCATION
21	< 10	rear inside*	 R. Generalized 7 Aug 07 	 R. Generalized 7 Aug 07 	
22	< 10	r. side inside*			
23	< 10	bases - top*			
24	< 10	bases - bottom*			

COMMENTS _____



RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 7 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS Release surveys
 INSTRUMENT(S) Bicron Frisk-Tech X Bicron Surveyor M Ludlum 3 LSCx
 Survey of plastic storage containers inside
 Container top sides bottom Lid sides ends bottom
 1-115 liter B-1 B-2 B-3 B-4 B-5 B-6 B-7
 2-53 L B-8 B-9 B-10 B-11 B-12 B-13 B-14
 3-30 L B-15 B-16 B-17 B-18 B-19 B-20 B-21
 4-53L 001364 B-22 B-23 B-24 C-1 C-2 C-3 C-4
 5-cooler C-5 C-6 C-7 C-8 C-9 C-10 C-11
 6-10L C-12 C-13 C-14 C-15 C-16 C-17 C-18
 7-10L C-19 C-20 C-21 C-22 C-23 C-24 D-1
 8-53L D-2 D-3 D-4 D-5 D-6 D-7 D-8

NO.	DPM	LOCATION	NO.	DPM	LOCATION
B-1	11	container #1	B-11	<10	Container #2
2	<10	↓	12	<10	↓
3	<10	↓	13	<10	↓
4	<10	↓	14	<10	↓
5	13	↓	15	<10	container #3
6	<10	↓	16	<10	↓
7	19	↓	17	<10	↓
8	<10	Container #2	18	<10	↓
9	<10	↓	19	<10	↓
10	<10	↓	20	<10	↓

COMMENTS No detectable contamination w/GMSM on #1-#7.
Bottom of #8 1,000-3,000 cpm (residue from broken sample bottles),
reduced to 100-300 cpm after several cleanings.



RADIOACTIVE CONTAMINATION SURVEY

LOCATION _____ DATE 7 Aug 07
SURVEYOR _____
CONDITIONS _____
INSTRUMENT(S) Bicon Frisk-Tech Bicon Surveyor M Ludlum 3 LSC

Table with 6 columns: NO., DPM, LOCATION, NO., DPM, LOCATION. Contains handwritten data for 16 measurements (B-21 to C-7, 8-16) with DPM values < 10 and various container locations.

COMMENTS Storage containers #1-5 released 8 Aug 07

RADIOACTIVE CONTAMINATION SURVEY

LOCATION _____ DATE 7 Aug 07
 SURVEYOR _____
 CONDITIONS _____
 INSTRUMENT(S) Bicon Frisk-Tech__ Bicon Surveyor M__ Ludlum 3__ LSC__

NO.	DPM	LOCATION	NO.	DPM	LOCATION
C-17	< 10	container #6	D-3	< 10	container #8
18	< 10	" "	P-4	< 10	
19	12	container #7	D-5	< 10	
20	< 10		P-6	< 10	
21	< 10		D-7	< 10	
22	< 10		D-8	1030	
23	23				
24	< 10				
D-1	48				
D-2	< 10	container #8			RWG 9 Aug 07

COMMENTS storage container #6 released 9 Aug 07

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 8 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS release survey
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 X LSC X

HPLC Pumps	front	sides & top	rear & bottom	Door
D1822	B-15	B-16	B-17	-
D1248 (UV)	B-18	B-19	B-20	B-21
D1247 (cont)	B-22	B-23	B-24	None
D1249 (LC)	C-1	C-2	C-3	-
D710	C-4	C-5	C-6	-
D445 (LC)	C-7	C-8	C-9	-
D1002 (UV)	C-10	C-11	C-12	C-13
D1013 (cont)	C-14	C-15	C-16	-

NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	< 10	CPU D381 side	B-11	< 10	power strip 1
B-2	< 10	" " front, rear	12	< 10	" " 2
3	< 10	" " top, bottom	13	< 10	" " 3
4	< 10	Keyboard D2639	14	< 10	" " 4
5	< 10	mouse	15	< 10	HPLC D1822
6	< 10	Monitor D2606 screen	16	< 10	" "
7	< 10	" sides & rear	17	< 10	" "
8	< 10	" top & bottom	18	< 10	D1248
9	< 10	syringe D2072 sides	19	< 10	"
10	< 10	" cover	20	< 10	"

COMMENTS Release D381, D2639, mouse, D2600, D2072
power strip 1, 2, 3, 4, D1822, D1248, RUV 10 Aug 07

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 8 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS Release surveys
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3X LSCX

NO.	DPM	LOCATION	NO.	DPM	LOCATION
21 B-21	< 10	D1248	C- 7	< 10	we RUST 9 Aug 07 DE D445
B-22	< 10	D1247	8	< 10	"
B-23	< 10	"	9	< 10	"
B-24	< 10	"	10	< 10	D1007
C-1	< 10	D1249	11	< 10	"
2	< 10	"	12	< 10	"
3	< 10	"	13	< 10	"
4	< 10	D710	14	< 10	D1013
5	< 10	"	15	< 10	"
6	< 10	"	16	< 10	"

COMMENTS Released D1247, D1249, D710, D445,
D1007, D1013 Aug 10 Aug 07

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 8 Aug 07
 SURVEYOR R. Geranlund
 CONDITIONS release survey
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 X LSC X

NO.	DPM	LOCATION	NO.	DPM	LOCATION
31 C-17	< 10	D-582 Injector top + sides	D-3	< 10	oven - doors
18	< 10	" rear + bottom	4	< 10	" interior
19	< 10	" door	5	< 10	syringe pump D597
20	< 10	" inside	6	< 10	" door
21	24	" Injector	7	< 10	tray for D-582 (cleaned)
22	156	" sampler	8		PLCY 9 Aug 07
23	78	" Injector swab	9		
24	166	" sampler swab	10		
D-1	< 10	D-1250 oven top + sides	11		
D-2	< 10	" rear + bottom	12		

COMMENTS D-582 sample tube holder 300 cpm (needle removed)
D-528 injector parts cleaned 9 Aug 07 + re-checked.
Release D1250 PLCY 10 Aug 07

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 9 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS Release survey
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3x LSCx

NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	< 10	CPU D357 front	11	< 10	D438 back+bottom
2	< 10	" " side+top	12	< 10	" Door inside
3	< 10	" " rear, bottom	13	< 10	" inside
4	< 10	Monitor D2478 front	14	< 10	" tray
5	< 10	" " side+top	15	< 10	" sampler
6	< 10	" " rear+bottom	16	< 10	" injector
7	< 10	key board D2617	17	< 10	D582 Injector tray
8	< 10	mouse	18	11	" sampler after cleaning
9	< 10	Injector D438 front	19	21	" injector
10	< 10	top + sides	20	< 10	B-RAM D2021 front

COMMENTS D438 500 cpm at sampling tube (cleaned w/acid) all others < 100 cpm w/GMSM
release D357, D2478, D2617, mouse, D438 RWJ 10 Aug 07
release D582, D2021, D2022 R227 13 Aug 07



PAGE 26

RADIOACTIVE CONTAMINATION SURVEY


LOCATION BI-101 DATE 9 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS Release surveys
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC

NO.	DPM	LOCATION	NO.	DPM	LOCATION
21	< 10	B-ran detector	31	< 10	D2022 Tray 2
22	< 10	" " point	 [Diagonal line through the right half of the table] RSG 9 Aug 07 		
23	< 10	" " Top + sides			
24	< 10	" " back + bottom			
25	< 10	D2022 Fract. Collector Front			
26	< 10	Top + sides			
27	< 10	rear + bottom			
28	< 10	gantry			
29	< 10	basin			
30	< 10	Tray 1			

COMMENTS _____

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 10 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS Release Survey
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 K LSC



Biological Oxidizer OX 500

NO.	DPM	LOCATION	NO.	DPM	LOCATION
32	< 10	D2200 BMO2 front	42	< 10	D938 L. side
33	< 10	Top	43	< 10	rear
34	< 10	R. side	44	< 10	holder sample boat 3
35	< 10	L. side	45	< 10	" " 4
36	< 10	rear	46	< 10	sample port
37	10	holder 1 sample boat	47	< 10	rear access chamber
38	14	" " 2	48	< 10	D2200 sample port
39	< 10	D938 BMO1 front	49	< 10	" rear access
40	< 10	top	50	< 10	" interior pump compartment
41	< 10	R. Side	51	< 10	D938 " " "

COMMENTS All items < 100 cpm w/GM5M
Release D2200, D938 Aug 13 Aug 07

RADIOACTIVE CONTAMINATION SURVEY

LOCATION BL-101 DATE 10 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS Release surveys
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC

NO.	DPM	LOCATION	NO.	DPM	LOCATION
52	< 10	U.V.lamp - D2081	62	< 10	D944 back + bottom
53	< 10	CM-10 Box outside	63	< 10	" chamber
54	< 10	" Box inside	64	< 10	D98 Controller front
55	< 10	" View port	65	< 10	↓ top + sides
56	< 10	UV lamp D2080 cover	66	< 10	↓ rear + bottom
57	< 10	" " " Top	67	< 10	D2020 pump front
58	< 10	" " " sides + bottom	68	< 10	↓ top + sides
59	< 10	UV safety glasses	69	< 10	↓ rear + bottom
60	< 10	D944 UV chamber front	70	< 10	D1009 UV det. front
61	< 10	↓ sides + top	71	< 10	" top + sides

COMMENTS all items < 100 cpm w/ GeMSM
Release D2081, CM-10 box, D2080, UV safety glasses
UV chamber, D944, D98, D2020, D1009 Rec'd 13 Aug 07



RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 10 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS release surveys
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor Mx Ludlum 3 LSC

NO.	DPM	LOCATION	NO.	DPM	LOCATION
72	< 10	D1009 back + bottom	82	< 10	D2014 rear + bottom
73	< 10	D1008 oven front	83	< 10	" sample rack
74	< 10	↓ top + sides	84	< 10	↓ sampler
75	< 10	↓ rear + bottom	85	< 10	↓ indicator
76	< 10	↓ oven chamber	86	< 10	↓ chamber
77	< 10	D2019 cooler front	 RILEY 10 Aug 07 		
78	< 10	↓ sides + top			
79	< 10	↓ rear + bottom			
80	< 10	D1014- Injector front			
81	< 10	top + sides			

COMMENTS all items < 100 cpm w/ GM54
Release D1008, D2019, D2014 13 Aug 07

RADIOACTIVE CONTAMINATION SURVEY

LOCATION BI-101 DATE 13 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS Release Survey
 INSTRUMENT(S) Bicron Frisk-Tech X Bicron Surveyor M Ludlum 3 LSCX


NO.	DPM	LOCATION	NO.	DPM	LOCATION
87	< 10	D597 Syringe pump front	97	< 10	TLC Jar outside
88	< 10	↓ sides & top	98	< 10	" inside
89	< 10	↓ rear & bottom	99	< 10	" rack
90	< 10	↓ cover	100	< 10	shield box outside
91	< 10	Syringe front D1817 Fract. Collector	1	< 10	" " inside
92	< 10	↓ top & side	2	< 10	Film holder outside
93	< 10	↓ rear & bottom	3	< 10	" " inside
94	< 10	↓ tray	15 Aug 07		
95	< 10	↓ plastic tray			
96	< 10	TLC Jar - top			

COMMENTS No detectable contamination w/ GEMMA.
Release D597, D1817, TLC jar, shield box, film holder - 15 Aug 07

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 15 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS Release Survey
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor Mx Ludlum 3 LSCx

Ambis TLC plate scanner plate holders



NO.	DPM	LOCATION	NO.	DPM	LOCATION
4	< 10	¹⁴ C test plate top	14	< 10	3/8 (1)
5	< 10	base	15	< 10	3/8 2
6	< 10	frames (1)	16	< 10	not used RWD 17 Aug 07
7	< 10	frames (2)			
8	< 10	bases 1			
9	< 10	" 2			
10	72	1/32 inserts			
11	190	1/16 "			
12	< 10	1/4 (1)			
13	< 10	1/4 (2)			

*Reed
17 Aug 07*

COMMENTS No detectable contamination w/ GM5M.
Release all items except 1/32 & 1/16 inserts. RWD 17 Aug 07
1/32 & 1/16 to be washed and rechecked.

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 17 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS Release surveys
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M^x Ludlum 3 LSC^x

NO.	DPM	LOCATION	NO.	DPM	LOCATION
16	< 10	B-Ram ¹⁶ C source S/N 2081	26	< 10	Hotplate D561 top
17	< 10	B-ram detector S/N 11921	27	< 10	" " front
18	< 10	B-ram " S/N 11783	28	< 10	" " sides + bottom
19	< 10	Hamilton Syringe (1)	29	< 10	" " Top after cleaning
20	< 10	" " (2)			
21	< 10	²⁵⁰⁻⁰⁴⁰³⁰⁸ O ₂ cylinder bottom			
22	< 10	" " top			
23	< 10	" " valve			
24	< 10	" " cover			
25	< 10	" " Reg. FZ36140			

COMMENTS D 561 ~ 50 cpm above bkg w/ GMSM.
Release B-Ram detectors, Hamilton syringes, O₂ cylinder,
regulator, hot plate, RWG 20 Aug 07

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 23 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS Release Surveys
 INSTRUMENT(S) Bicron Frisk-Tech X Bicron Surveyor M Ludlum 3 LSC X

NO.	DPM	LOCATION	NO.	DPM	LOCATION
30	< 10	AMBIS-plates 1/32"	40	< 10	S.S. TAC plateholder ^{outside}
31	< 10	" " 1/16"	41	< 10	" " " " ^{inside}
32	< 10	1000ml bottles outside	42	< 10	auto transformer ^{D 2163}
33	< 10	" " lids	43	< 10	Fraction collector ^{bases} trays
34	< 10	" " " inside-swab	44	< 10	" " " " 1
35	< 10	" " " " -swab	45	< 10	" " " " 2
36	< 10	Plastic TAC plateholder ^{outside}	46	< 10	" " " " 3
37	< 10	" " " " ^{inside}	47	< 10	" " " " 4
38	< 10	" " " " [*] side plates	48	< 10	" " " " 5
39	< 10	" " " " tray	49	< 10	" " " " 6
			50	< 10	" " " " 7

COMMENTS *side plate 500 cpm w/ GM SM before cleaning, No detectable contamination w/ GM SM.
 Release all items page 142 - 24 Aug 07 RGR

RGR

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-106 DATE 23 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS Release Survey 3
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC

NO.	DPM	LOCATION	NO.	DPM	LOCATION
51	< 10	Block heater D 982	23 Aug 07 24 Aug 07		
52	< 10	block 1			
53	< 10	2			
54	< 10	3			
55	< 10	block 1 swab			
56	< 10	" 2 "			
57	< 10	" 3 "			
58	< 10	Sieve #60 mesh			
59	< 10	" #10			
60	< 10	magnetic stirrer			

COMMENTS No detectable contamination w/ GEMMA

Handwritten signature

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 27 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS Release Surveys
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC

NO.	DPM	LOCATION	NO.	DPM	LOCATION
61	< 10	propane torch	71	< 10	1000 ml Erlenmeyer (2)
62	< 10	DG65 stirrer	72	< 10	" " (3)
63	30	Lab Jack	73	< 10	mini-impinger
64	< 10	B-shield box	74	< 10	Lid Freezer 32 D95B
65	< 10	Filter funnels (3)	75	< 10	Front
66	< 10	Empty columns (2)	76	< 10	inside lid
67	< 10	1000 ml cylinder	77	< 10	" compartment
68	< 10	250 ml beaker	78	< 10	1 ml H ₂ O from ice
69	< 10	10 ml cylinder			
70	< 10	1000 ml erlenmeyer			

COMMENTS No detectable contamination w/ GMSM.
RAM label removed from freezer in sample room, Room 105
Release all items except lab jack Riley 28 Aug 07

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 30 Aug 07
 SURVEYOR release survey R. Granlund
 CONDITIONS _____
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3x LSCx

NO.	DPM	LOCATION	NO.	DPM	LOCATION
78	< 10	4-liter bottle carrier	88	< 10	evap dish
79	25	Lab-Jack-cleansed	89	40	glass 5ml syringe
80	< 10	Blot-tray	90	< 10	stir bar retriever
81	< 10	Tray lds - 2			
82	26	poly sample cont.			
83	< 10	poly beakers			
84	< 10	glass beakers			
85	< 10	R1-tube holder			
86	< 10	2-L bottle			
87	< 10	3/4" tubing			

R. Granlund
21 Aug 07

COMMENTS Discard glass syringe, reclean lab jack, ret-
discard poly sample containers, release other items.
No detectable contamination w/ GM5M.

RADIOACTIVE CONTAMINATION SURVEY


LOCATION B1-101 DATE 31 Aug 07
 SURVEYOR R. Granlund
 CONDITIONS Release Survey
 INSTRUMENT(S) Bicron Frisk-Tech x Bicron Surveyor M Ludlum 3 LSCx

NO.	DPM	LOCATION	NO.	DPM	LOCATION
91	< 10	1ml island sink trap	_____		
92	< 10	1ml wall sink trap			
93	< 10	lab jack end plates			
94	< 10	↓ ↓ scissors			
95	< 10	↓ ↓ lateral supports			

COMMENTS No detectable contamination w/ GeMS4
lab jack released 4 Sep 07 RLG

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 10 Sep 07
 SURVEYOR R. Granlund
 CONDITIONS FSS E. wall cabinets
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC & Ludlum 2360 S/N 177116 / Ludlum 43-89 SKV PR184173
M=800 Bonly Bkg 27B c/min C-17 Check 10,555 c/min





NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	< 10	side	11	< 10	back knockhole
2	< 10	door front	12	< 10	side "
3	< 10	" rest	13	< 10	drawer front
4	< 10	cavity bottom	14	< 10	" inside
5	< 10	" sides	15	< 10	shelf
6	< 10	drawer front	16	< 10	cavity - bottom
7	< 10	" inside	17	< 10	" sides
8	< 10	drawer & front	18	< 10	door front
9	< 10	drawer inside	19	< 10	" rear
10	< 10	side knockhole	20		area to survey

COMMENTS 17-279cpm, 4-268, 7-265, 9-267 static measurements, 1min counts 100% survey w/ ludlum 2360 - no detectable contamination.

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 11 Sep 07
 SURVEYOR R. Granlund
 CONDITIONS FSS S.wall cabinet
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC
Ludlum 2360 S/N 177116 / Ludlum 43-89 S/N PR184173
HV800 B only Bkg 296cpm C-14 check 10,460cpm




NO.	DPM	LOCATION	NO.	DPM	LOCATION
21	< 10	L. front	 R. Granlund 11 Sep 07 		
22	< 10	R. front			
23	< 10	L. inside			
24	< 10	R. inside			
25	< 10	base board.			
26	< 10	base			
27	< 10	R. side			
28	< 10	L. side			
29	< 10	Back board			
30	< 10	shelf.			

COMMENTS static #26-30cpm, #29-274cpm, #30-290cpm
No detectable contamination on 90+% scan (pp1-5)

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 11 Sep 07
 SURVEYOR R. Granlund
 CONDITIONS F55 E. side E. island cabinets
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC





NO.	DPM	LOCATION	NO.	DPM	LOCATION
31	< 10	Front	41	< 10	Knockhole rear
32	< 10	Front	42	< 10	" right
33	< 10	Drawer, inside	43	< 10	Front
34	< 10	Door, inside	44	< 10	"
35	< 10	base	45	< 10	Drawer inside
36	< 10	side + rear	46	< 10	Door inside
37	< 10	shelf	47	< 10	base
38	< 10	Front	48	< 10	side + rear
39	< 10	Drawer inside	49	< 10	shelf.
40	< 10	Knockhole L	50	< 10	Front

COMMENTS Static #33-290cpm, #37-306cpm, #39-276, #45-323cpm,
#35-311cpm, #47-304, #49-319cpm, #51-299cpm

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 11 Sep 07
 SURVEYOR R. Granlund
 CONDITIONS FSS. E. side E. Island cabinets, E. Island sink
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC





NO.	DPM	LOCATION	NO.	DPM	LOCATION
51	< 10	Drawer inside	61	< 10	base L.
52	< 10	knockhole L.	62	< 10	" R
53	< 10	" Rear	63	< 10	sides
54	< 10	" R.	64	< 10	rear
55	< 10	Front.			
56	< 10	Front			
57	< 10	L. door front			
58	< 10	R " "			
59	< 10	L. door inside			
60	< 10	R. " "			

*Rec'd
11 Sep 07*

COMMENTS static #61-313cpm, #62-294cpm

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 11 Sep 07
 SURVEYOR R. Granlund
 CONDITIONS F.S.S. W. side E. island cabinets
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSCx




NO.	DPM	LOCATION	NO.	DPM	LOCATION
65	< 10	Front	75	< 10	Knockhole L.
66	< 10	Drawer, inside	76	< 10	Front
67	< 10	Front	77	< 10	drawer inside
68	< 10	Door inside	78	< 10	front
69	< 10	base	79	< 10	door w/ 11 sep or 1007 drawer inside
70	< 10	side + rear	80	< 10	base
71	< 10	Front	81	< 10	sides + rear
72	< 10	drawer, inside	82	< 10	front
73	< 10	Knockhole R	83	< 10	drawer inside
74	< 10	" rear	84	< 10	front

COMMENTS static #66-280cpm, #69-282cpm #72-290cpm
#77-294cpm #80-316cpm, #88-269cpm, #83-339cpm, #89-276

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 11 Sep 07
 SURVEYOR R. Geranlund
 CONDITIONS FSS w/ side E. island cabinets
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC





NO.	DPM	LOCATION	NO.	DPM	LOCATION
85	< 10	kneehole L	 R. Geranlund 11 Sep 07 		
86	< 10	" rear			
87	< 10	" right			
88	< 10	shelf			
89	< 10	shelf			

COMMENTS _____

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 12 Sep 07
 SURVEYOR R. Granlund
 CONDITIONS FSS s. wall sink
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC X
Ludlum 2360 SIN 179116 / Ludlum 93-89 SIN PR 184173
HV 800 B only Bkg 289cpm (10min) ¹⁴C check 10,402cpm





NO.	DPM	LOCATION	NO.	DPM	LOCATION
91	< 10 < 100	RE RUM 12 Sep 07 front	1		front
92	< 10	L. door outside	2		drawer inside
93	< 10	R. " "	3		door outside
94	< 10	L. " inside	4		" inside
95	< 10	R. " "	5		base
96	< 10	L. base	6		side + back
97	< 10	R. "	7		knee hole L.
98	< 10	sides	8		" rear
99	< 10	rear	9		" R
100	< 10	Trap outside	10		shelf

COMMENTS Static 1 min counts #96-285cpm #97-201cpm

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 12 Sep 07
 SURVEYOR R. Granlund
 CONDITIONS FSS E. side w. island
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC





NO.	DPM	LOCATION	NO.	DPM	LOCATION
11	< 10	front	21	< 10	kneehole h.
12	< 10	drawer inside	22	< 10	" rear
13	< 10	front	23	< 10	" R.
14	< 10	drawer inside	24	< 10	front.
15	< 10	door outside			
16	< 10	" inside			
17	< 10	base			
18	< 10	side + rear			
19	< 10	front			
20	< 10	drawer			

R. Granlund
14 Sep 07

COMMENTS Static / min counts # 2-265cpm, # 5-315cpm, # 12-292cpm
14-279cpm, # 17-282cpm # 10-276cpm, # 20-293cpm

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 14 Sep 07
 SURVEYOR R. Granlund
 CONDITIONS FSS Lab cabinets w. island end section
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC
Ludlum 2360 S/N 177116 / Ludlum 43-B9 S/N PR184173
800V, β only; Bkg 277 cpm (10 min), ^{137}C check 10, 531 cpm




NO.	DPM	LOCATION	NO.	DPM	LOCATION
25	< 10	front	35	< 10	base
26	< 10	drawer inside	36	< 10	sides and rear
27	< 10	front	37	< 10	shelf
28	< 10	drawer inside	38		
29	< 10	door outside	39		
30	< 10	" inside	40		
31	< 10	" outside	41		
32	< 10	" inside	42		
33	< 10	base	43		
34	< 10	sides + rear	44		

COMMENTS Static /min counts #26-275 cpm, #28-297 cpm, #33-210 cpm
#35-278 cpm #37-301

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 14 Sep 07
 SURVEYOR R. Granlund
 CONDITIONS FSS - Lab cabinets
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC X



NO.	DPM	LOCATION	NO.	DPM	LOCATION
38	< 10	front	48	< 10	drawer, inside
39	< 10	front	49	< 10	front
40	< 10	drawer inside	50	< 10	drawer, inside
41	< 10	front	51	< 10	51 door front
42	< 10	drawer inside	52	< 10	" inside
43	< 10	door front	53	< 10	shelf
44	< 10	" inside	54	< 10	base
45	< 10	sides & rear	55	< 10	base
46	< 10	shelf	56	< 10	sides & rear
47	< 10	front	57	< 10	knee hole rear

COMMENTS 1 min static counts #40-297, #42-270, #55-278, #46-251, #48-276, #50-298, #54-276, #53-289

RADIOACTIVE CONTAMINATION SURVEY



LOCATION B1-101

DATE 14 Sep 07

SURVEYOR R. Granlund

CONDITIONS FSS - s. end of w. aisle

INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC





NO.	DPM	LOCATION	NO.	DPM	LOCATION
58	< 10	kneehole, R	68	< 10	base R.
59	< 10	" L	69	< 10	sides
60	< 10	" rear	70	< 10	rear
61	< 10	" R	71	< 10	door, front
62	< 10	" L.	72	< 10	" inside
63	< 10	door front	73	< 10	" front
64	< 10	" inside	74	< 10	" inside
65	< 10	" "	75	< 10	base L.
66	< 10	" front	76	< 10	base R
67	< 10	base L.	77	< 10	sides

COMMENTS static 1min count # 67-305, # 68-310, # 77-

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 14 Sep 07
 SURVEYOR R. Gronlund
 CONDITIONS F.S.S. - W. wall cabinets
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC




Scan - 17 Sep 07
 Ludlum 2360 S/N 177116
 Ludlum 43-89/S/N PR184173
 800V
 * C check 10,466 c/min
 Bkg - 278 cpm (10min)

NO.	DPM	LOCATION	NO.	DPM	LOCATION
78	< 10	rear	88	< 10	base R.
79	< 10	shelf	89	< 10	sides
80	< 10	kneehole L.	90	< 10	rear
81	< 10	" rear	91	< 10	shelf
82	< 10	" R	92	< 10	drawer front
83	< 10	door front	93	< 10	" inside
84	< 10	" inside	94	< 10	" front
85	< 10	" front	95	< 10	" inside
86	< 10	" inside	96	< 10	kneehole L.
87	< 10	base L.	97	< 10	" rear

COMMENTS static /min counts #75-337, #76-311, #79-284
#87-326, #88-293, #91-287, #93-275, #95-314, #93-316
#1-336, #10-301 Bkg - 316 c/min

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 14 Sep 07
 SURVEYOR R. Geranlund
 CONDITIONS E.S.S. Lab cabinets, bookcase
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC X



NO.	DPM	LOCATION	NO.	DPM	LOCATION
98	< 10	kneehole near R	8	< 10	bookcase sides
99	< 10	door front	9	< 10	" rear
100	< 10	" inside	10	< 10	" shelves
1	< 10	" front	11	< 10	" back
2	< 10	" inside			
3	< 10	base L			
4	< 10	base R			
5	< 10	sides			
6	< 10	rear			
7	< 10	bookcase top			

Recy
14 Sep 07

COMMENTS release bookcase 17 Sep 07

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 17 Sep 07


SURVEYOR R. Granlund

CONDITIONS F.S.S. File cabinet

INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC

Ludlum 2360 S/N 177116
Ludlum 43-89 S/N PR184173
800V

¹³⁷C check 10,466 c/min
Bkg 278 cpm(10 min)




NO.	DPM	LOCATION	NO.	DPM	LOCATION
12	< 10	top	 R. Granlund 18 Sep 07 		
13	< 10	side R			
14	< 10	" L			
15	< 10	rear			
16	< 10	front			
17	< 10	top drawer inside			
18	< 10	bottom " "			
19	< 10	bottom			
20	< 10	book case bottom			
		R. Granlund 18 Sep 07			

COMMENTS 1 min static count #18-283, #17-270, #12
#87(p.49) 2970/10min bkg 2992c/10min, #12-291,
Release file cabinet 18 Sep 07

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 18 Sep 07
 SURVEYOR R. Granlund
 CONDITIONS F.S.S. walls
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSCA
Ludlum 2360 SIN 177116 Harvey H-3
Ludlum 13-89 SIN PR 184173
800V 11,211, 11,147 cpm
¹⁴C check 9285, 9398 cpm
Bkg 270.1cpm (10min)




NO.	DPM	LOCATION	NO.	DPM	LOCATION
21	< 10	E. Wall	31	< 10	W. Wall window
22	< 10	"	32	< 10	shade
23	< 10	"	33	< 10	window
24	< 10	"	34	< 10	shade
25	< 10	"	35	< 10	window
26	< 10	"	36	< 10	shade
27	< 10	"	37	< 10	window
28	< 10	"	38	< 10	↓ shade
29	< 10	"	39	< 10	W. Wall
30	< 10	"	40	< 10	↑

COMMENTS *scanned area behind paper coating experiment.
No detectable contamination on wall w/scan.

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 18 Sep 07
 SURVEYOR R. Granlund
 CONDITIONS F.S.S.
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC





NO.	DPM	LOCATION	NO.	DPM	LOCATION
41	< 10	W. Wall	51	< 10	N. Wall
42	< 10	"	52	< 10	" door
43	< 10	"	53	< 10	"
44	< 10	"	54	< 10	"
45	< 10	"	55	< 10	"
46	< 10	"			
47	< 10	S. Wall			
48	< 10	"			
49	< 10	"			
50	< 10	"			

COMMENTS * scanned area

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 18 Sep 07
 SURVEYOR R. Granlund
 CONDITIONS F.S.S.
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
47			47		
48			48		
49			49		
50			50		
51			51		
52			52		
53			53		
54			54		

Recy 18 Sep 07

COMMENTS * scanned area behind sink and extraction manifold.

RADIOACTIVE CONTAMINATION SURVEY




LOCATION B1-101 DATE 19 Sep 07

SURVEYOR R. Granlund

CONDITIONS A.S.S. ceiling

INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC X

Ludlum 2300 S/N 127116
Ludlum 43-89 S/N PR184123
800V Bkg. 2684/10min
1 σ check 10,644cpm
Richard Altma



NO.	DPM	LOCATION	NO.	DPM	LOCATION
56	< 10	light cover	 R. Granlund 21 Sep 07 		
57	< 10	" "			
58	< 10	tile			
59	< 10	light cover			
60	< 10	tile			
61	< 10	light cover			
62	< 10	tile			
63	< 10	light cover			
64	< 10	tile			
		check 21 Sep 07			

COMMENTS No detectable contamination on scan of light covers or tile. Tile background is ~ 400cpm from natural activity (²³²Th). Normal background for other lab surfaces is 279±10(1σ)

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 1 Oct 07
 SURVEYOR R. Granlund
 CONDITIONS release survey
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC

NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	< 100	20-L carboy #3 outside	11	< 100	5 gal carboy #6 outside
2	7140	#3 inside	12	710	" " #6 inside
3	< 100	#4 outside	13	< 100	siphon hose outside
4	< 100	↓ ↓ * + inside	14	< 100	" " inside
5	< 100	10-L carboy #1 outside	15	< 100	" " tap
6	< 100	#1 inside	16	< 100	cap 1
7	< 100	#2 outside	17	< 100	2
8	< 100	↓ ↓ #2 inside	18	< 100	3
9	< 100	5-gal carboy #5 outside	19	< 100	4
10	< 100	" " #5 inside	20	< 100	↓ 5,6

COMMENTS _____

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 1 Oct 07
 SURVEYOR R. Gonzalez
 CONDITIONS release survey
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC

NO.	DPM	LOCATION	NO.	DPM	LOCATION
21	< 100	carboy 3- inside *	Reel 2 Oct 07		
22	< 100	" 6 "			
23	< 100	sip kton #2 p			
24	< 100	poly chips from cutting			
25	< 100	waste bucket poly			
26	< 100	" " rubber			

COMMENTS * inside after cutting open and cleaning.
Carboy #3 & #6 discarded after cleaning.



3058 Research Drive Phone: 814-272-1039
 State College, PA 16801 Fax: 814-231-1580

RADIOACTIVE CONTAMINATION SURVEY

Page 58


Location: Radioisotope Lab (B1-101) Date: 3 Oct 07

Surveyor: R. Granlund

Conditions: FSS Benchtops

Instrument(s): Bicron Frisk-Tech Bicron Surveyor M
 Ludlum 3 LSC

Ludlum 2360/43-89 S/N 177116/PR 184123 800V Bkg-2854/10min
 MC 95509/10min



No.	dpm	Location	No.	dpm	Location
71	< 10	benchtop	81	< 10	benchtop
72	< 10	"	82	< 10	"
73	< 10	"	83	< 10	"
74	< 10	"	84	< 10	"
75	< 10	"	85	< 10	"
76	< 10	"	86	< 10	"
77	< 10	"	87	< 10	"
78	< 10	"	88	< 10	"
79	< 10	"	89	< 10	"
80	< 10	"	90	< 10	"

Comments:

No detectable contamination w/ ludlum 2360/
43-89 on ~100% surface scan w/ 1 min integrate during
scan.




3058 Research Drive Phone: 814-272-1039
 State College, PA 16801 Fax: 814-231-1580

RADIOACTIVE CONTAMINATION SURVEY

Page 59

Location: Radioisotope Lab (B1-101) Date: 30 Oct 07
 Surveyor: R. Granlund
 Conditions: FSS Benchtops
 Instrument(s): Bicron Frisk-Tech Bicron Surveyor M
 Ludlum 3 LSC
Ludlum 2360/43-89



No.	dpm	Location	No.	dpm	Location
91	< 10	benchtop	Ready 9 Oct 07		
92	< 10	"			
93	< 10	"			
4	< 10				
5	< 10				
6	< 10				
7	< 10				
8	< 10				

Ready 9 Oct 07

Ready 9 Oct 07

Comments:



3058 Research Drive Phone: 814-272-1039
 State College, PA 16801 Fax: 814-231-1580

RADIOACTIVE CONTAMINATION SURVEY

Page 60


Location: Radioisotope Lab (B1-101) Date: 4 Oct 07

Surveyor: R. Granlund

Conditions: FSS counter tops

Instrument(s): Bicon Frisk-Tech Bicon Surveyor M

Ludlum 3 LSC
 Ludlum 2360/43-89 SIN 177116/PR184173
 Rtg 275.5 cpm (omix)
 1st C 95096/10 min




No.	dpm	Location	No.	dpm	Location
9	< 10	benchtap	19	< 10	bench top
10	< 10	"	20	< 10	"
11	< 10	"			
12	< 10	"			
13	< 10	"			
14	< 10	"			
15	< 10	"			
16	11	"			
17	< 10	sink			
18	< 10	benchtap			

Comments:

sample #41 sink drain (S.wall) surface of trap
#42 bottom of trap 2ml sample in 20 ml packard
opto-gold #41 = 1.1×10^{-6} uCi/ml #42 = 1.6×10^{-6} uCi/ml*
No detectable contamination w/ ludlum 2360/43-89

RADIOACTIVE CONTAMINATION SURVEY

LOCATION B1-101 DATE 4 Oct 07
 SURVEYOR R. Granlund
 CONDITIONS Release Survey LSC Counters
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC X



NO.	DPM	LOCATION	NO.	DPM	LOCATION
21	< 10	LSC # 2 front-lower	31	< 10	LSC # 1 - same as 42 front-lower
22	< 10	" upper lid	32	< 10	" upper lid
23	< 10	top	33	< 10	top
24	< 10	side L.	34	< 10	side L.
25	< 10	side R	35	< 10	side R
26	< 10	rear	36	< 10	rear
27	< 10	inside tray L	37	< 10	inside tray L
28	< 10	" " R	38	< 10	" " R
29	< 10	monitor	39	< 10	monitor
30	< 10	printer	40	< 10	printer


COMMENTS No detectable contamination on scan with Ludlum 2360/43-89

RADIOACTIVE CONTAMINATION SURVEY

Page 62

Location: Radioisotope Lab (B1-101) Date: 10 Oct 07
 Surveyor: R. Granlund
 Conditions: FSS Floor
 Instrument(s): Bicron Frisk-Tech Bicron Surveyor M
 Ludlum 3 LSC

Ludlum 2360/43-89
 S/N 177116/PR 18+173
 Bkg - 28+9/10 min
 4°C check - 97347/10 min



No.	dpm	Location	No.	dpm	Location
43	< 10	floor	53	< 10	floor
44	< 10	"	54	< 10	"
45	35	"	55	< 10	"
46	32	"	56	< 10	"
47	< 10	"	57	< 10	"
48	< 10	"	58	< 10	"
49	< 10	"	59	< 10	"
50	< 10	"	60	< 10	"
51	< 10	"	61	< 10	"
52	< 10	"	62	< 10	"


Comments:

Smears cover 1 ft² (1 tile) shaded areas indicate scra (~10 ft²
in 2 min) No detectable contamination in scan with
Ludlum 2360/43-89.

RADIOACTIVE CONTAMINATION SURVEY

Page 63

Location: Radioisotope Lab (B1-101) Date: 10 Oct 07
 Surveyor: R. Granlund
 Conditions: FSS of floor
 Instrument(s): Bicron Frisk-Tech Bicron Surveyor M
 Ludlum 3 LSC



No.	dpm	Location	No.	dpm	Location
63	< 10	floor	73	< 10	floor
64	< 10	"	74	< 10	"
65	< 10	"	75	< 10	"
66	< 10	"	76	< 10	"
67	< 10	"	77	< 10	"
68	< 10	"	78	< 10	"
69	< 10	"	79	< 10	"
70	< 10	"			
71	< 10	"			
72	< 10	"			


RG
10 Oct 07

Comments:

RADIOACTIVE CONTAMINATION SURVEY

Page 64

Location: Radioisotope Lab (B1-101) Date: 11 Oct 07
 Surveyor: R. Granlund
 Conditions: FSS floors
 Instrument(s): Bicon Frisk-Tech Bicon Surveyor M
 Ludlum 3 CLSC



No.	dpm	Location	No.	dpm	Location
80	< 10	floor	90	< 10	shelves freezer #10
81	< 10	"	91	28	Frost from freezer #35
82	< 10	"			
83	< 10	" hallway			
84	< 10	" "			
85	< 10	" "			
86	< 10	floor - walk in freezer room			
87	< 10	floor - inside freezer #10			
88	< 10	" " " "			
89	< 10	" " " "			

Relly
16 Oct 07

Comments:

#80, 81, 82 @ locations # 45, 46 in 10 Oct 07 survey
after cleaning: #86 floor outside freezer #10, #91 2ml
melted frost from -80°C freezer #35. Freezers #10
& #35 used at one time to store RAM samples.




3058 Research Drive Phone: 814-272-1039
 State College, PA 16801 Fax: 814-231-1580

RADIOACTIVE CONTAMINATION SURVEY

Page 65

Location: Radioisotope Lab (B1-101) Date: 16 Oct 07
 Surveyor: R. Granlund
 Conditions: Release survey.
 Instrument(s): Bicron Frisk-Tech Bicron Surveyor M
 Ludlum 3 LSC



No.	dpm	Location	No.	dpm	Location
92	< 10	Trash can 1 outside	66	< 10	trash can 2-inside
93	< 10	" " " inside	67	< 10	chair 2-upper
94	< 10	Chair 1-top	68	< 10	" 2 under
95	< 10	" " - under	69	< 10	" 2 base.
96	< 10	" " - base	70	< 10	Printer D196
97	< 10	5-gal hood mat pair/out	71	< 10	computer D194
98	< 10	" " " inside	72	< 10	paper cutter
99	< 10	20-L carboy 2.98	73	< 10	letter trays
100	< 10	" " " 2.96	74	< 10	misc office supplies
65	< 10	Trash can 2-outside	75	< 10	balance

Comments:

No detectable contamination w/GM5M.




3058 Research Drive Phone: 814-272-1039
 State College, PA 16801 Fax: 814-231-1580

RADIOACTIVE CONTAMINATION SURVEY

Page 46

Location: Radioisotope Lab (B1-101) Date: 16 Oct 07
 Surveyor: R. Gramlund
 Conditions: Release Survey
 Instrument(s): Bicron Frisk-Tech Bicron Surveyor M
 Ludlum 3 LSC




No.	dpm	Location	No.	dpm	Location
76	< 10	balance cover	86	< 10	chair 3 - under
77	< 10	balance tray top	87	< 10	" " base
78	< 10	" " bottom	88*	< 10	pipetor IN 980
79	< 10	pipetor IN 516			
80	23	" IN 980			
81	< 10	" IN 984			
82	< 10	" IN 982			
83	< 10	Trash can 3 - outside			
84	< 10	" inside			
85	< 10	chair 3 - upper			

Comments:

* pipetor IN 980 cleaned and recalibrated 18 Oct 07.

RADIOACTIVE CONTAMINATION SURVEY

LOCATION II-111 DATE 26 May 05
 SURVEYOR R. W. Granlund
 CONDITIONS Survey after Hercules synthesis
 INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	<100	water bath	11	<100 <100	floor
2	<100	ring stand	12	<100 <100	"
3	<100 <100	overhead stirrer	13	<100 <100	"
4	<100 <100	hood	14	<100 <100	"
5	<100	Thermos	15	<100 <100	"
6	<100	Lab Jack & Hotplate	16	<100	Sink
7	<100	Trays	17	<100	benchtop
8	<100 <100	hood	18	<100	"
9	<100 <100	"			
10	<100	Sash			

R. W. Granlund
27 May 05

COMMENTS GMSK check α - count # 3, 4, 8, 9, 11-15 w/ Frisk-Tech then count all overnight w/ LSC




^{check} RADIOACTIVE CONTAMINATION SURVEY

LOCATION Bldg 2 - 111 DATE 14 Jun 05

SURVEYOR R.W. Granlund

CONDITIONS survey after ¹⁴C synthesis by Hercules

INSTRUMENT(S) Bicron Frisk-Tech Bicron Surveyor M Ludlum 3 LSC



NO.	DPM	LOCATION	NO.	DPM	LOCATION
1	< 100	^{check} hood benchtop	11	< 100	floor
2	< 100	"	12	< 100	"
3	< 100	"	13	< 100	"
4	= 100	sink	14	< 100	hood
5	< 100	hood	15	< 100	hood apron
6	< 100	balance	16	< 100	hood
7	< 100	di. H ₂ O table	17	< 100	"
8	< 100	back sink	18	< 100	floor
9	< 100	floor			
10	< 100	"			

COMMENTS GMSM check @

KEVIN LLOYD
(814) 231-8032
MPI RESEARCH - STATE COLLEGE
3048 RESEARCH DRIVE
STATE COLLEGE PA 16801

2 LBS


1 OF 1

SHIP TO:

LICENSING ASSISTANT SECTION
NUCLEAR MATERIALS SAFETY BRANCH
US NUCLEAR REGULATORY COMMISSION, I
475 ALLENDALE ROAD
KING OF PRUSSIA PA 19406-1415

141361

NUCLEAR MATERIALS SAFETY BRANCH
475 ALLENDALE RD
KING OF PRUSSIA PA 19406-1415
P:RED S:09 I:4F
0211-RDC
1Z4657X50148430017 103
QTG9QSM PAWES170 Nov 23 06:00:12 2007
TB 1940 HIP 6.2.1 INT4420




PA 193 9-02




UPS NEXT DAY AIR

TRACKING #: 1Z 465 7X5 01 4843 0017

1



EXTREMELY URGENT

UPS Next Day Air®

BILLING: P/P

Dept No.: Administration
REF 2: Report

WS 9 0 49

LP2442 72 0A 10/2007



International Shipping Notice - Cargo hereunder may be subject to the rules relating to liability and other terms and/or conditions established by the Convention for the Unification of Certain Rules Relating to International Carriage by Air (the "Warsaw Convention") and/or the Convention on the Contract for the International Carriage of Goods by Road (the "CMR Convention"). These conventions, technology or software were exported from the U.S. in accordance with the Export Administration Regulations. Diversion contrary to U.S. law prohibited. For shipping pieces, call 1-800-750-7622. United Parcel Service, Louisville, KY

This is to acknowledge the receipt of your letter/application dated

11/15/2007, and to inform you that the initial processing which includes an administrative review has been performed.

TEAM. 37-30095-01
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 141361.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.