

GEORGE HARMS CONSTRUCTION
PO BOX 817
FARMINGDALE, NJ 07727

RECEIVED
REGION 1

2007 DEC -3 PM 12:36

United States Nuclear Regulatory Commission
Regional Administrator Region I
475 Allendale Road
King Of Prussia PA 19406
And
Nuclear Regulatory Commission
ATTN Document Control Desk
Washington DC 20555

03022207/2007001

Reply to a Notice of Violation

License # 29-23456-01

In regards to violation concerning a use not authorized by our material license response is as noted


The reason for violation is I referenced a Troxler drawing that included the 4640 gauge but it was not included on our last license renewal, which I did not notice upon review.

License has since been amended to correct this deficiency.

As far as labeling of UN2974 package. I was under impression that domestic travel was not included in this change but realize I was incorrect and all shipping papers and labeling has been corrected to UN3332 as needed.

If there is any further questions or comments please feel free to contact me at 908-618-6331

Sincerely,


Roberto J. Garcia

NMSS/RGNI MATERIALS-004