



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory
Certificate #0972-01

Certificate of Calibration

Submitted By: DIV20

Address: B57

Contact: DON BANNON

Manufacturer / Model: SARTORIUS / ME215S

Description: BALANCE

Serial No: 12809099

Asset No: 008780

Procedure: BALANCES & SCALES - 1 DEC 06

Work Order: 303077426

Date Issued: Nov 13, 2007

Calibration Date: Nov 13, 2007

***Calibration Due:** May 13, 2008

Calibration Location: B57

Environment: Temp. 71.0°F Hum. 50 %RH

****Data Type:** FOUND-LEFT

DivID/Location: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

Remarks: None

Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
001708	C863	RICE LAKE	10G	WEIGHT, CLASS S	Aug 09, 08
001709	C864	RICE LAKE	20G	WEIGHT, CLASS 1	Aug 09, 08
001710	C865	RICE LAKE	20G	WEIGHT, CLASS 1	Aug 09, 08
001711	C866	RICE LAKE	50G	WEIGHT, CLASS 1	Aug 09, 08
001712	C867	RICE LAKE	100G	WEIGHT, CLASS S	Aug 09, 08
001714	C869	RICE LAKE	200G	WEIGHT, CLASS S	Aug 09, 08

Scott Kuehl

Reviewed by: () wgh (✓) srk () jrg () blt () pwc

Metrology Technician

m:\a2la1.rpt Rev date August 15, 2005

Mark Romero

Measurements by Mark Romero

Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303077426	Mfr:	Sartorius	Technician:	Mark Romero
Asset No:	008780	Model:	ME215S	Cal Date:	13-Nov-07
Serial No:	12809099	Type:	Balance		

Remarks:
Ambient Conditions 71 °F 50 % RH 14.44 PSIA

Function/Range	Applied	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
	grams	grams	grams	grams	grams	Result
Corner Load						
Reference	100.00000					
Front	0.00000	-0.00002	-0.00002	0.00020		Pass
Rear	0.00000	0.00000	0.00000	0.00020		Pass
Left	0.00000	0.00000	0.00000	0.00020		Pass
Right	0.00000	0.00000	0.00000	0.00020		Pass
Repeatability						
1	100.00000	100.00000				
2	100.00000	100.00005				
3	100.00000	100.00003				
4	100.00000	100.00006				
5	100.00000	100.00007				
6	100.00000	100.00006				
7	100.00000	100.00007				
8	100.00000	100.00006				
9	100.00000	100.00009				
10	100.00000	100.00006				
Std Deviation		0.000025		0.000050		Pass
Linearity						
Nominal	Conventional					
Value (g)	Mass					
0.00000	0.00000	0.00000	0.00000	0.00020	0.00011	Pass
20.00000	20.00003	20.00000	-0.00003	0.00020	0.00011	Pass
40.00000	40.00012	40.00004	-0.00008	0.00020	0.00011	Pass
60.00000	59.99994	59.99990	-0.00004	0.00020	0.00011	Pass
80.00000	79.99997	79.99993	-0.00004	0.00020	0.00011	Pass
100.00000	99.99990	99.99976	-0.00014	0.00020	0.00011	Pass
120.00000	119.99993	119.99981	-0.00012	0.00020	0.00011	Pass
140.00000	140.00002	139.99988	-0.00014	0.00020	0.00011	Pass
160.00000	159.99984	159.99970	-0.00014	0.00020	0.00011	Pass
180.00000	179.99987	179.99975	-0.00012	0.00020	0.00011	Pass
200.00000	200.00003	199.99997	-0.00006	0.00020	0.00011	Pass

End of Report