

U.S. NUCLEAR REGULATORY COMMISSION		Date: 12/19/07
TELEPHONE CONVERSATION RECORD		Time: 12:35pm
Mail Control 141424 or Report No(s).	License No(s). 37-30768-01	Docket No(s). 030-36111
Name of Licensee:	Exelon PowerLabs	
Name of Participant(s):	John Moore Jenny M. Johansen, Health Physicist	
Telephone No.	610-380-2464 610-337-5071	
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	SS Model No CDC 26205 not listed as source no. on SSD sheet for IRD-2000 calibrator.	
Summary:	John stated Manufacturer had told him the wrong source no. It should be IPL Cs7.PO3 as listed on the SSD for the IRD-2000	
Action Required:	Correct Model No of the Sealed source requested to IPL Cs7.PO3.	
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date	
Prepared & SISP Review Completed By:	/ RA / Johansen	Date: 12/19/07