

Marvin H. Mann, Assistant Director
Division of Inspection, Washington

DEC 31 1957

Robert W. Kirkman, Director
Inspection Division, WFO

TELETYPE REPORT OF LICENSEE COMPLIANCE INSPECTION REPORT - 10 CFR 30

DEBOL: INB:PBK

Transmitted herewith is the following inspection report involving
"non-compliance":

The Martin Company
Baltimore 3, Maryland

License No. 19-1393-1 through 9 except -4 which is
superseded by -6

w/amend #1, 2 to -1
w/amend #1 to -2
w/amend #1 to -3
w/amend #1 to -6
w/amend #1 to -7
w/amend #1 to -8
w/amend #1 to -9

The following items of non-compliance were brought to the attention
of Mr. Paul Quinn, who stated that these items will be corrected:

20.203 "Caution signs, labels and signals" -(f)(1) "Containers":
labeling of waste containers in waste storage area.

20.3 "License requirements": Licensee possesses a 1 millicurie
Pu-170 source not presently covered by a specific license.

Violation of Condition 14 of license 19-1393-2 and Condition 13
of license 19-1393-9 requiring leak testing of Sr-90 sealed sources.

We feel that a slight hazard exists from the licensee's failure to
leak test Sr-90 sources as required by licenses -2 and -9. We feel
that essentially no hazard exists from the licensee's failure to
label waste containers and require a license for the one millicurie
Pu-170 source.

It should be noted that this licensee 19-1393-2 has not been cited
for failure to provide a visible or audible alarm system as re-
quired by 20.203(c)(2). In view of the physical set-up of the

Al 6

DEC 31 1957

Irvin H. Mann

- 2 -

radiology laboratory and associated facilities; and the fact that Mr. Mann stated that door "A" to the radiography cubicle is either locked or a man posted at this door during all radiographic exposures, we do not feel that this type of device is necessary. Door "B" to the cubicle is normally locked during all radiographic exposures (See item 13, enclosure -2).

It is recommended that a letter be sent to the Martin Company - Attention: Mr. Paul Gatum, Radiological Safety Officer - requiring correction of all items of non-compliance.

enclosure:
non. rpt. (4)

