



January 8th, 2008

U.S. Nuclear Regulatory Commission
ATTN: Document Control Desk
Washington, D.C. 20555

Subject: Reply to a Notice of Violation
License No.: 49-18276-01

To Whom It May Concern:

This letter is to reply to a notice of a violation sent by Vivian Campbell, dated December 18th, 2007. There was one violation; it is addressed in the following:

Violation: Failing to lock and secure hot lab in nuclear medicine area.

1. Reason for violation:
When technologist went to cardiopulmonary to inject a patient the door was not locked upon leaving the room.
2. Corrective steps taken:
We contacted our maintenance department to have a keypad lock installed.
3. Action to avoid further violations:
The door automatically locks every time you leave the room.
4. Date when full compliance achieved.
Full compliance was achieved on December 3rd, 2007, when the keypad lock was installed.

If there are any further questions or concerns please contact Jennifer Heffron, CNMT at 307-739-7298 or Michelle Pask, Imaging Director at 307-739-7295 .

Sincerely,

Robert C. Berlin, M.D.
Radiation Safety Officer

Cc:Regional Administrator, Region IV

jlh

IE07
R&N IV