

Phone: (810) 732-5400 Fax: (810) 733-1624

Ethiraj G. Raj, MD, FACC, FSCAl Mohamadali H. Amlani, MD, FACC Michelle Rossie, PA-C

Date: January 07, 2008

United States Nuclear Regulatory Commission Region III, Materials Licensing Section 2443 Warrenville Road, Suite 210 Lisle, Illinois 60532-4352

RE:

Notification of NRC License

Close-out survey

NRC License No.: 21-26540-01

Facility Name: Mid-Michigan Cardiology Associates, PLC

The purpose of this letter is to close-out our old Nuclear Medicine Department on our current NRC license.

Item#1: A close-out survey was performed on December 12, 2007 and is enclosed for your review.

The last date of use of radioactive material in the old Nuclear Medicine Department location was December 7, 2007 for TI-201 and December 3, 2007 for Tc-99m. No other radioactive materials are used at this facility.

A current copy of the leak test results for the sealed sources used at our facility is enclosed for your review. Sealed Source Leak tests demonstrated no removable contamination. These sources include a Co-57 flood source, Cs-137 and Ba-133 vials for constancy and accuracy testing.

We did not dispose of radioactive materials by release to sewers or incineration. In addition, there were no spills of any long-lived radioactive material or on-site burials of radioactive materials.

Item #2: We would like to change our mailing address to: 1165 South Linden Road, Flint, Michigan 48 32.

Please contact our physics consultant, Laura M Luna of the Medical Physics Consultants, Inc. at 734-662-3197, if you have any questions concerning these requests.

Thank you again for your cooperation.

Cordially

Ethiraj O Raj, MD, FACC, FSCAI

Executive Administration

Mid-Michigan Cardiology Associates

#### Item #1

# Close Out Survey Mid-Michigan Cardiology Associates 2329 Stonebridge, Bldg E Flint, MI 48532

Date Performed: December 12, 2007

Performed By: Laura M. Luna, Physics Consultant

#### Last date of byproduct radiopharmaceutical use:

The last date of use of radioactive material in the old Nuclear Medicine department location site was December 7, 2007 for TI-201 and December 3, 2007 for Tc-99m. No other radioactive materials used at this facility. All radioactive sources were transferred to the new location prior to closeout.

#### Well

Wipe tests were analyzed with a Captus 3000 well counter (S/N 900271). Efficiency for Cesium-137 is 7.46 dpm/cpm. MDA is 203.9 dpm.

#### Survey Meter

Survey performed with Biodex 14C survey meter (S/N 101587) with end window probe calibrated on 03/21/07. Check source reading of 0.8 mR/hr. Range used was 0-.20 mR/hr.

#### Visual Check

The area was checked to ensure that all radioactive waste had been removed. No radioactive material was located.

#### Sealed Source Leak Testing

There is no history of leaks from any sealed sources.


#### Radiation Level Survey

No area within the department demonstrated radiation levels in excess of the background reading of 0.02 mR/hr.

### Removable Contamination

1 minute counts of a wipe sample covering at least a100 square centimeter area

# Close Out Survey License Number 21-26540-01



- 1. Acquisition Station
- Camera
   Acquisition Station
- 4. Camera 5. Counter

- Radioactive Storage
   Package Receipt
   Treadmill #1
- 8. Treadmill #1 9. Treadmill#2

N

# Mid-Michigan Cardiology Associates, P.L.C. License # 21-26540-01

Date: 12/12/07
Well Counter: Captus 3000
Survey Meter: Bicron 14C

Performed By: L. Luna MDA: 203.9 dpm

Check Value: 2.1 mR/hr

Calibrated 03/21/07

	dpm	,	dpm		dpm
1	527	24	508	47	491
2	527	25	470	48	491
3	527	26	470	49	479
4	511	27	470	50	479
5	511	28	490	51	479
6	511	29	490	52	500
7	478	30	490	53	500
8	478	31	497	54	500
9	478	32	497	55	463
10	487	33	497	56	463
11	487	34	483	57	463
12	487	35	483	58	498
13	500	36	483	59	498
14	500	37	496	60	498
15	500	38	496	61	501
16	482	39	496	62	501
17	482	40	506		
18	482	41	506		
19	500	42	506		
20	500	43	512		
21	500	44	512		
22	508	45	512	Bkg.	498
23	508	46	491		

ACTION LEVELS: 1000 dpm Efficiency (dpm/cpm): 7.46

< 0.2 mR/hr

Maximum removeable contamination: 527 dpm -498 dpm = 29 dpm

COMMENTS: No evidence of removable contamination.

## Medical Physics Consultants, Inc.

# Sealed Source Leak Test

Licensee: Mid-Michigan Cardiology Associates, PLC Date: <u>10/11/07</u>

Performed by:

Laura Luna

Nuclide	Туре	Calibration Activity	Calibration Date	n Location	M/N	S/N
Cs-137	Vial	215 uCi	10/28/1993	Hot Lab	NES-356	\$356026080
Ba-133	Vial	271 uCi	5/27/1993	Hot Lab	NES-358	S358015041
Co-57	Flood	10 mCi	6/18/2007	Hot Lab	MED3709	110004
Co-57	Flood	10 mCi	1/9/2006	Hot Lab	MED3709	68774

Comment: The sources listed above were leak tested using a dry wipe technique and were found to have less than 0.005 uCi removable activity. The following Minimum Detectable Activities are based upon a background at the indicated value. Background was at or below these levels when the above tests were completed.

Well Counter: Captus 3000

<u>Nuclide</u> MDA 2.0 x 10 -4 uCi Cs-137 8.8 x 10<sup>-5</sup> uCi Ba-133 1.7 x 10<sup>-5</sup> uCi Co-57

Background 145 counts/1 min

204,counts/1 min

RADIATION SAFETY OFFICER:

Mid-Michigan Cardiology Associates, PLC 1165 South Linden Road Flint, Michigan 48532



United States Nuclear Regulatory Commission Region III, Materials Licensing Section 2443 Warrenville Road, Suite 210 Lisle, Illinois 60532-4352 Fed Ex.
TRK# 8632 0474 2389

NY-ENLA

FRI - 11 JAN A2 PRIORITY OVERNIGHT

> ORD 60532

FedEx Retrieval Cop 0200 irbill 8632 0474 2389 291440827 Date 1/10/08 Sender'S HID-MICHIGAN CARDIOLOGY 810 738-54003 1,800,GoFedEx 1,800,463,3339 1165 S. LINCLEN Company U.S. NUCLEAR REGULATORY COMM. 35:00 PROCESSING SECTION MATERIALS LICENSING SECTION WE CARROLLE VICTORIAN JULIA PROCESSING SECTION AND COMMENTS AND THE PROCESSING SECTION AND COMMENTS AND THE PROCESSING SECTION OF THE P fedex.com 1 0 TU or 60532-433 0

C 2005 FedEx 155476/155475 REV 9/05 RT