


## Telephone Record

Golden Valley Memorial Hospital

24-16597-01, 030-11317, CN 316743, ML 073510243

Conversation with Phil James, Technical Director of Radiology on  
1/7/08.

The reviewer spoke to PJ concerning the leak test history on the one Gd-153 source that was disposed of September 10, 2007 by Bionomics Inc. and was accepted by under South Carolina Radioactive Waste Transport Permit No. 5374-24-07. The licensee has no leak test history for this source since it has decayed down to less than 100 microcuries. At the time of waste pickup, the source had decayed to less than .00027 microcuries.  $1\text{mBq} = 27\text{ microcuries}$ , the source was  $1\text{E-5 mBq}$ . The reviewer explained to PJ that there were no other issues with his NRC license at this time.



---

James Mullauer, 1/7/08