



WEST MICHIGAN CANCER CENTER

Passionate Science. Compassionate Care

14 January 2008

Colleen Carol Casey
Materials Licensing Section
U.S. Nuclear Regulatory Commission, Region III
2443 Warrenville Rd.
Lisle, IL 60532-4352

Subject: Amendment requests for license number 21-32501-01

This is a request of The West Michigan Cancer Center for the following amendment to its materials license no. 21-32501-01.

This letter is in response to your letter of 31 December 2007.


The following section of our license should be changed to read as follows:

Section 8 A: Two sources, one not to exceed 12 Ci and the second not to exceed 6 Ci.

Section 9 A: One source for medical use, as permitted by 10 CRF 35.600, in a Varian Medical Systems, Inc. Model GammaMed Plus remote afterloading brachytherapy device. The source activity may not exceed 10.0 Ci at the time of installation. This source, which is not to exceed 12 Ci while stored pending installation, is to be held in a shipping container for source replacement.

If there are any questions, please do not hesitate to contact me at (269) 373-7407.

Thank you.



Paul Jursinic, Ph.D.
Radiation Safety Officer
West Michigan Cancer Center

P 269.382.2500
P 269.373.7450
P 269.382.2500
P 269.382.2500
P 269.382.2500
P 269.382.2500
P 269.382.2500

