Aguadilla X Ray Office & Body Imaging Center

Aguadilla Medical Building, Ofic, 302, Progreso #2 y #3 P.O. Box 418, Aguadilla, PR 00605 Tcl/Fax 787-891-6165 - Imaging Center 787-891-6565, Fax 787-891-6566

JANUARY 16, 2008

NRC US NUCLEAR REG. COM 475 ALLENDALE ROAD KING OF PRUSIA, PA 19406-1415 NMSBZ

03036369

RE: AMENDMENT TO NRC # 52-25632-01

DEAR MS. PAMELA HENDERSON:

PLEASE AMEND OUR NRC LICENSE TO ADD THE FOLLOWING SOURCE:

ISOTOPOACTIVITYUSECS-13710mCiCAMERA QCNA-22300uCiCAMERA QC

ATTACHED COPIES OF SOURCE DATA SHEET.

THANK YOU FOR YOUR ATTENTION TO THIS MATTER, ANY QUESTION YOU CAN CONTACT US AT THE TELEPHONE NUMBER ABOVE.

SINCERELY YOURS,

We provide the following medical imaging services: Open MRI, Spiral CT-Scan, Diagnostic Ultrasound, Mammography (ACR Accredited and FDA Certified), Bone Densitometry, Fluoroscopy and Conventional Pediatric and Adult Radiology.

Vasting 10

José E. Rivera Rodriguez, M.D. Diagnostic Radiologist Diplomate American Board of Radiology

141618

NMSS/RGN1 MATERIALS-002

/lmm

PAGE Ø2



24937 Avenue Tibbitts Valencia, Callfornia 91355

An Eckert & Ziegler Company

Tel 661·309·1010 Fax 661·257·8303

NOMINAL SOURCE DATA SHEET

NUCLIDE: Na-22QUANTITY: 1ACTIVITY: 100 µCi (3.7 MBq)REFERENCE DATE: 1 Aug 06CATALOG NO.: GF-0227CAPSULE: IPL D-discCOVER: AcrylicBACKING: EpoxyACTIVE DIAMETER: 0.039" (1 mm)ACTIVE LENGTH: 0.031" (0.8 mm)NATURE OF ACTIVE DEPOSIT: Sodium Chloride in ceramic matrix.

 ADAC UGM Part:
 Source ID:

 Source-NA-0.15
 1142-82-23

LEAK TEST INFORMATION IS ON THE REVERSE SIDE.

IMPURITIES: None detected.

LAB BOOK-PAGE: 1142-82

23-Jun-06 90hn hu SIGNATURE DATE

- ISO 9001 CERTIFIED

■ .	". D	. 14	al Imaging Laboratory		•	
🖻 Isotope	e Product	S 24937	c <mark>al Imaging Laboratory</mark> Avenue Tibbitts – Valencia, C			
	Laboratorie	γς Tel 6	61 309 1010 Fax 661 257	8303		
Aur Falsaut D	7:	1800 1	trial Gauging Laboratory North Keystone Street – Burb			
An Eckert & Ziegler Company		IY Tel 6	Tel 661 309 1010 Fax 661 257 8303		x	
	Nomina		RCE DATA	SHEET		
NUCLIDE:	Cs-137		QUANTITY: 1			
ACTIVITY: 5 mCi (185 MBq)			REFERENCE DATE: 1 Aug			
CATALOG NO.: HEG-0096			CAPSULE: IPL A3015			
COVER: Stainless Steel			BACKING: Stainless Steel			
ACTIVE DI	AM.: 0.125" (3	3.18 mm)	ACTIVE LEN		' (2.29 mm)	
OVERALL [DIAM.: 0.236	" (6.0 mm)	OVERALL LE	ENGTH: 0.3	15" (8.0 mm	
NATURE O	E ACTIVE D	EPOSIT: Ce	sium Chloride in co	eramic matrix		
					<u> </u>	
	ADA	C UGM PART:	SOURCE ID:			
ć	Sc	ource-CS-5	CZ-585			
		•		·, *.		
I FAK TOOT						
,			THE REVERSE	SIDE.	• ,	
LEAK TEST			THE REVERSE			
,			THE REVERSE	SIDE.		
,			THE REVERSE			
,			THE REVERSE :			
,			THE REVERSE :			
,			THE REVERSE			
,			THE REVERSE			
,			THE REVERSE			
,			THE REVERSE		•	
,			THE REVERSE			
,	None detecte	ed.				
,	None detecte	ed.	THE REVERSE			
,	None detecte	ed.				
,	None detecte	ed.	PAGE: 1104-45			
,	None detecte	-AB BOOK-F	PAGE: 1104-45			
,	None detecte	ed.	PAGE: 1104-45			
,	None detecte	-AB BOOK-F	PAGE: 1104-45			

BEBIG Isotopen- und Medizintechnik GmbH Robert-Rössle-Str.10 Tel: +49-30-941084-0 Fax: +49-30-941084-112 .

This is to acknowledge the receipt of your letter/application dated

1/16/2008, and to inform you that the initial processing which includes an administrative review has been performed.



x

There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned Mail Control Number _	141618
When calling to inquire about this action, please refer to	
You may call us on (610) 337-5398, or 337-5260.	

NRC FORM 532 (RI) (6-96)

Sincerely, Licensing Assistance Team Leader