

**BALA MEDICAL CENTER**  
**Wayne V. Arnold, D.O., F.A.C.O.I.**

CARDIOLOGY - PERIPHERAL VASCULAR DISEASES

15 NORTH PRESIDENTIAL BLVD.  
SUITE 100  
BALA CYNWYD, PA. 19004

(610) 667 - 2746  
(610) 667- 2749  
FAX#(610) 667 - 9420

January 23, 2008

*NMSB 2*

U.S. Nuclear Regulatory Commission, Region 1  
Licensing assistance Team  
475 Allendale Road  
King of Prussia, PA 19409-1415

*03030949*

Re: License No. 37-28333-01

Dear Sir or Madam:

Please remove the following site as a facility of use from our license:

123 Whitehorse-Mercerville Road, Bldg. C, Suite 317, Mercerville, NJ 08619.

Please find attached a copy of the close-out survey revealing no removable contamination or ambient radiation levels above background. The dose calibrator QC sources (Cs-137, <200uCi; Co-57, <5 mCi) and flood source (Co-57, <10mCi) have all been transported via appropriate courier to other facilities under our license.

The facility has been released to the existing practice for general use.

Thank you for your kind attention to this matter.

If you have any questions or concerns regarding this amendment, please do not hesitate to contact me at 610-667-2746, or our Nuclear Operations Manager, Terry Day, RT, at 856-786-4700.

Sincerely,

*Wayne V. Arnold, DO*  
Wayne V. Arnold, D.O.  
Radiation Safety Officer

Enclosures

2008 JAN 24 PM 12:33  
RECEIVED  
REGION 1

*141664*

OMNI DIAGNOSTICS  
CLOSE-OUT SURVEY

FACILITY: Mercerville, NJ (Trenton)

DATE: 1-11-08

PERFORMED BY: Terry Day

SURVEY METER: Ludlum 140 188291

WELL: 15W 170520 (SN)

STANDARD SOURCE: CS-137 ACT: 0.5  $\mu$ Ci

**SURVEY RESULTS:**

AREA	BKG	MR/HR	WIPE	BKG	DPM
1.	.02	.02		261 cpm	21
2.		.02			33
3.		.02			3
4.		.02			16
5.		.02			10
6.		.02			26
7.		.02			23
8.		.02			-12
9.		.02			01
10.		.02			14
11.		.02			28
12.		.02			-2
13.		.02			15
14.		.02			07
15.		.02			14
16.		.02			0
17.		.02			04
18.		.02			-6
19.		.02			12
20.		.02			16

**RESULTS:**

NO REMOVABLE CONTAMINATION DETECTED

NO RADIATION LEVELS ABOVE BACKGROUND DETECTED

REMOVABLE CONTAMINATION DETECTED

AREA DECONTAMINATED TO BACKGROUND

FINAL SURVEY ACCEPTABLE

AREA RELEASED FOR GENERAL PUBLIC USE

RSO REVIEW

Terry Day RT, Co-RSO  
W. Ahmed, Co RSO  
1-22-08

DATE:

1-11-08

This is to acknowledge the receipt of your letter/application dated  
1/23/2008, and to inform you that the initial processing which  
includes an administrative review has been performed.

Amendo. 37-28333-01  
There were no administrative omissions. Your application was assigned to a  
technical reviewer. Please note that the technical review may identify additional  
omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

---

A copy of your action has been forwarded to our License Fee & Accounts Receivable  
Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 141664.  
When calling to inquire about this action, please refer to this control number.  
You may call us on (610) 337-5398, or 337-5260.