

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

| | | | |
|--|--|--|--|
| 1. LICENSEE/LOCATION INSPECTED: Oakland Hills Cardiology 44038 Woodward Ave., Suite 200 Bloomfield Hills, MI 43802 REPORT NUMBER(S) <u>2008-001</u> | | 2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III 2443 Warrenville Road Suite 210 Lisle, Illinois 60532-4351 | |
| 3. DOCKET NUMBER(S) 030-35911 | 4. LICENSEE NUMBER(S) 21-32366-01 | 5. DATE(S) OF INSPECTION <u>January 15, 2008</u> | |

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:


- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- ☐ 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

| Title | Printed Name | Signature | Date |
|---------------------------|-----------------------|--|------------------------------|
| LICENSEE'S REPRESENTATIVE | | | |
| NRC INSPECTOR | Robert P. Hays |  | <u>1/15/08</u> <u>RPH</u> |

**SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION**

| | | | |
|---|---|--|--|
| 1. LICENSEE Oakland Hills Cardiology REPORT NUMBER(S) 2008-001 | | 2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532 | |
| 3. DOCKET NUMBER(S) 03035911 | 4. LICENSE NUMBER(S) 21-32366-01 | 5. DATE(S) OF INSPECTION January 15, 2008 | |
| 6. INSPECTION PROCEDURES USED 87130 (10/24/02) | 7. INSPECTION FOCUS AREAS 03.01-03.07 | | |

SUPPLEMENTAL INSPECTION INFORMATION

| | | | |
|------------------------------------|-------------------------|---|--|
| 1. PROGRAM CODE(S) 02201 | 2. PRIORITY 5 | 3. LICENSEE CONTACT Peggy Passiak, CNMT | 4. TELEPHONE NUMBER 248-335-1064 |
|------------------------------------|-------------------------|---|--|

☒ Main Office Inspection Next Inspection Date: **January 2013**

☐ Field Office _____

☐ Temporary Job Site
Inspection _____

PROGRAM SCOPE

The licensee was a medical clinic authorized by the license to use any byproduct material identified in 10 CFR 35.200 for any imaging and localization procedure approved in 10 CFR 35.200 (excluding xenon-133 and aerosols) limited to cardiovascular clinical procedures at the location specified on the license. The nuclear medicine department was staffed with one nuclear medicine technologist (NMT) who routinely conducted an average of 6 patient procedures per day, Monday thru Wednesday, and Friday, each week. The licensee received Myoview® unit doses as ordered from a local nuclear pharmacy. All waste was held for decay-in-storage (DIS) or returned to the nuclear pharmacy as a limited quantity shipment. No change in NMT or RSO since the previous inspection.

Performance Observations

During the inspection, the licensee's NMT demonstrated/discussed: (1) package check-in procedures and wipe test counting; (2) dosimetry; (3) dose calibrator tests; (4) unit dose handling procedures; (5) security of licensed material; (6) radiation safety program audits; (7) rad waste procedures; and (8) contamination events and procedures.