

Subject : GAI Consultants, Inc., License Amendment , Control No. ~~141469~~

141815 mod

Dear Jenny Johansen

P-2
37-20584-01
03019945

This letter is to inform that all nuclear gauges have been removed from our prior facility, that being 6 Berkeley Road , Devon PA 1933, and that all gauges will now be stored at our new laboratory located at 470 Drew Court, King of Prussia PA 19406.

Leak tests were performed on the sealed sources for all the devices in our inventory. These results are Stated individually for each devices.

If you need any additional information , please feel free to contact me .

Respectfully yours,



Joseph P. Lowry
Radiation Safety Officer

cell # 610-761-8366



Gauge #1

North East Technical Services, Inc.
75 Aileron Court, Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

Valley Forge Labs, Inc.
1055 Westlakes Dr, Suite 200
Berwyn, PA 19312
ATTN: Tim Bowers

Shipping Address: 1055 Westlakes Dr, Suite 200
Berwyn, PA 19312

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3440	Gauge S/N	31331	Leak Test Date	1/29/2008
Source		Reading in microCurics			
	750-6016		0.00010024		
	47-1710		0.00000		

Note: 0.005 microCurics (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by:

Douglas Sims

Date:

FEB 04 2008



Gauge # 3

North East Technical Services, Inc.
75 Aileron Court, Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

Valley Forge Labs, Inc.
1055 Westlakes Dr, Suite 200
Berwyn, PA 19312
ATTN: Tim Bowers

Shipping Address: 1055 Westlakes Dr, Suite 200
Berwyn, PA 19312

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3440	Gauge S/N	28145	Leak Test Date	1/28/2008
Source					Reading in microCuries
	750-2095				0.00010024
	47-24854				0.00000

Note: 0.005 microCuries (J85 Rq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Douglas Sims

Date: FEB 04 2008



Gauge #4

North East Technical Services, Inc.
75 Aileron Court, Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

Valley Forge Labs, Inc.
1055 Westlakes Dr, Suite 200
Berwyn, PA 19312
ATTN: Tim Bowers

Shipping Address: 1055 Westlakes Dr, Suite 200
Berwyn, PA 19312

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3430	Gauge S/N	33184	Leak Test Date	1/28/2008
Source	Reading in microCuries				
750-8557	0.00007518				
47-12523	0.00000135				

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Douglas Sims

Date: FEB 04 2008



North East Technical Services, Inc.
75 Alleron Court, Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

#5

Valley Forge Labs. Inc.
1055 Westlakes Dr. Suite 200
Berwyn, PA 19312
ATTN: Tim Bowers

Shipping Address: 1055 Westlakes Dr, Suite 200
Berwyn, PA 19312

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3440	Gauge S/N	32811	Leak Test Date	1/28/2008
Source	Reading in microCuries				
750-8163	0.00007518				
47-29024	0.00000135				

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Doug L. Sims

Date: FEB 04 2008



Gauge # 6

North East Technical Services, Inc.
75 Aileron Court, Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

Valley Forge Labs, Inc.
1055 Westlakes Dr, Suite 200
Berwyn, PA 19312
ATTN: Tim Bowers

Shipping Address: 1055 Westlakes Dr, Suite 200
Berwyn, PA 19312

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3411	Gauge S/N	15835	Leak Test Date	1/28/2008
Source					
	50-4306		0.00007518		
	47-10773		0.00000135		

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Douglas Sims

Date: FEB 04 2008



North East Technical Services, Inc.
75 Allerton Ct, Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

4 7

Valley Forge Labs, Inc.
1055 Westlakes Dr, Suite 200
Brynyn, PA 19312
ATTN: Tim Bowers

*Shipping Address: 1055 Westlakes Dr, Suite 200
Brynyn, PA 19312*

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	3411	Gauge S/N	15532	Leak Test Date	1/28/2008
	Source		Reading in microCuries		
	50-4493		0.00007518		
	47-10929		0.00000135		

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: *Douglas Sims*

Date: FEB 04 2008



North East Technical Services, Inc.
75 Allerton Ct., Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

#8

Valley Forge Labs, Inc.
1055 Westlakes Dr, Suite 200
Berwyn, PA 19312
ATTN: Tim Bowers

Shipping Address: 1055 Westlakes Dr, Suite 200
Berwyn, PA 19312

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model	3411	Gauge S/N	16096	Leak Test Date	1/28/2008
Source			Reading in microCuries		
	50-5165		0.00007518		
	47-11504		0.00000135		

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Douglas Sims

Date: FEB 04 2008



#9

North East Technical Services, Inc.
75 Alleron Ct, Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

Valley Forge Labs, Inc.
1055 Westlakes Dr, Suite 200
Berwyn, PA 19312
ATTN: Tim Bowers

Shipping Address: 1055 Westlakes Dr, Suite 200
Berwyn, PA 19312

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model	3440	Gauge S/N	31529	Leak Test Date	1/28/2008
Source		Reading in microCuries			
	750-6013		0.00008413		
	47-21655		0.00000269		

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Douglas Simon

Date: FEB 04 2008

This is to acknowledge the receipt of your letter/application ~~dated~~ *undated*
Fax RECEIVED 2/14/2008, and to inform you that the initial processing which
includes an administrative review has been performed.

AMC 37-20584-01 There were no administrative omissions. Your application was assigned to a
technical reviewer. Please note that the technical review may identify additional
omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable
Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 141815.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.