

U.S. NUCLEAR REGULATORY COMMISSION**TELEPHONE CONVERSATION RECORD**

Date: 2/26/08

Time: 3:40 PM

Mail Control 141618 License No(s). 52-25632-01 Docket No(s). 03036369
or Report No(s).

Name of Licensee: Aguadilla X-Ray & Body Imaging Center

Name of Participant(s): David Rowe, Consultant
Dennis Lawyer R1, DNMS

Telephone No. 787-245-7248

Subject: RAI, Sources Request

(NOTE: This will be used as the
Documents Title in ADAMS)

Summary: Are either of the two sources requested installed in a camera? Yes, the Cs-137 source is installed in the camera. Is it used as a transmission source where the source is exposed to the patient? No, it is only a quality assurance check source and does not involve the patient. Then you may obtain the sources through 10 CFR 35.65 authorization and the license does not need to be amended.

Action Required: Send notification letter.

Document Availability: Publicly Available Non-Publicly Available
 Non-Sensitive Non-Sensitive Copyright Sensitive Sensitive Copyright
 Immediate Release Normal Release Delay Release Date

Prepared & SUNSI Review Completed By: Dennis Lawyer / RA / Date: 2/27/08