INSTRUMENT CALIBRATION RECORD

Inst. Type	Inst. Model	Asset No.	S/N
Geiger Counter	14C	003535	83561
Range	**Year**	**User Name**	
X0.1 - X1000	1991	Paul Bertetti	

Date Calibrated	Calibrated by		Calculated Readings	Actual Readings	Remarks
2-6-06	Diane Hebbs	X0.1 Kcpm	50, 200	50, 200	
		X1 Kcpm	500 2K	500, 2K	
		X10 Kcpm	5K 20K	5K, 20K	
		X100 Kcpm	50K 200K	50K, 200K	
		X1000 mR/h	1 R/hr, 500mR/hr	1 R/hr, 500mR/h	
		X100 mR/h	50 mR/h	50 mR/hr	probe
2-6-07	Diane Hebbs	X0.1 Kcpm	50, 200	50, 200	
		X1 Kcpm	500, 2K	500, 2K	
		X10 Kcpm	5K, 20K	5K, 20K	
		X100 Kcpm	50K 200K	50K 200K	
		X1000 mR/h	1 R/hr, 500mR/hr	1 R/hr 500mR/h	
		X100 mR/h	50 mR/hr	50 mR/hr	probe
2-7-08	Diane Hebbs	X0.1 Kcpm	50, 200	50, 200	
		X1 Kcpm	500, 2K	500, 2K	
		X10 Kcpm	5K, 20K	5K, 20K	
		X100 Kcpm	50K, 200K	50K, 200K	
		X1000 mR/h	1 R/hr, 500mR/hr	1 R/hr. 500mR/h	
		X100 mR/h	50 mR/hr	50 mR/hr	probe