INSTRUMENT CALIBRATION RECORD

Inst. Type	Inst. Model	Asset No.	S/N
Beta + Alpha Probe	2200 + Probes 43-65 + 43-1	005068	110912
Range	**Year**	**User Name**	
0-500K cpm	1994	Paul Bertetti	

Date Calibrated	Calibrated by		Calculated Readings	Actual Readings	Remarks
2/7/06	Diane Gibbs	X 0.1	digital 10, 40	9, 40	
			meter 100, 400	100, 380	
		X 1	100, 400	100, 400	
		X10	1000 4000	499, 4010	
		X100	10K 40K	digital 996 + 10K meter	digital 3990 + 3800 meter
		X1K	100K 400K	digital 9943 9800 meter	digital 39957 + 3800 meter
		probe 43-1	SN PR 112636	(3067) Thorium check source	
		probe 44-40	SN PR 76050	X100 (15K) Thorium check source 390	
02/06/07	Diane Gibbs	X0.1	digital 10, 40	10, 40	
			meter 100, 400	100, 390	
		X 1	100, 400	99, 400	
		X10	1000, 4000	digital 998 meter 1000	digital 4007 meter 3900
		X100	10K, 40K	digital 10,000 meter 1000	digital 3998 meter 3800
		X1K	100K 400K	9952 100K	39915 3900
		probe 43-1	SNPR 112636	(3022) Thorium Check Source	
02/07/08	Diane Gibbs	X0.1	digital 10, 40	10, 40	
			meter 100, 400	100 380	
		X 1	100, 400	digital 99, 398	meter 100 380
		X10	1000, 4000	digital 992 3981	meter 1000 3900
		X100	10K 40K	digital 9959 39865	meter 1000 4000
		X1K	100K 400K	digital 9969 39842 100K	meter 100K 38000
		~~alpha~~ probe 43-1	SN 112636	(2804) Thorium	#897H220637 Check Source