

03/10/2008 09:35 FAX 860 561 4823

IMAGING CTR W HTPD

→ HART ST

002

**MANDELL & BLAU MDs, PC**

40 Hart Street, New Britain, CT 06052

Tel. 860-229-2058

NH562

To: USNRC Region I  
Medical Licensing  
475 Allendale Road  
King of Prussia, PA 19406

March 5, 2008

Re: Termination of NRC Materials License #06-31037-01

03036916

Dear Sir or Miss:

The breast imaging practice at our office known as Imaging Center of West Hartford, with address of 1216 Farmington Ave. Suite 101, West Hartford, CT 06107 has ceased to use radioactive materials. The office did not possess any calibration sources or sealed sources and we have determined to not carry the license any longer.

We last received radioactive material for the breast imaging service in October 2007. The final area and contamination surveys were performed December 18. The following data summarizes the final survey for ambient radiation (mR/hr) levels and removable contamination from selected areas of the office.

Area	mR/hr	Wipe Test
Exam (prep) room	0.02	< 200 dpm
Imaging tray	0.02	< 200 dpm
Flooring, counter tops	0.02	< 200 dpm
Hot Lab & counter tops (including waste storage spaces)	0.02	< 200 dpm

Office unrestricted Areas:

Hallway	0.02	< 200 dpm
Bathroom	0.02	< 200 dpm

The background radiation survey ranged from 0.02 - 0.03 mR/hr, and as expected, no contamination from Tc-99m was found on site. Area survey measurements were made with a Ludlum Model 14C GM survey meter with pancake probe, and, contamination measurements were made with a Ludlum Model 2200 scaler/ratemeter with a two inch sodium iodide probe with well.

142177

UNRESTRICTED MATERIALS-002

03/11/2008 TUE 7:54 FAX 860 231 6185 W.H.S.C. Blueback Square  
03/10/2008 09:43 FAX 860 229 8495 Mandell & Blau

003/003

003

03/10/2008 09:35 FAX 860 581 4823


IMAGING CTR W HTFD

→ HART ST

003

Should you have any questions or require additional information regarding the closure of our breast imaging service, please contact me, or, Dr. Jean Weigert at 860-229-2039.  
Thank you.

Respectfully submitted,



Jeffrey S. Blau, MD  
President

This is to acknowledge the receipt of your letter/application dated

3/5/2008, and to inform you that the initial processing which includes an administrative review has been performed.

☒ 06-31037-C1 There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

☐ Please provide to this office within 30 days of your receipt of this card

---

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 142177.  
When calling to inquire about this action, please refer to this control number.  
You may call us on (610) 337-5398, or 337-5260.