

U.S. NUCLEAR REGULATORY COMMISSION**TELEPHONE CONVERSATION RECORD**

Date: 03/31/08

Time: 4:00 pm

Mail Control 142045 License No(s). 45-25425-01 Docket No(s). 030-34650
or Report No(s).

Name of Licensee: American Type Culture Collection

Name of Participant(s): Marian McKee, Ph.D.

Telephone No. 703-365-2792

Subject: Clarification on renewal request

(NOTE: This will be used as the
Documents Title in ADAMS)

Summary: The licensee has a Cs-137 self-shielded Gammator 50B Irradiator and licensed for CHIPS material. The licensee requested to place the irradiator in a storage only mode. However, the irradiator has been out of service since September 2006. I explained to the licensee they can place the irradiator on a possession only status if they had no intentions of repairing the irradiator and had intentions of discarding it. I also explained the difference between a possession only and storage only license status. Dr. McKee said she would discuss these options with the committee members.
I also indicated that for the license renewal of the CHIPS material they could follow the guidance in NUREG 1556 Vol. 7. I explained that a renewal is similar to a new license in that they need to commit to certain policies and procedures. I further explained that if they had intentions of using the irradiator they could follow the guidance in NUREG 1556 Vol 5.

Action Required: Place in ADAMS under MC142045

Document Availability:	X	Publicly Available	Non-Publicly Available	
<input checked="" type="checkbox"/> Non-Sensitive	<input type="checkbox"/>	Non-Sensitive Copyright	<input type="checkbox"/> Sensitive	<input type="checkbox"/> Sensitive Copyright
<input type="checkbox"/> Immediate Release	<input checked="" type="checkbox"/>	Normal Release	<input type="checkbox"/>	Delay Release Date

Prepared & SUNSI Review Completed By: Lizette Roldán / RA / Date: 03/31/08