



TELECOPY TRANSMITTAL

DATE: 4/7/08

TO: Bill Reichold

Fax Number: 630-515-1078

FROM: Michelle Kritzman

RE: control #316854

COMMENTS:

This transmittal consists of 43 pages (including transmittal cover sheet). Please contact me if you have any questions or problems with this transmittal.

2309 Shelby Avenue Ann Arbor, MI 48103 (734) 662-9224 Fax (734) 662-3197	70 E. 91 st Street, Suite 106 Indianapolis, IN 46240 (317) 581-1931 Fax (317) 581-1911	4806 Mile High Drive Salt Lake City, UT 84124 (801) 272-2952 Office & Fax	N7375 Crystal Ridge Drive Beaver Dam, WI 53916 (920) 885-9872 Fax (920) 885-9870
---	--	---	---

08-0008 SB

FORM 840 UNIFORM LOW-LEVEL RADIOACTIVE WASTE MANIFEST SHIPPING PAPER		ADCO SERVICES, INC.		5. SHIPPER - NAME AND FACILITY DR. MOHAMED RABIBANQOR, THE ACCOUNT OF ADCO SERVICES 823 BYRON ROAD, STE 200 HOWELL, NJ 08843			SHIPMENT NUMBER 81460			7. FORM 840 AND 848A FORM 541 AND 541A FORM 842 AND 842A ADDITIONAL INFORMATION			PAGE 1 OF 1 PAGES(1) PAGE(2) PAGE(3) PAGE(4)			8. MANIFEST NUMBER (Use this number on all continuation pages) 08-0008 SB				
1. EMERGENCY TELEPHONE NUMBER (Include Area Code) N/A				6C PERMIT N/A		SHIPMENT NUMBER 08-0008 SB		<input checked="" type="checkbox"/> COLLECTOR <input type="checkbox"/> PROCESSOR			8. CORROSION - Name and Facility Address ADCO SERVICES, INC. 794-437-3486 TORLEY PARK, IL 60477			CONTACT LEN WARBANY-FACILITY MANAGER TELEPHONE NUMBER (Include Area Code) 708-428-1660						
2. IS THIS AN 'EXCLUSIVE USE' SHIPMENT? <input checked="" type="checkbox"/> YES <input type="checkbox"/> NO				3. TOTAL NUMBER OF PACKAGES IDENTIFIED ON THIS MANIFEST 1		6. CARRIER - Name and Address ARBON EXPRESS, INC. 17640 DURNAN DRIVE TORLEY PARK, IL 60477			EPA ID NUMBER ILD64281384			SIGNATURE - Authorized carrier acknowledging waste receipt <i>Len Warbany</i>			DATE 1/14/08					
4. DOES EPA REGULATED WASTE REQUIRING A MANIFEST ACCOMPANY THIS SHIPMENT? <input checked="" type="checkbox"/> YES <input type="checkbox"/> NO If "Yes," provide Manifest Number 840-842.				EPA MANIFEST NUMBER NA		CONTACT ROBERT BRASSETT SIGNATURE - Authorized sender acknowledging waste receipt <i>Robert Brassett</i>			TELEPHONE NUMBER (Include Area Code) 708-428-3813			DATE 1-7-08			18. CERTIFICATION This is to certify that the hereinafter materials are properly classified, classified, packaged, marked, and labeled and are in proper condition for transportation according to the applicable regulations of the Department of Transportation. This also certifies that the materials are classified, packaged, marked, and labeled and are in proper condition for transportation and disposal as described in accordance with the requirements of 10 CFR Parts 20 and 61, or equivalent state regulations.					
11. U.S. DEPARTMENT OF TRANSPORTATION DESCRIPTION (including proper shipping name, hazard class, UN ID number, and any additional instructions)				12. DOT LABEL "RADIOACTIVE"		13. WAREHOUSE INDEX		14. PHYSICAL AND CHEMICAL FORM		15. INDIVIDUAL RADIONUCLIDES			16. TOTAL PACKAGE ACTIVITY MBq		17. LEAKAGE CLASS		18. TOTAL WEIGHT OR VOLUME (Use appropriate Units)		19. IDENTIFICATION NUMBER OF PACKAGE	
Radioactive Material, Type A Package, 7, UN2818 ERG1P183				I		NA		Solid SEALED SOURCES		Cs-137 Cs-137 Cs-137			5.3148E+01 1.4307E+06		NA		26 LBS; 1.6 FT3		08-0008-01	
FOR LEADERS USE ONLY				20. Check appropriate items: <input type="checkbox"/> Container represents and warrants that all data set forth in this Uniform Low-Level Radioactive Material Manifest is true and correct to the best of my knowledge. <input type="checkbox"/> Packages listed as "Limited Quantity of Radioactive Material" on this manifest conform to the conditions and limitations specified in 49 CFR 173.401 for radioactive materials transported in limited quantity of a sealed UN2810. <input type="checkbox"/> Packages listed as "NON-REGULATED MATERIAL" on this manifest are classified in accordance with 49 CFR 172.405 (Definition of Radioactive Material). These materials must still be disposed of at a licensed facility.																

FORM 840 (10-05)

04/08/08 TUE 10:33 FAX 7346629224 /BR4299/55

MPC, Inc. ADCO

PAGE 02 002

Medical Physics Consultants, Inc.

Sealed Source Leak Test

Licensee: Mohammed Rabbanl, M.D., P.C.

Date: 10/27/07

Performed by: Michelle Kritzman

Nuclide	Type	Calibration Activity	Calibration Date	Location	M/N	S/N
Cs-137	Vial	222 uCi	11/30/1978	Hot Lab	NES-356	3560878C05
Ba-133	Vial	276.3 uCi	5/1/1984	Hot Lab	BDC.VI	2297MA
Co-57	Flood	10 mCi	8/1/2006	Hot Lab	NES8400	1187-073

Comment: The sources listed above were leak tested using a dry wipe technique and were found to have less than 0.005 uCi removable activity. The following Minimum Detectable Activities are based upon a background at the indicated value. Background was at or below these levels when the above tests were completed.

Well Counter: Captus 3000

Nuclide	MDA	Background
Cs-137	2.0×10^{-4} uCi	145 counts/1 min
Ba-133	8.8×10^{-5} uCi	204 counts/1 min
Co-57	1.7×10^{-5} uCi	53 counts/1 min

RADIATION SAFETY OFFICER: _____