



GILES



ENGINEERING ASSOCIATES, INC.

- 3990 Flowers Road, Suite 530 / Atlanta, GA 30360 / 770-458-3399 / FAX: 770-458-3998
- 10553 Olympic Drive, Suite 102 / Dallas, TX 75220 / 214-358-5885 / FAX: 214-358-5884
- 1521 E. Orangethorpe Avenue, Suite B / Fullerton, CA 92831 / 714/879-8436 / FAX: 714-879-4582
- N8 W22350 Johnson Drive, Suite A1 / Waukesha, WI 53186 / 262-544-0118 / FAX: 262-549-5868
- 514 Progress Drive, Suite J / Linthicum, MD 21090 / 410-636-9320 / FAX: 410-636-9325
- 4155 St. Johns Parkway, Suite 1200 / Sanford, FL 32771 / 407-321-5356 / FAX: 407-321-6604
- Corporate Office & Accounting / 262-544-0118 / FAX: 262-544-0501

TO:	LOREN HUBER	FROM:	CHARLIE GRESSER
FAX NO:	630-515-1078	DATE/TIME:	4/22/08
REFERENCE:	LEAK TESTS		

THIS TRANSMISSION WILL CONSIST OF ~~4~~¹⁵ PAGE(S) (INCLUDING THIS SHEET)
 IF THERE ARE ANY PROBLEMS WITH THIS TRANSMISSION,
 PLEASE CONTACT US AT THE ABOVE REFERENCED OFFICE

MESSAGE:

CONTROL # 316973

The information contained in this fax message is intended only for the personal and confidential use of the designated recipient named above. This message may be an attorney/client communication, and as such, is privileged and confidential. If the reader of this message is not the intended recipient or an agent responsible for delivering it to the intended recipient, you are hereby notified that you have received this document in error, and that any review, dissemination, distribution or copying of this message is strictly prohibited. If you have received this communication in error, please notify us immediately by telephone and return the original message to us by mail.

**GILES ENGINEERING ASSOCIATES, INC.
NUCLEAR DENSITY GAUGE INVENTORY***

Type: CPN Nuclear Density Gauge


Nuclear Material: Cesium 137 - 10mCi
 Americium 241/BE - 50mCi

No.	Model	Serial No.	Mfg. Date	Last Calibration	Leak Test Done	Lk/TST Due
1	MC - 3	M370703817	1997	2/9/07	02-Feb-08	02-Feb-09
2	MC - 1DR-P	MD60703306	1996	11/7/06	02-Jan-08	02-Jan-09
3	MC - 1DR	MD9089039	1989	2/9/07	02-Feb-08	02-Feb-09
4	MC - 1DR	MD20500899	1992	5/8/06	26-Apr-07	26-Apr-08
5	MC - 1DR	MD30701544	1993	1/29/07	15-Jan-08	15-Jan-09
6	MC - 1DR	MD30701545	1993	2/12/07	04-Feb-08	04-Feb-09
7	MC-1DR-P	M341202389	1994	2/12/07	04-Feb-08	04-Feb-09
8	MC - 1DR	MD50502625	1995	2/18/08	11-Feb-08	11-Feb-09
9	MC - 1DR-P	MD71103991	1997	2/13/08	11-Feb-08	11-Feb-09
10	MC - 3	M370503723	1997	3/1/07	04-Feb-08	04-Feb-09
11	MC-1DR-P	MD80404216	1998	5/8/06	11-Mar-08	11-Mar-09
12	MC - 1DR	MD60703327	1996	2/28/07	04-Feb-08	04-Feb-09
13	MC-1DR	MD20506604	2002	1/29/07	15-Jan-08	15-Jan-09

By RSO: Charles S. Gresser P.E.

* All gauges (unless otherwise noted) are stored at N8 W22350 Johnson Drive, Suite A1, Waukesha WI 53186

CS 4/21/08



Richard Aguinaga
Aguinaga Technical Services
1924-G Calumet Drive
Sheboygan, WI 53081
Phone: (920) 803-8789
Fax: (920) 803-6707

Where Customer Service is a Priority!

LEAK TEST CERTIFICATE

Attn: Charles Gresser
Giles Engineering Associates, Inc.
N8W22350 Johnson Drive, STE A-1
Waukesha WI 53186

A leak test has been performed on the following gauge:

Manufacturer: **CPN International**
Model: **MC3**
Serial Number: **M370703817**
Cs137 Serial Number: **C-3817**
Am241/Be Serial Number: **A-3817**

Date Sample Collected: **2/2/2008**

Date Sample Analyzed: **2/12/2008**

Analyzed By: **Richard Aguinaga**

Sample Activity α : **< .000004 μ Ci**

Sample Activity β : **< .00005 μ Ci**


A sample activity of 0.005 μ Ci or greater is considered by Federal and most Agreement state regulations to be a leaking source and must be removed from service and reported to the applicable regulatory agency within five days of the test.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by WI License 117-1005-01


Richard Aguinaga
Radiation Safety Officer

REVIEWED BY RADIATION SAFETY OFFICER
(if required by your license sign below)



Richard Aguinaga
Aguinaga Technical Services
1924-C Calumet Drive
Sheboygan, WI 53081
Phone: (920) 803-8789
Fax: (920) 803-6707

Where Customer Service is a Priority!

LEAK TEST CERTIFICATE

Attn: Charles Gresser
Giles Engineering Associates, Inc.
N8W22350 Johnson Drive, STE A-1
Waukesha WI 53186

A leak test has been performed on the following gauge:

Manufacturer: CPN International
Model: MC1DRP
Serial Number: MD9089039
Cs137 Serial Number: C-9039
Am241/Be Serial Number: A-9039

Date Sample Collected: 2/2/2008

Date Sample Analyzed: 2/12/2008

Analyzed By: Richard Aguinaga

Sample Activity α : < .000004 μ Ci

Sample Activity β : < .00005 μ Ci


A sample activity of 0.005 μ Ci or greater is considered by Federal and most Agreement state regulations to be a leaking source and must be removed from service and reported to the applicable regulatory agency within five days of the test.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by WI License 117-1005-01


Richard Aguinaga
Radiation Safety Officer

REVIEWED BY RADIATION SAFETY OFFICER
(if required by your license sign below)



Richard Aguinaga
Aguinaga Technical Services
1924-C Calumet Drive
Sheboygan, WI 53081
Phone: (920) 803-8789
Fax: (920) 803-6707

Where Customer Service is a Priority!

LEAK TEST CERTIFICATE

Attn: Charles Gresser
Giles Engineering Associates, Inc.
N8W22350 Johnson Drive, STE A-1
Waukesha WI 53186

A leak test has been performed on the following gauge:

Manufacturer: CPN International

Model: MC1DR

Serial Number: MD20500899

Cs137 Serial Number: C-0899

Am241/Be Serial Number: A-0899

Date Sample Collected: 4/26/2007

Date Sample Analyzed: 5/7/2007

Analyzed By: Richard Aguinaga

Sample Activity α : < .000004 μ Ci


Sample Activity β : < .00005 μ Ci

A sample activity of 0.005 μ Ci or greater is considered by Federal and most Agreement state regulations to be a leaking source and must be removed from service and reported to the applicable regulatory agency within five days of the test.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by WI License 117-1005-01


Richard Aguinaga
Radiation Safety Officer



Richard Aguinaga
Aguinaga Technical Services
1924-G Galumet Drive
Sheboygan, WI 53081
Phone: (920) 803-8789
Fax: (920) 803-6707

Where Customer Service is a Priority!

LEAK TEST CERTIFICATE

Attn: Charles Gresser
Giles Engineering Associates, Inc.
N8W22350 Johnson Drive, STE A-1
Waukesha WI 53186

A leak test has been performed on the following gauge:

Manufacturer: **CPN International**

Model: **MC3**

Serial Number: **M370503723**

Cs137 Serial Number: **C-3723**

Am241/Be Serial Number: **A-3723**

Date Sample Collected: **2/4/2008**

Date Sample Analyzed: **2/12/2008**

Analyzed By: **Richard Aguinaga**

Sample Activity α : **< .000004 μ Ci**

Sample Activity β : **< .00005 μ Ci**


A sample activity of 0.005 μ Ci or greater is considered by Federal and most Agreement state regulations to be a leaking source and must be removed from service and reported to the applicable regulatory agency within five days of the test.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by WI License 117-1005-01


Richard Aguinaga
Radiation Safety Officer

REVIEWED BY RADIATION SAFETY OFFICER
(if required by your license sign below)



Richard Aguinaga
Aguinaga Technical Services
1924-C Calumet Drive
Sheboygan, WI 53081
Phone: (920) 803-8789
Fax: (920) 803-6707

Where Customer Service is a Priority!

LEAK TEST CERTIFICATE

Attn: Charles Gresser
Giles Engineering Associates, Inc.
N8W22350 Johnson Drive, STE A-1
Waukesha WI 53186

A leak test has been performed on the following gauge:

Manufacturer: CPN International

Model: MC1DR

Serial Number: MD60703306

Cs137 Serial Number: C-3306

Am241/Be Serial Number: A-3306

Date Sample Collected: 1/2/2008

Date Sample Analyzed: 1/8/2008

Analyzed By: Richard Aguinaga

Sample Activity α : $< .000004\mu\text{Ci}$

Sample Activity β : $< .00005\mu\text{Ci}$


A sample activity of $0.005\ \mu\text{Ci}$ or greater is considered by Federal and most Agreement state regulations to be a leaking source and must be removed from service and reported to the applicable regulatory agency within five days of the test.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by WI License 117-1005-01


Richard Aguinaga
Radiation Safety Officer

REVIEWED BY RADIATION SAFETY OFFICER
(if required by your license sign below)



Richard Aguinaga
Aguinaga Technical Services
1924-C Calumet Drive
Sheboygan, WI 53081
Phone: (920) 803-8789
Fax: (920) 803-6707

Where Customer Service is a Priority!

LEAK TEST CERTIFICATE

Attn: Charles Gresser
Giles Engineering Associates, Inc.
N8W22350 Johnson Drive, STE A-1
Waukesha WI 53186

A leak test has been performed on the following gauge:

Manufacturer: CPN International

Model: MC1DRP

Serial Number: MD60703327

Cs137 Serial Number: C-3327

Am241/Be Serial Number: A-3327

Date Sample Collected: 2/4/2008

Date Sample Analyzed: 2/12/2008

Analyzed By: Richard Aguinaga

Sample Activity α : < .000004 μ Ci

Sample Activity β : < .00005 μ Ci


A sample activity of 0.005 μ Ci or greater is considered by Federal and most Agreement state regulations to be a leaking source and must be removed from service and reported to the applicable regulatory agency within five days of the test.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by WI License 117-1005-01


Richard Aguinaga
Radiation Safety Officer

REVIEWED BY RADIATION SAFETY OFFICER
(if required by your license sign below)



RECEIVED
JAN 8 2008

Richard Aguinaga
Aguinaga Technical Services
1924-C Calumet Drive
Sheboygan, WI 53081
Phone: (920) 803-8789
Fax: (920) 803-6707

Where Customer Service is a Priority!

LEAK TEST CERTIFICATE

Attn: Charles Grosser
Giles Engineering Associates, Inc.
N8W22350 Johnson Drive, STE A-1
Waukesha WI 53186

A leak test has been performed on the following gauge:

Manufacturer: CPN International

Model: MC1DRP

Serial Number: MD20506604

Cs137 Serial Number: C-6604

Am241/Be Serial Number: A-6604

Date Sample Collected: 1/15/2008

Date Sample Analyzed: 1/24/2008

Analyzed By: Richard Aguinaga

Sample Activity α : < .000004 μ Ci

Sample Activity β : < .00005 μ Ci

A sample activity of 0.005 μ Ci or greater is considered by Federal and most Agreement state regulations to be a leaking source and must be removed from service and reported to the applicable regulatory agency within five days of the test.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by WI License 117-1005-01


Richard Aguinaga
Radiation Safety Officer

REVIEWED BY RADIATION SAFETY OFFICER
(if required by your license sign below)





Richard Aguinaga
Aguinaga Technical Services
1924-C Calumet Drive
Sheboygan, WI 53081
Phone: (920) 803-8789
Fax: (920) 803-6707

Where Customer Service is a Priority!

LEAK TEST CERTIFICATE

Attn: Charles Gresser
Giles Engineering Associates, Inc.
N8W22350 Johnson Drive, STE A-1
Waukesha WI 53186

A leak test has been performed on the following gauge:

Manufacturer: **CPN International**
Model: **MC1DRP**
Serial Number: **MD71103991**
Cs137 Serial Number: **C-3991**
Am241/Be Serial Number: **A-3991**

Date Sample Collected: **2/11/2008**
Date Sample Analyzed: **2/12/2008**
Analyzed By: **Richard Aguinaga**

Sample Activity α : **< .000004 μ Ci**

Sample Activity β : **< .00005 μ Ci**


A sample activity of 0.005 μ Ci or greater is considered by Federal and most Agreement state regulations to be a leaking source and must be removed from service and reported to the applicable regulatory agency within five days of the test.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by WI License 117-1005-01


Richard Aguinaga
Radiation Safety Officer

REVIEWED BY RADIATION SAFETY OFFICER
(if required by your license sign below)



Where Customer Service is a Priority!

Richard Aguinaga
Aguinaga Technical Services
1924-C Calumet Drive
Sheboygan, WI 53081
Phone: (920) 803-8789
Fax: (920) 803-8707

LEAK TEST CERTIFICATE

Attn: Charles Gresser
Giles Engineering Associates, Inc.
N8W22350 Johnson Drive, STE A-1
Waukesha WI 53186

A leak test has been performed on the following gauge:

Manufacturer: CPN International

Model: MC1DRP

Serial Number: MD80404216

Cs137 Serial Number: C-4216

Am241/Be Serial Number: A-4216

Date Sample Collected: 3/11/2008

Date Sample Analyzed: 3/19/2008

Analyzed By: Richard Aguinaga

Sample Activity α : < .000004 μ Ci

Sample Activity β : < .00005 μ Ci


A sample activity of 0.005 μ Ci or greater is considered by Federal and most Agreement state regulations to be a leaking source and must be removed from service and reported to the applicable regulatory agency within five days of the test.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by WI License 117-1005-01


Richard Aguinaga
Radiation Safety Officer

REVIEWED BY RADIATION SAFETY OFFICER
(if required by your license sign below)



Richard Aguinaga
Aguinaga Technical Services
1924-C Calumet Drive
Sheboygan, WI 53081
Phone: (920) 803-8789
Fax: (920) 803-6707

Where Customer Service is a Priority!

LEAK TEST CERTIFICATE

Attn: Charles Gresser
Giles Engineering Associates, Inc.
N8 W22350 Johnson Drive
Waukesha WI 53186

A leak test has been performed on the following gauge:

Manufacturer: CPN International

Model: MC3

Serial Number: M341202389

Cs137 Serial Number: C-2389

Am241/Be Serial Number: A-2389

Date Sample Collected: 2/9/2007

Date Sample Analyzed: 2/16/2007

Analyzed By: Richard Aguinaga

Sample Activity α : < .000004 μ Ci


Sample Activity β : < .00005 μ Ci

A sample activity of 0.005 μ Ci or greater is considered by Federal and most Agreement state regulations to be a leaking source and must be removed from service and reported to the applicable regulatory agency within five days of the test.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by WI License 117-1005-01


Richard Aguinaga
Radiation Safety Officer



Richard Aguinaga
Aguinaga Technical Services
1924-C Calumet Drive
Sheboygan, WI 53081
Phone: (920) 803-8789
Fax: (920) 803-6707

Where Customer Service is a Priority!

LEAK TEST CERTIFICATE

Attn: Charles Gresser
Giles Engineering Associates, Inc.
N8W22350 Johnson Drive, STE A-1
Waukesha WI 53186

A leak test has been performed on the following gauge:

Manufacturer: CPN International

Model: MC1DRP

Serial Number: MD30701544

Cs137 Serial Number: C-1544

Am241/Be Serial Number: A-1544

Date Sample Collected: 1/15/2008

Date Sample Analyzed: 1/24/2008

Analyzed By: Richard Aguinaga

Sample Activity α : < .000004 μ Ci

Sample Activity β : < .00005 μ Ci


A sample activity of 0.005 μ Ci or greater is considered by Federal and most Agreement state regulations to be a leaking source and must be removed from service and reported to the applicable regulatory agency within five days of the test.


Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by WI License 117-1005-01


Richard Aguinaga
Radiation Safety Officer

REVIEWED BY RADIATION SAFETY OFFICER
(if required by your license sign below)





Richard Aguinaga
Aguinaga Technical Services
1924-C Galumet Drive
Sheboygan, WI 53081
Phone: (920) 803-8789
Fax: (920) 803-6707

Where Customer Service is a Priority!

LEAK TEST CERTIFICATE

Attn: Charles Gresser
Giles Engineering Associates, Inc.
N8W22350 Johnson Drive, STE A-1
Waukesha WI 53186

A leak test has been performed on the following gauge:

Manufacturer: GPN International

Model: MC1DRP

Serial Number: MD30701545

Cs137 Serial Number: C-1545

Am241/Be Serial Number: A-1545

Date Sample Collected: 2/4/2008

Date Sample Analyzed: 2/12/2008

Analyzed By: Richard Aguinaga

Sample Activity α : < .000004 μ Ci

Sample Activity β : < .00005 μ Ci


A sample activity of 0.005 μ Ci or greater is considered by Federal and most Agreement state regulations to be a leaking source and must be removed from service and reported to the applicable regulatory agency within five days of the test.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by WI License 117-1005-01


Richard Aguinaga
Radiation Safety Officer

REVIEWED BY RADIATION SAFETY OFFICER
(if required by your license sign below)



Richard Aguinaga
Aguinaga Technical Services
1824-C Calumet Drive
Sheboygan, WI 53081
Phone: (920) 803-8789
Fax: (920) 803-8707

Where Customer Service is a Priority!

LEAK TEST CERTIFICATE

Attn: Charles Gresser
Giles Engineering Associates, Inc.
N8W22350 Johnson Drive, STE A-1
Waukesha WI 53186

A leak test has been performed on the following gauge:

Manufacturer: CPN International

Model: MC1DRP

Serial Number: MD50502625

Cs137 Serial Number: C-2625

Am241/Be Serial Number: A-2625

Date Sample Collected: 2/11/2008

Date Sample Analyzed: 2/12/2008

Analyzed By: Richard Aguinaga

Sample Activity α : < .000004 μ Ci

Sample Activity β : < .00005 μ Ci

A sample activity of 0.005 μ Ci or greater is considered by Federal and most Agreement state regulations to be a leaking source and must be removed from service and reported to the applicable regulatory agency within five days of the test.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by WI License 117-1005-01


Richard Aguinaga
Radiation Safety Officer

REVIEWED BY RADIATION SAFETY OFFICER
(If required by your license sign below)
