

CENTRAL CT CARDIOLOGISTS, LLC
19 Woodland St, Suite 35, Hartford, CT 06105

NM182

USNRC Region I
Commercial and R&D Branch
475 Allendale Road
King of Prussia, PA 19406

April 22, 2008

06-27909-01
03029568

Re: Amendment request for the decommissioning and closure of the licensed site at 9 Cranbrook Blvd, Enfield, CT

Dear Sir or Miss:

We have closed the 9 Cranbrook Blvd, Enfield, CT office of our nuclear cardiology imaging practice. The sealed sources were removed and transferred to the recently licensed site of 7 Elm Street, Suite 201, Enfield, CT. Wastes were allowed to decay for greater than 10 half-lives and can now be removed as normal or biomedical trash, as appropriate.

The last day for the receipt of radioactive material and patient imaging was April 17, 2008. The final area and contamination surveys were performed on April 19. The following data summarizes the final survey for ambient radiation (mR/hr) levels and removable contamination from selected areas of the closed office. The background radiation survey ranged from 0.02 - 0.03 mR/hr. The removable contamination background level (75 - 225 keV) was 76 cpm.

<u>Hot Lab Area</u>	<u>mR/hr</u>	<u>Net Wipe Test</u>
Upper cabinets	≤ 0.03	≤ 12 cpm total
Countertops	≤ 0.03	≤ 12 cpm total
Lower cabinets	≤ 0.03	≤ 12 cpm total
Co-57 flood source (8/00)	≤ 0.03	≤ 12 cpm total

<u>Imaging Room</u>	<u>mR/hr</u>	<u>Net Wipe Test</u>
Gamma Camera	≤ 0.03	≤ 12 cpm total
camera table	≤ 0.03	≤ 12 cpm total
technologist workspaces	≤ 0.03	≤ 12 cpm total
room sink	≤ 0.03	≤ 12 cpm total

<u>Stress Room</u>	<u>mR/hr</u>	<u>Net Wipe Test</u>
Treadmill	≤ 0.03	≤ 12 cpm total
Exam table	≤ 0.03	≤ 12 cpm total
EKG recorder unit	≤ 0.03	≤ 12 cpm total
room sink	≤ 0.03	≤ 12 cpm total

Please expedite this amendment request; the building owner has secured a new tenant for the space. If you have any questions, do not hesitate to contact me at 860-525-8901.

Respectfully submitted,


John T. Cardone, MD, RSO

142346

NMSS/RGN1 MATERIALS-002