

CONVERSATION RECORD

TIME | DATE  
5-08-2008

VISIT       CONFERENCE       TELEPHONE  
  
 INCOMING  
 OUTGOING

NAME OF PERSON(S) CONTACTED OR IN CONTACT	ORGANIZATION (OFFICE, DEPT. ETC.)	TELEPHONE NO.
Minh Hoac, RSO	Hercules, Incorporated Louisiana, MO	573-754-6211 ext 362 FAX -3182


**SUBJECT**  
Control No. 316967 – Re - Amendment request (NRC License No. 24-03060-01) by letter dated February 8, 2008 (with attachments), regarding the recent transfer/disposal of 9 gauges to Ohmart/Vega, a licensed and certified contractor, and the receipt of 2 replacement gauges, all containing sealed sources of Cs-137, ranging in activity from 20 millicuries to 300 millicuries each.

**SUMMARY**  
In a telephone conversation with Mr. Hoac on this date, the amendment request described above was discussed and Mr. Hoac stated or confirmed the following:

1. The licensee no longer possesses any, and is requesting deletion from the license, of Kay-Ray Model 7062BP gauges as well as Ohmart Model SR-1 (also designated as ED-12) gauges. Further, the only gauges remaining in their possession are Ohmart Models SH-F1A (one with 15 millicuries Cs-137), SHRM-PA (three with 100 millicuries Cs-137 each and three with 10 millicuries Cs-137 each), SHLG-1 (three with 50 millicuries Cs-137 each and one with 20 millicuries Cs-137), SHLG-2 (one with 20 millicuries Cs-137), and Kay-Ray Model 7062B (one with 50 millicuries of Cs-137). Although total current possession is 585 millicuries, the license requested a total possession limit of 800 millicuries to allow for replacement sources or limited expansion involving make and model of gauges already authorized on the license.
2. Also, Mr. Hoac confirmed that the licensee has never had in their possession any of these gauges which developed a leaking source, as confirmed by routine leak test analysis.
3. Mr. Hoac also confirmed that the serial number of the source listed as S96B0501 in their letter dated February 8, 2008, is incorrect and that the correct serial number is S96B2901.

**ACTION REQUIRED**  
License reviewer to utilize above information in evaluation of subject amendment request.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this conversation record will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html> (the Public Electronic Reading Room).

NAME OF PERSON DOCUMENTING CONVERSATION	SIGNATURE	DATE
Loren J. Hueter		5-08-2008