



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory
Certificate #0972-01

Certificate of Calibration

Submitted By: DIV20

Address: B51

Contact: DON BANNON

Manufacturer / Model: ORION / EA940

Description: EXPANDABLE IONANALYZER

Serial No: 4274

Asset No: 003325

Procedure: PH METERS MV ONLY - 11 SEP 07

Work Order: 303080483

Date Issued: Apr 17, 2008

Calibration Date: Apr 17, 2008

***Calibration Due:** Oct 17, 2008

Calibration Location: Bldg. 64

Environment: Temp. 70.0°F Hum. 43 %RH

****Data Type:** FOUND-LEFT

DivID/Location: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NC SL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

Remarks: Limited Cal: mVolts Only

Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
004164	6380025	FLUKE	5500A/SC300	CALIBRATOR	Aug 24, 08

Reviewed by: srk mar wgh

Measurements by: Clint Rowe
Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303080483	Mfr:	ORION	Technician:	CER
Asset No:	003325	Model:	EA940	Cal Date:	17-Apr-08
Serial No:	4274	Type:	ION ANALYZER		

Remarks:

LIMITED CAL: mVolts Only

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Ch 1	mVolts	mVolts	mVolts	mVolts	mVolts	Result
	1900.0	1900.2	0.2	1.0	0.12	Pass
	-1900.0	-1900.4	-0.4	1.0	0.12	Pass
Ch 2	mVolts	mVolts	mVolts	mVolts	mVolts	
	1900.0	1900.3	0.3	1.0	0.12	Pass
	-1900.0	-1900.4	-0.4	1.0	0.12	Pass

END OF REPORT