



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory
Certificate #0972-01

Certificate of Calibration

Submitted By: DIV20

Address: B51

Contact: DON BANNON

Manufacturer / Model: FLUKE / 54 II

Description: THERMOCOUPLE THERMOMETER

Serial No: 90810070

Asset No: 012159

Procedure: DIGITAL THERMOMETERS - 17 MAR 08

Work Order: 303080796

Date Issued: Apr 29, 2008

Calibration Date: Apr 29, 2008

***Calibration Due:** Apr 29, 2009

Calibration Location: Bldg. 64

Environment: Temp. 74.0°F Hum. 40 %RH

****Data Type:** FOUND-LEFT

DivID/Location: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

Remarks: Calibrated Type JKTE Only

Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
004164	6380025	FLUKE	5500A/SC300	CALIBRATOR	Aug 24, 08

Reviewed by: () srk (x) mar () wgh

Measurements by: Bob Trollinger
Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303080796	Mfr:	Fluke	Technician:	blt
Asset No:	012159	Model:	54 II	Cal Date:	29-Apr-08
Serial No:	90810070	Type:	Temperature Meter		

Remarks:

Calibrated Type JKTand E only.

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left	
Type J	°F	°F	°F	°F	°F	Result	
	-300	-299.5	0.5	1.1	0.57	Pass	
	110	110.0	0.0	0.6	0.31	Pass	
	525	525.1	0.1	0.8	0.37	Pass	
	940	939.9	0.1	1.0	0.37	Pass	
	1350	1350	0.0	1.2	0.37	Pass	
	°C	°C	°C	°C	°C		
	-200	-199.6	0.4	0.7	0.33	Pass	
	40	40.1	0.1	0.3	0.21	Pass	
	275	275.0	0.0	0.4	0.32	Pass	
	510	510.0	0.0	0.6	0.32	Pass	
	750	749.9	-0.1	0.7	0.32	Pass	
	Type K	°F	°F	°F	°F	°F	
		-300	-299.1	0.9	1.1	0.57	Pass
		390	390.1	0.1	0.7	0.55	Pass
1075		1075	0.0	1.0	0.55	Pass	
1760		1760	0.0	1.4	0.55	Pass	
2450		2450	0.0	1.7	0.84	Pass	
°C		°C	°C	°C	°C		
-150		-149.7	0.3	0.6	0.40	Pass	
210		210.0	0.0	0.4	0.32	Pass	
575		575.0	0.0	0.6	0.32	Pass	
940		939.9	-0.1	0.8	0.32	Pass	
1300		1300	0.0	1.0	0.47	Pass	
Type T		°F	°F	°F	°F	°F	
		-370	-369.1	0.9	2.2	0.57	Pass
		0	0.1	0.1	0.5	0.55	Pass
	212	212.1	0.1	0.6	0.55	Pass	
	392	392.0	0.0	0.7	0.55	Pass	
	730	729.9	0.1	0.9	0.84	Pass	
	°C	°C	°C	°C	°C		
	-225	-224.4	0.6	1.4	0.40	Pass	
	32	32.1	0.1	0.3	0.32	Pass	
	100	100.1	0.1	0.4	0.32	Pass	
	200	200.0	0.0	0.4	0.32	Pass	
	390	390.0	0.0	0.5	0.47	Pass	

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303080796	Mfr:	Fluke	Technician:	blt
Asset No:	012159	Model:	54 II	Cal Date:	29-Apr-08
Serial No:	90810070	Type:	Temperature Meter		

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left	
Type E	°F	°F	°F	°F	°F	Result	
	-230	-229.9	0.1	0.8	0.57	Pass	
	32	31.9	0.1	0.3	0.55	Pass	
	100	99.9	0.1	0.4	0.55	Pass	
	900	899.7	0.3	0.8	0.55	Pass	
	1800	1800	0.0	1.2	0.84	Pass	
	°C	°C	°C	°C	°C		
	-145	-144.8	0.2	0.7	0.40	Pass	
	0	0.0	0.0	0.5	0.32	Pass	
	40	40.0	0.0	0.5	0.32	Pass	
Difference	°F	°F	°F	°F	°F		
	T1-T2	0	-0.3	-0.3	1.0	.012	Pass
	°C	°C	°C	°C	°C		
	T1-T2	0	0.0	0.0	0.6	.012	Pass
	T2 Type J	°F	°F	°F	°F	°F	
-300		-299.4	0.6	1.1	0.57	Pass	
110		110.0	0.0	0.6	0.31	Pass	
525		525.1	0.1	0.8	0.37	Pass	
940		939.9	0.1	1.0	0.37	Pass	
1350		1350.0	0.0	1.2	0.37	Pass	
°C		°C	°C	°C	°C		
-200		-199.5	0.5	0.7	0.33	Pass	
40		40.2	0.2	0.3	0.21	Pass	
275		275.1	0.1	0.4	0.32	Pass	
Type K	°F	°F	°F	°F	°F		
	-300	-299.2	0.8	1.1	0.57	Pass	
	390	390.1	0.1	0.7	0.55	Pass	
	1075	1075.0	0.0	1.0	0.55	Pass	
	1760	1760.0	0.0	1.4	0.55	Pass	
	2450	2450.0	0.0	1.7	0.84	Pass	
	°C	°C	°C	°C	°C		
	-150	-149.5	0.5	0.6	0.40	Pass	
	210	210.1	0.1	0.4	0.32	Pass	
	575	575.0	0.0	0.6	0.32	Pass	
940	939.9	-0.1	0.8	0.32	Pass		
1300	1300.0	0.0	1.0	0.47	Pass		

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303080796	Mfr:	Fluke	Technician:	blt
Asset No:	012159	Model:	54 II	Cal Date:	29-Apr-08
Serial No:	90810070	Type:	Temperature Meter		

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Type T	°F	°F	°F	°F	°F	Result
	-370	-368.3	1.7	2.2	0.57	Pass
	0	0.1	0.1	0.5	0.55	Pass
	212	212.1	0.1	0.6	0.55	Pass
	392	392.1	0.1	0.7	0.55	Pass
	730	729.9	0.1	0.9	0.84	Pass
	°C	°C	°C	°C	°C	
	-225	-224.2	0.8	1.4	0.40	Pass
	32	32.2	0.2	0.3	0.32	Pass
	100	100.1	0.1	0.4	0.32	Pass
	200	200.0	0.0	0.4	0.32	Pass
	390	389.9	-0.1	0.5	0.47	Pass
Type E	°F	°F	°F	°F	°F	
	-230	-229.7	0.3	0.8	0.57	Pass
	32	32.1	0.1	0.3	0.55	Pass
	100	100.0	0.0	0.4	0.55	Pass
	900	899.9	0.1	0.8	0.55	Pass
	1800	1800.0	0.0	1.2	0.84	Pass
	°C	°C	°C	°C	°C	
	-145	-144.7	0.3	0.7	0.40	Pass
	0	0.1	0.1	0.5	0.32	Pass
	40	40.1	0.1	0.5	0.32	Pass
	500	500.0	0.0	0.8	0.32	Pass
	950	949.9	-0.1	1.0	0.47	Pass

END OF REPORT