



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory
Certificate #0972-01

Certificate of Calibration

Submitted By: DIV20

Address: B51

Contact: DON BANNON

Manufacturer / Model: FISHER SCIENTIFIC / 15-166A

Description: THERMOMETER, GLASS

Serial No: H98-179

Asset No: 007169

Procedure: THERMOMETERS, GLASS - 11 SEP 06

Work Order: 303080495

Date Issued: Apr 25, 2008

Calibration Date: Apr 25, 2008

***Calibration Due:** Apr 25, 2009

Calibration Location: Bldg. 64

Environment: Temp. 74.0°F Hum. 40 %RH

****Data Type:** FOUND-LEFT

DivID/Location: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NC SL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of $k=2$ to approximate a 95% confidence level.

Remarks: None

Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
009137	A21208	HART SCIENTIFIC	1575	SUPER THERMOMETER	Jun 17, 08
013908	1641	HART SCIENTIFIC	5628	SPRT	Feb 20, 10

Reviewed by: (✓) srk () mar () wgh

Measurements by: Bob Trollinger
Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303080495	Mfr:	Fisher-Scientific	Technician:	blt
Asset No:	007169	Model:	15-166A (ASTM 1C)		
Serial No:	H98-179	Type:	THERMOMETER	Cal Date:	25-Apr-08
Remarks:	Accuracy Reference ANSI Z236.1-1983				

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
	°C	°C	°C	°C	°C	Result
Temperature	-19.75	-20.0	-0.3	0.5	0.29	Pass
	-0.01	0.0	0.0	0.5	0.29	Pass
	50.01	49.8	-0.2	0.5	0.29	Pass
	100.03	99.5	-0.5	0.5	0.29	Pass
	150.00	149.5	-0.5	0.5	0.29	Pass
END OF REPORT						