

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Aggarwal & Associates, M.D.'s, P.C.
29610 Ryan Street
Warren, MI 48092

REPORT NUMBER(S) **2008-001**

2. NRC/REGIONAL OFFICE

U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4351

3. DOCKET NUMBER(S)
030-36037

4. LICENSEE NUMBER(S)
21-32405-01

5. DATE(S) OF INSPECTION
April 30, 2008

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:


- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- ☐ 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Robert P. Hays		4/30/08

**SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION**

1. LICENSEE

Aggarwal & Associates, M.D.'s, P.C.

REPORT

NUMBER(S) **2008-001**

2. NRC/REGIONAL OFFICE

Region III**2443 Warrenville Road, Suite 210
Lisle, IL 60532**

3. DOCKET NUMBER(S)

03036037

4. LICENSE NUMBER(S)

21-32405-01

5. DATE(S) OF INSPECTION

April 30, 2008

6. INSPECTION PROCEDURES USED

87130 (10/24/02)

7. INSPECTION FOCUS AREAS

03.01-03.07**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S)

02201

2. PRIORITY

5

3. LICENSEE CONTACT

I. Aggarwal, M.D., RSO

4. TELEPHONE NUMBER

810/751-6270

Main Office Inspection

Next Inspection Date: **April 2013**

Field Office

Temporary Job Site
Inspection**PROGRAM SCOPE**

The licensee was a cardiology medical clinic with locations as specified on the license in the Detroit, Michigan vicinity. The Warren, MI location was staffed with two nuclear medicine technologists (NMT) who rotate on Mondays and Wednesdays each week and each NMT conducts an average of 7-8 cardiac studies each day tests are scheduled. Licensed material is received from a nearby Livonia, MI pharmacy.

Performance Observations

During the inspection, the licensee's NMT demonstrated/discussed: (1) unit dose prep and safe use; (2) package check-in and return procedures; (3) wipe test counting; (4) dose calibrator tests; (5) security of license materials; (6) radiation safety program reviews; (7) surveys; (8) sealed source inventory; and (9) any minor contamination events.