

BB

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

**Community Health Center
of Branch County
274 East Chicago Street
Coldwater, MI 49036**
REPORT NUMBER(S)

2. NRC/REGIONAL OFFICE

**U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4351**

3. DOCKET NUMBER(S)
030-11390

4. LICENSEE NUMBER(S)
21-16642-01

5. DATE(S) OF INSPECTION
April 28, 2008

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Robert P. Hays		4/28/08

**SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION**

1. LICENSEE Comm. Health. Ctr. of Branch Co.		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532	
REPORT NUMBER(S) 2008-001			
3. DOCKET NUMBER(S) 03011390	4. LICENSE NUMBER(S) 21-16642-01	5. DATE(S) OF INSPECTION April 28, 2008	
6. INSPECTION PROCEDURES USED 87130 (10/24/02))		7. INSPECTION FOCUS AREAS 03.01-03.07	

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02121	2. PRIORITY 5	3. LICENSEE CONTACT Linda Brown, CNMT, RSO	4. TELEPHONE NUMBER 517-279-5441
------------------------------------	-------------------------	--	--

Main Office Inspection Next Inspection Date: **April 2013**

Field Office _____

Temporary Job Site Inspection _____

PROGRAM SCOPE

The licensee was a medical institution staffed with one nuclear medicine technologist (NMT) who also serves as the licensee's radiation safety officer. Licensee staff also includes one NMT in training who assists the RSO with diagnostic studies that average 10 diagnostic patient studies per day, Monday-Friday, with an occasional weekend scan when needed. Licensed material is delivered by a nearby Ft. Wayne, IN nuclear pharmacy. No change in NMT since the previous inspection.

Performance Observations

During the inspection, the licensee's NMTs demonstrated/discussed: (1) dose prep and safe use; (2) package check-in procedures and wipe test counting; (3) dosimetry; (4) dose calibrator tests; (5) security of license materials; (6) delivery and return transportation procedures; (7) radiation safety program reviews; and (8) minor contamination events.