

BB

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Med-Share Incorporated
26222 Telegraph Road
Suite 100
Southfield, MI 48034
REPORT NUMBER(S)

2. NRC/REGIONAL OFFICE

U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4351

3. DOCKET NUMBER(S)
030-31441

4. LICENSEE NUMBER(S)
21-26098-01

5. DATE(S) OF INSPECTION
April 30 - May 1, 2008

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:


- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Robert P. Hays		5/1/08

**SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION**

1. LICENSEE Med-Share Incorporated		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532	
REPORT NUMBER(S) 2008-001			
3. DOCKET NUMBER(S) 03031441	4. LICENSE NUMBER(S) 21-26098-01	5. DATE(S) OF INSPECTION April 30-May 1, 2008	
6. INSPECTION PROCEDURES USED 87130 (10/24/02)	7. INSPECTION FOCUS AREAS 03.01-03.07		

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02201	2. PRIORITY 5	3. LICENSEE CONTACT Pam Bradford, RSO	4. TELEPHONE NUMBER 248-335-1064
------------------------------------	-------------------------	---	--

Main Office Inspection Next Inspection Date: **May 2013**

Field Office **3363 Fort Street, Wyandotte, MI; 20905 Greenfield, Ste. 105, Southfield, MI**

Temporary Job Site Inspection

PROGRAM SCOPE

The licensee was a multisite medical provider authorized by the license to use any byproduct material permitted by 10 CFR 35.100 and 35.200,(excluding xenon-133) at the locations specified on the license. Each nuclear medicine suite was staffed with one or two assigned nuclear medicine technologists (NMTs) who routinely conducted an average of 6 patient procedures depending on the day studies are scheduled. Monday thru Friday each week. The licensee also services 37 medical locations in the Michigan area with four mobile medical vans using unit doses only. The licensee receives unit doses delivered directly to each location specified on the license. All waste was held for decay-in-storage (DIS) or returned to the unit dose prep facility as a limited quantity shipment for pickup by the vendor. The RSO performs all quarterly and annual testing of nuclear medicine equipment as part of her RSO duties.

Performance Observations

During the inspection, the licensee's NMTs demonstrated/discussed: (1) package check-in procedures and wipe test counting; (2) dosimetry and fetal monitoring; (3) dose calibrator tests; (4) unit dose handling procedures; (5) security of licensed material; (6) radiation safety program audits; (7) rad waste procedures; (8) contamination events and procedures; (9) transportation procedures and (10) DOT requirements including hazmat training.