U.S. NUCLEAR REGULATORY COMMISSION

TELEPHONE CONVERSATION RECORD

Date: June 13 and 17, 2008

Time: Various

Mail Control or Report No(s). 2008-001 License No(s). 29-30077-01 Docket No(s). 03033306

Name of Licensee:	Datachem, Incorporated
Name of Participant(s):	James Menoutis, Vice President Steven R. Courtemanche, Health Physicist NRC RI
Telephone No.	732-248-3335 610-948-2161
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	Priority 5T Inspection
Summary:	6/13/2008 13:10 I was informed that he was at lunch and I left a voice-mail for him to get back in touch with me. 6/17/2008 10:15 I was informed that he was in a meeting and I left a message with the receptionist for him to get back in touch with me. 14:20 I spoke with Mr. Menoutis about his licensed program. He stated that he still possesses the two Tracor Model 540 gas chromatographs and that both have been put into storage as of May 10, 2003. They were last leak tested on May 2, 2003 by the licensee using the manufacturer's leak test kit and analyzed by the manufacturer. The licensee intends to transfer one unit to the manufacturer for disposal and converted the second unit to the manufacturer's general license program in 2006. When the first unit is transferred to the manufacturer and documentation is received, the licensee shall begin the termination process for the license.
Action Required:	Place in ADAMS

Document Availability: X **Publicly Available** ____ Non-Publicly Available

X Non-Sensitive ____ Non-Sensitive Copyright ____ Sensitive ____ Sensitive Copyright

____ Immediate Release X Normal Release ____ Delay Release Date

Prepared & SUNSI Review Completed By: Steven R. Courtemanche ***/ RA /*** Date: June 17, 2008