

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

<p>1. LICENSEE/LOCATION INSPECTED: <i>Road Commission for Oakland County</i> <i>Waterford, Michigan 48328</i></p> <p>REPORT NUMBER(S) <i>2008-001</i></p>	<p>2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III 2443 Warrenville Road Suite 210 Lisle, Illinois 60532-4351</p>
-------------------------------------------------------------------------------------------------------------------------------------------------------------------	-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------


<p>3. DOCKET NUMBER(S) <i>030-09514</i></p>	<p>4. LICENSEE NUMBER(S) <i>21-15646-01</i></p>	<p>5. DATE(S) OF INSPECTION <i>June 5, 2008</i></p>
-------------------------------------------------	-----------------------------------------------------	---------------------------------------------------------

LICENSEE:
The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)



Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Deborah A. Piskura	<i>Deborah A. Piskura</i>	<i>6/5/08</i>

**SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION**

1. LICENSEE Road Commission for Oakland County REPORT NUMBER(S) 2008-001		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-09514	4. LICENSE NUMBER(S) 21-15646-01	5. DATE(S) OF INSPECTION June 5, 2008	
6. INSPECTION PROCEDURES USED 87124,		7. INSPECTION FOCUS AREAS 03.01 – 03.08	

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 03121	2. PRIORITY 5	3. LICENSEE CONTACT Kenneth Ballard, RSO	4. TELEPHONE NUMBER 248-858-4864
-----------------------------	------------------	---------------------------------------------	-------------------------------------

Main Office Inspection Next Inspection Date: June 2013

Field Office _____

Temporary Job Site Inspection _____

PROGRAM SCOPE

The licensee was large municipal office employing 400 individuals. The licensee possessed four moisture/density gauges, two of which were used daily for road testing projects during the construction season. The licensee approved 6 individuals (all successfully completed the manufacturer's training course) as authorized gauge users. The licensee was authorized by License Condition 18. to remove the Cs-137 source rod for cleaning and maintenance.

During this inspection, the inspector toured the gauge storage area and surveyed the gauge storage room, interviewed select licensee personnel, and reviewed select records. Licensee personnel discussed and demonstrated security measures in the field, transportation procedures and emergency procedures. At the time of this inspection the gauge was not in use due to inclement weather. Therefore, the inspector was unable to conduct a field inspection.