

INGHAM

REGIONAL MEDICAL CENTER

A McLAREN HEALTH SERVICE

June 18, 2008

United States Nuclear Regulatory Commission
Region III, Office of Materials Licensing
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

RE: Amendment to NRC License No. 21-04073-01
Ingham Regional Medical Center

Dear Sir/Madam:

The purpose of this letter is to amend our current NRC license to reflect the following changes.

Item #1

We previously requested and received an amendment for re-location of our department. Enclosed are the results of the close-out survey of our Nuclear Medicine department at the above location.

A close-out survey was performed on June 4, 2008 of the above address and is enclosed for your review.


The last date of receipt and use of radioactive material at this site was May 19, 2008 for Tc-99m. Other radioactive materials used at this facility include, I-131 therapy capsule January 15, 2008, In-111 on March 4, 2008, Ga-67 on April 22, 2008, I-123 on May 7, 2008, and Y-90 on March 11, 2008. No other radioactive materials used at this facility.

A current copy of the leak test results for the sealed sources used at our facility is enclosed for your review. These sources include a Co-57 flood source, Cs-137, and Ba-133 vials for constancy and accuracy testing. These sources were removed to the new department prior to performing the close-out.

We did not dispose of radioactive material by release to sewers or incineration. In addition, there were no spills of any long-lived radioactive material or on-site burials of radioactive materials.

Thank you for your cooperation. If you have any questions or require additional information, please contact our physics consultant, Kevin B. Miller at 734-662-3197.

Respectfully Yours,



Patrick Salow
IROH Administrator
Ingham Regional Medical Center

Close-out Survey

Date performed: June 4, 2008

Performed by: Kevin B. Miller, Medical Physics Consultants

Comments: Sealed sources and radioactive trash were transferred to our new location prior to performing the close-out survey.

Instruments

Wipe tests analyzed with a Captus 3000 Well Counter S/N#900271.

Instrument: Well Counter

Radionuclide: Full Spectrum MDA for Full Spectrum: 261dpm

Date completed: 06/06/08 Conversion factor: 2.38 cpm/dpm

Efficiency=40%

Area survey performed with the following survey meter:

Manufacturer: Ludlum

Type: GM

Model Number: 14C

Serial Number: 81890

Probe Model: End Window

Annual Calibration Due: 03/19/09

Battery check acceptable: YES

Operational check acceptable: YES

Current reading: 7.5 mR/hr

History of Radionuclides Used

Unsealed: The last date of receipt and use of radioactive material at this site was May 19, 2008 for Tc-99m. Other radioactive materials used at this facility include, I-131 therapy capsule January 15, 2008, In-111 on March 4, 2008, Ga-67 in April 22, 2008, I-123 on May 7, 2008, and Y-90 on March 11, 2008. No other radioactive materials used at this facility.

Sealed

Cs-137 Vial, Ba-133 Vial, and Co-57 Flood last used May 19, 2008.

Last date of byproduct radiopharmaceutical use: May 19, 2008.

Close-out Survey - continued

Visual Check: The area was checked to ensure that all sealed sources and radioactive waste had been removed. No evidence of radioactive material was noted.

Sealed Source Leak Testing: No history of leaks from Sealed Sources.



Radiation Level Survey: No area within the department demonstrated radiation levels in excess of the background reading of 0.02 mR/hr.

Removable Contamination: No area within the department demonstrated removable contamination in excess of 60 dpm's. There is no history of any spills of long-lived radioisotopes.

Conclusion: No radioactive materials remain in this department. No removable contamination is present.

Ingham Regional Medical Center
401 W. Greenlawn Avenue
Lansing, Michigan 48910-2819

Old Nuclear Medicine Department



Removable Contamination

Hot Lab

Ingham Regional Medical Center

License # 21-04073-01

Date: 6/4/2008

Performed By: Kevin B. Miller

Well Counter: Captus 3000

MDA: 261 dpm

Survey Meter: Ludlum 14C GM

Check Value: 7.5 mR/hr

Calibrated 03/19/08

		cpm	net dpm	mR/h
1	Hot Lab Fume Hood	481.00	0	0.02
2	Hot Lab Fume Hood	485.00	0	0.02
3	Hot Lab Dose Prep	529.00	60	0.02
4	Hot Lab Dose Prep	529.00	60	0.02
5	Hot Lab Dose Prep	529.00	60	0.02
6	Sealed Source Storage	485.00	0	0.02
7	Sealed Source Storage	485.00	0	0.02
8	Hot Lab Floor	485.00	0	0.02
9	Hot Lab Floor	487.00	0	0.02
10	Hot Lab Floor	487.00	0	0.02
11	Hot Lab Floor	487.00	0	0.02
12	Hot Lab Floor	497.00	0	0.02
13	Hot Lab Floor	497.00	0	0.02
14	Hot Lab Waste Storage	497.00	0	0.02
15	Hot Lab Waste Storage	494.00	0	0.02
16	Waste Storage Floor	494.00	0	0.02
17	Waste Storage Floor	525.00	50	0.02
18	Waste Storage Floor	525.00	50	0.02
19	Waste Storage Floor	525.00	50	0.02
20	Waste Storage Floor	494.00	0	0.02
	Background	505		0.02

ACTION LEVELS: 2000 dpm

**Efficiency
(dpm/cpm):**

2.5

< 0.2 mR/hr

COMMENTS: No evidence of removable contamination.

RSO Signature: Byron G. Tollenear 6/17/08

Maximum removable contamination:

60.0 net dpm/100 cm²

Camera Room 1

Ingham Regional Medical Center				
License # 21-04073-01				
Date:		6/4/2008		Performed By: Kevin B. Miller
Well Counter:		Captus 3000		MDA: 261 dpm
Survey Meter:		Ludlum 14C EWGM		Check Value: 7.5 mR/hr
Calibrated 03/19/08				
		cpm	net dpm	mR/h
1	Rm 1 Counter	498.00	0	0.02
2	Rm 1 Floor	498.00	0	0.02
3	Rm 1 Floor	498.00	0	0.02
4	Rm 1 Floor	495.00	0	0.02
5	Rm 1 Floor	495.00	0	0.02
6	Rm 1 Floor	490.00	0	0.02
7	Rm 1 Floor	490.00	0	0.02
8	Rm 1 Floor	495.00	0	0.02
9	Rm 1 Floor	490.00	0	0.02
10	Rm 1 Floor	490.00	0	0.02
	Background	505		0.02
ACTION LEVELS: 2000 dpm Efficiency (dpm/cpm): 2.5 < 0.2 mR/hr				
COMMENTS: No evidence of removable contamination.				
RSO Signature: <u>Bryan G. Tolleran 6/17/08</u>				

Maximum removable contamination: 0.0 net dpm/100 cm²

Camera Room 2

Maximum removable contamination: 112.5 net dpm/100 cm²

Camera Room 3

Maximum removable contamination: 25.0 net dpm/100 cm²

Removable Contamination

Injection/Stress Room

Ingham Regional Medical Center				
License # 21-04073-01				
Date: 6/4/2008		Performed By: Kevin B. Miller		
Well Counter: Captus 3000		MDA: 261 dpm		
Survey Meter: Ludlum 14C EWGM		Check Value: 7.5 mR/hr		
Calibrated 03/19/08				
		cpm	net dpm	mR/h
1	Counter/Bed	508.00	7.5	0.02
2	Update/Inj. Floor	508.00	7.5	0.02
3	Update/Inj. Floor	508.00	7.5	0.02
4	Update/Inj. Floor	487.00	0	0.02
5	Update/Inj. Floor	487.00	0	0.02
6	Update/Inj. Floor	487.00	0	0.02
7				
8				
9				
10				
	Background	505		0.02
ACTION LEVELS: 2000 dpm Efficiency (dpm/cpm): 2.5 < 0.2 mR/hr				
COMMENTS: No evidence of removable contamination.				
RSO Signature: <u>Benjamin G. Tollenon 6/17/08</u>				

Maximum removable contamination: **7.5 net dpm/100 cm²**

Conclusion

As of June 4, 2008 all radioactive materials were removed from the site and no removable contamination was present.

Medical Physics Consultants, Inc.

Sealed Source Leak Test

Licensee: Ingham Reg'l. Med. Cntr.-Greenlawn Campus

Date: 06/04/08

Performed by:

Kevin Miller

Nuclide	Type	Calibration Activity	Calibration Date	Location	M/N	S/N
Cs-137	Vial	1000 uCi	01/01/72	Hot Lab	MALLIN	1LK
Current Activity: 430.9 uCi						
Ba-133	Vial	270 uCi	06/06/95	Hot Lab	NES-358	358019035
Current Activity: 116.7 uCi						
Co-57	FF	10 mCi	10/01/07	Hot Lab	PF24R-057-10M	1254-151
Current Activity: 5.316 mCi						

Comment: The sources listed above were leak tested using a dry wipe technique and were found to have less than 0.005 uCi removable activity. The following Minimum Detectable Activities are based upon a background at the indicated value. Background was at or below these levels when the above tests were completed.

Well Counter: Captus 3000

Nuclide	MDA	Background
Cs-137	2.0×10^{-4} uCi	145 counts/1 min
Ba-133	8.8×10^{-5} uCi	204 counts/1 min
Co-57	1.7×10^{-5} uCi	53 counts/1 min

RADIATION SAFETY OFFICER:

Bryan G. Tolleran 6/17/08

Medical Physics Consultants, Inc.

Sealed Source Leak Test

Licensee: Ingham Reg'l. Med. Cntr.-Greenlawn Campus

Date: 06/04/08

Performed by:

Kevin Miller

Nuclide	Type	Calibration Activity	Calibration Date	Location	M/N	S/N
Cs-137	Vial	1000 uCi	01/01/72	Hot Lab	MALLIN	1LK
Current Activity: 430.9 uCi						
Ba-133	Vial	270 uCi	06/06/95	Hot Lab	NES-358	358019035
Current Activity: 116.7 uCi						
Co-57	FF	10 mCi	10/01/07	Hot Lab	PF24R-057-10M	1254-151
Current Activity: 5.316 mCi						

Comment: The sources listed above were leak tested using a dry wipe technique and were found to have less than 0.005 uCi removable activity. The following Minimum Detectable Activities are based upon a background at the indicated value. Background was at or below these levels when the above tests were completed.

Well Counter: Captus 3000

Nuclide	MDA	Background
Cs-137	2.0×10^{-4} uCi	145 counts/1 min
Ba-133	8.8×10^{-5} uCi	204 counts/1 min
Co-57	1.7×10^{-5} uCi	53 counts/1 min

RADIATION SAFETY OFFICER: Royce G. Tolleran 6/17/08

Radiation Detection Efficiency

Licensee: Medical Physics Consultants, Inc.

Date: 06/06/08

Instrument: Well

Model: Captus 3000

Window: Full Spectrum

Serial #: 900271

Source Activity: 0.4539 uCi of Cs-137
 Activity: 1007761 dpm
 Counts: 424600 cpm
 Counts: 422800 cpm
 Counts: 424000 cpm
 Average Counts: 423800 cpm
 Counting Time: 1 Minutes (T_t)
 Net Source Counts: 423275 cpm

Background Counting Time: 1 Minutes (T_b)
 Counts: 525 cpm
 Counts: 0 cpm
 Counts: 0 cpm
 Average Background: 525 cpm (R_b)

Efficiency **2.38 dpm/cpm**
 932433 cpm/uCi

MINIMUM DETECTABLE ACTIVITY

$$MDA = (K^2/T_t) + 2K\{R_b/T_b\} (1 + T_b/T_t)^{1/2}$$

where: K = 1.65 at 95% confidence level

T_t = counting time of samples in minutes

T_b = counting time of background in minutes

R_b = background count rate in cpm

MDA for: Cs-137 **110 cpm**
 0.00012 uCi
 261 dpm

MDA Test for Well

06/06/2008 15:00

Isotope: Cs137

Minimum Detectable Activity: 240.4dpm

$$MDA = \frac{(f\sqrt{N} + C)}{(Eff * T)}$$

 f = Precision Factor C = Correction Factor Eff = Efficiency T = Counting Time

N = Counts = 40

T = Live Time = 60.0sec

f = 4.65

C = 2.71

Eff = 0.134

MDA Test for Well

06/06/2008 14:59

Isotope: I 131

Minimum Detectable Activity: 151.1dpm

$$MDA = \frac{(f\sqrt{N} + C)}{(Eff * T)}$$

 f = Precision Factor C = Correction Factor Eff = Efficiency T = Counting Time

N = Counts = 35

T = Live Time = 60.0sec

 f = 4.65 C = 2.71 Eff = 0.200

UPS CampusShip: View/Print Label

1. **Print the label(s):** Select the Print button on the print dialog box that appears. Note: If your browser does not support this function select Print from the File menu to print the label.
2. **Fold the printed label at the dotted line.** Place the label in a UPS Shipping Pouch. If you do not have a pouch, affix the folded label using clear plastic shipping tape over the entire label.

3. **GETTING YOUR SHIPMENT TO UPS**





Customers without a Daily Pickup

- o Schedule a same day or future day Pickup to have a UPS driver pickup all your CampusShip packages.
- o Hand the package to any UPS driver in your area.
- o Take your package to any location of The UPS Store®, UPS Drop Box, UPS Customer Center, UPS Alliances (Office Depot® or Staples®) or Authorized Shipping Outlet near you. Items sent via UPS Return ServicesSM (including via Ground) are accepted at Drop Boxes.
- o To find the location nearest you, please visit the Resources area of CampusShip and select UPS Locations.

Customers with a Daily Pickup

- o Your driver will pickup your shipment(s) as usual.

FOLD HERE

GREENLAWN CAMPUS 5173342285 INGHAM REGIONAL MED CENTER GRN 401 W GREENLAWN AVE LANSING MI 48910	LTR	1 OF 1
SHIP TO: REGION 111, OFFICE OF MATERIALS LIC UNITED STATES NUCLEAR REG. COMM. SUITE 210 2443 WARRENVILLE ROAD LISLE IL 60532-4352		
	IL 603 9-03 	
UPS NEXT DAY AIR		1
TRACKING #: 1Z 421 845 01 9480 5509		
		
BILLING: P/P		
CS 10.1.06 W001E60 75.0A 01/2008		 TM