U.S. NUCLEAR REGULATORY COMMISSION

TELEPHONE CONVERSATION RECORD

Date: 7-10-08

Time: 4 pm

Mail Control or Report No(s). 142411 License No(s). 52-17273-01 Docket No(s). 030-12483

Name of Licensee:
NUSCAN Center

Name of Participant(s):
Hiram Rivera Luna, M.D.

Telephone No.
787-378-3901

Subject: RAI for Amendment to License
(NOTE: This will be used as the Documents Title in ADAMS)

Summary:
The licensee is requesting to change their name from NUSCAN Center to NUSCAN. A review of the docket file indicated that the owner in 2000 was Sonuvac Corp. Dr. Luna indicated that the Corporation name had been changed by him to NUSCAN and that he was the owner of Sonuvac and NUSCAN. He will submit a description of this change. With regard to the PET facilities, Dr. Luna indicated that the ceiling is 6 inches of concrete with steel reinforcement and a patient quiet room is located adjacent to the PET camera and has walls with 8 inches of concrete.

Action Required: Response

Document Availability: x **Publicly Available** ____ Non-Publicly Available

x Non-Sensitive ____ Non-Sensitive Copyright ____ Sensitive ____ Sensitive Copyright

____ Immediate Release x Normal Release ____ Delay Release Date

Prepared & SUNSI Review Completed By: P. Lanzisera ***/ RA /*** Date: 7-10-08