

From: James Montgomery
Sent: Monday, July 14, 2008 2:47 PM
To: 'yning@provak.org'
Subject: NRC license amendment request

Mr. Ning: regarding your May 28, 2008 letter requesting an amendment to your NRC license no. 50-17838-01 to add an LDR Ir-192 source, please describe how the patient and device will be monitored during treatment to ensure that the sources and catheter guide tube are not disturbed during treatment and to provide for prompt detection of any operational problems with the LDR device during treatment.

Also, regarding your April 18, 2008, letter, we cannot add Dr. Halligan for 35.600 uses at this time. Although he is certified in Radiation Oncology by the ABR, his certificate does not contain the words "AU Eligible" about the ABR Seal. Please refer to the NRC web site at <http://www.nrc.gov/materials/miau/med-use-toolkit/spec-board-cert.html> for further information concerning this requirement.

Jim Montgomery, NRC Region IV

E-mail Properties

Mail Envelope Properties (CA7A6E1C5CB1204FA832458074C4571BAF1CE119C7)

Subject: NRC license amendment request
Sent Date: 7/14/2008 2:46:50 PM
Received Date: 7/14/2008 2:46:00 PM
From: James Montgomery

Created By: James.Montgomery@nrc.gov

Recipients:
yning@provak.org ('yning@provak.org')
Tracking Status: None

Post Office:
R4CLSTR01.nrc.gov

Files	Size	Date & Time
MESSAGE	3629	7/14/2008

Options
Expiration Date:
Priority: olImportanceNormal
ReplyRequested: False
Return Notification: False

Sensitivity: olNormal
Recipients received: