# SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834


## Certificate of Calibration

ACCREDITED
Calibration Laboratory
Certificate #0972-01

**Submitted By:** DIV20
**Address:** B51
**Contact:** DON BANNON
**Manufacturer / Model:** BARNSTEAD INTL. / E896X5
**Description:** RESISTOR
**Serial No:** 896020321720
**Asset No:** 009582
**Procedure:** DECADE RESISTOR - 21 MAR 06

**Work Order:** 303081684
**Date Issued:** Jun 27, 2008
**Calibration Date:** Jun 27, 2008
***Calibration Due:** Jun 27, 2009
**Calibration Location:** Bldg. 64
**Environment:** Temp. 72.0°F Hum. 45 %RH
****Data Type:** FOUND-LEFT
**DivID/Location:** N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

**Remarks:** Tested to +/- 1 %

### Standards Used

| Asset No. | Serial No. | Manufacturer | Model | Description | Cal Due |
|---|---|---|---|---|---|
| 001505 | 2823A07741 | HEWLETT-PACKARD | 3458A/OPT 002 | MULTIMETER | May 08, 09 |

Reviewed by: ( ) srk (✓) mar ( ) wgh

Measurements by: Clint Rowe
Metrology Technician

Southwest Research Institute
Calibration Laboratory
Metrology Report

| | | | | | |
|---|---|---|---|---|---|
| Work Order: | 303081684 | Mfr: | Barnstead Int. | Technician: | CER |
| Asset No: | 009582 | Model: | E896X5 | | |
| Serial No. | 896020321720 | Type: | Resistor | Cal Date: | 27-Jun-08 |
| Remarks: | | | | | |
| | | | Tested to +/- 1 % | | |

| Function/Range | Test Point | TI Reading | Difference | +/-Limit | +/-Uncertainty | Found/Left |
|---|---|---|---|---|---|---|
| | MOhm | MOhm | MOhm | MOhm | MOhm | Result |
| 1.5 Mohm | 1.5000 | 1.5051 | 0.0051 | 0.0150 | 0.00023 | Pass |
| | Ohm | Ohm | Ohm | Ohm | Ohm | |
| 552.95 Ohm | 552.950 | 553.415 | 0.464 | 5.530 | 0.0085 | Pass |

END OF REPORT