

**SOUTHWEST RESEARCH INSTITUTE
CALIBRATION LABORATORY
MEMORANDUM**

June 11, 2008

To: DON BANNON DIV20 B51

From: Institute Calibration Laboratory

Subject: Status of Calibration Supplier

Manufacturer/Model: OMEGA 044-013786-02

Description: THERMOCOUPLE

Serial Number: 12767

Asset Number: 012767

Work Order Number: 303081004

Date Calibrated: June 3, 2008

Supplier: BROADVIEW INST. VIEW OH - A2LA - 800-547-8570

Remarks: CAL'D @ 900, 1000, 1100° C

☒ Supplier is on the Approved Suppliers List (ASL).

☐ Supplier is not on the Approved Suppliers List.


☒ Calibration is ISO 17025 accredited.

☐ Calibration is not ISO 17025 accredited.

☐ There is no known supplier to meet ISO 17025 accreditation at this time.

Please contact the Institute Calibration Laboratory, extension 5215, if you have any questions about the condition of this equipment or calibration documentation.

Attachment(s) 1



Reviewed by: ☐ srk ☒ mar ☐ wgh



Calibration Certificate #2046.01

10048

7632B Hub Parkway Valley View, OHIO 44125

216-525-0050

NAME:	SOUTHWEST RESEARCH INSTITUTE		PURCHASE OR.	
ADDRESS:	6220 CULEBRA		LABOR	
CITY/STATE:	SAN ANTONIO	TX	MATERIAL	
ZIP:	78238		TOTAL COST	\$0.00
CONTACT:	CYNTHIA LUNA		ON-SITE	No
REC'D COND.	Within Spec		RETURNED	Within Spec
MANUFACTURER:	Omega		CAL. DATE	3-Jun-08
MODEL:	TYPE U	'ERATURE INDICA	CAL DUE:	3-Jun-09
SERIAL / ASSET:	12767		TEMP/ RH%	23.9°C / 49.4%

[illegible]

EQUIPMENT USED IN CALIBRATION:

QC#	Manufacturer	Model	Date Due
493	FLUKE	525	18-Dec-08

BIS TEST REPORT NO: 396023111

Traceability information: This calibration was conducted using an unbroken chain of standards to National Institute of Standards and Technology.

Customer approval:

Reviewed By:

Metrologist:

Jeff Pais

All equipment used for this calibration is traceable to National Standards (e.g. NIST). Calibration meets ISO/IEC 17025, The Uncertainty of standards utilized for calibration meet a Test Uncertainty Ratio (TUR) of 4:1 (at k=2, 95% confidence interval) or better unless otherwise stated.

While the calibration due date is recommended, the client is advised that any number of factors may cause the item to drift out of calibration before the recommended due date expires. It is the responsibility of the client to monitor the calibration status while using the equipment

FORM #10C Dec '07 REV.1.10 Note this form may not be reproduced except in full without written permission of Broadview Instrumentation Service, Inc.