WOODLAND MEDICAL GROUP, P.C.
MEDICAL
CENTERS

October 18, 1983

United States Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137

ATTENTION: D. J. Sreniawski, Chief
Materials Radiation Protection Section II

Dear Sir:

This letter is in response to the inspection of our nuclear facility made on August 30 and August 31, 1983 by Mr. R. J. Caniano of your office, inspection report No. 30-02149/83-01 (.D.R.M.S.P.).

Please be advised that in accordance with regulations 49 CFR 173.393 (h) and 49 CFR 173.397 (a) (I), we have already instituted the taking and recording of measurements of radioactivity on the external surface of packages being returned to Pharmatopes on a daily basis.

Thank you for calling this important matter to my attention.

Very truly yours,

Harold J. Daitch, M.D.
Radiation Safety Officer

/lsw

A/47

41935 Twelve Mile Road
22341 Eight Mile Road
Novi, Michigan 48050
Detroit, Michigan 48219
313/855-3222
313/538-4700

OCT 21 1983