

*Located at
HQ
Detroit Office
Circuit Board
Daisy Wheel
Printer*



BICRON®

BICRON CORPORATION

12345 Kinsman Road, Newbury, Ohio 44065

(216) 564-2251

Telex 980474

(216) 564-2251

TELEX 980474

CERTIFICATE OF INSTRUMENT CALIBRATION

Customer: WOODLAND MEDICAL CENTER

Q-34533

Instrument Model: surveyor 2000

Serial No. A603N

Probe Model: EW6M

Serial No. A156N

Calibration Data:

Range	Exposure Rate (mR/h)	Inst. Reading (mR/h)	Exposure Rate (mR/h)	Inst. Reading (mR/h)
x1000	1300	1300	700	740
x100	130	124	70	74
x10	13	13	7	6.6
x1	1.3	1.3	.7	.76
x0.1	.13	.13		

Calibration Source: 1 Ci ¹³⁷Cs

Procedures have been approved by the U.S. Nuclear Regulatory Commission, and are on file in License No. 34-13845-01.

Calibrated by Roy Compton Date 8-18-88

A154