F

Woodland Medical Group
ATTN: Harold Daitch, M.D.
Radiation Safety Officer
22341 West Eight Mile Road
Detroit, MI 48219

Gentlemen:

We have reviewed your application dated March 15, 1990 requesting renewal of NRC License Number 21-13255-01 and find that we will need additional information as follows:

1. On a detailed revision of your facility diagram, please indicate the type, dimensions, position, and thickness of shielding that you will use for:

   a. Use and storage of Tc-99m generators.

   b. Storage of radiopharmaceuticals (refrigerated and non-refrigerated).

   c. Storage of radioactive waste, including decay-in-storage prior to disposal as nonradioactive waste. This area should be large enough to handle an accumulation of used Tc-99m generators as well as other solid waste.

2. The memoranda to the Receiving Dock and Security for radioactive package delivery were not received. Please submit them for review.

We will continue our review of your application upon receipt of this information. Please reply in duplicate, within 30 days, and refer to Control Number 89025.

Upon failure to file a response within the specified time, we will consider that you have abandoned your request and will void this action. This is without prejudice to resubmission of the application.

If you have any questions or require clarification on any of the information stated above, you may contact us at (708) 790-5625.

Sincerely,

Robert G. Gattone, Jr.
Materials Licensing Section

RIII
BG
Gattone/ib
05/3 /90

A162