NOTICE OF VIOLATION

Woodland Medical Group
Detroit, Michigan

License No. 21-13285-01
Docket No. 030-02149

During an NRC inspection conducted on October 13, 1992, a violation of NRC requirements was identified. In accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C (1992), the violation is listed below:

10 CFR 35.51(a) requires that a licensee calibrate the survey instruments used to show compliance with 10 CFR Part 35 before first use, annually, and following repair.

Contrary to the above, as of October 13, 1992 the licensee was using a Bicron 2000 survey instrument to show compliance with 10 CFR Part 35, and this survey instrument had not been calibrated from November 8, 1990 through March 26, 1992.

This is a Severity Level IV violation (Supplement VI).

This is a repeat violation.

Pursuant to the provisions of 10 CFR 2.201, Woodland Medical Group is hereby required to submit a written statement or explanation to the U.S. Nuclear Regulatory Commission, Region III, 799 Roosevelt Road, Glen Ellyn, Illinois, 60137, within 30 days of the date of the letter transmitting this Notice of Violation (Notice). This reply should be clearly marked as a "Reply to a Notice of Violation" and should include for each violation: (1) the reason for the violation, or, if contested, the basis for disputing the violation, (2) the corrective steps that have been taken and the results achieved, (3) the corrective steps that will be taken to avoid further violations, and (4) the date when full compliance will be achieved. If an adequate reply is not received within the time specified in this Notice, an order or a demand for information may be issued as to why the license should not be modified, suspended, or revoked, or why such other action as may be proper should not be taken. Where good cause is shown, consideration will be given to extending the response time.

11/4/92
Dated

S. J. Mulay FOR GLS
Gary L. Shear, Chief
Nuclear Materials Inspection
Section 2

A168

9211160088 921105
PDR ADOCK 03002149
C PDR

DCD/DCB

DEC 23 1992

Woodland Medical Group
ATTN: Robert Bennett
Administrative Director
22341 W. Eight Mile Road
Detroit, MI 48219

License No. 21-13255-01
Docket No. 030-02149

Dear Mr. Bennett:

SUBJECT: NOTICE OF VIOLATION DATED NOVEMBER 5, 1992

This acknowledges receipt of your letter dated November 30, 1992, in response to our letter dated November 5, 1992, transmitting a Notice of Violation.

We have reviewed your corrective actions, which appear to be adequate, and have no further questions at this time. These corrective actions will be examined during a future inspection.

Sincerely,

Gary L. Shear, Chief
Nuclear Materials Inspection
Section 2

cc w/ltr dtd 11/5/92:
DCD/DCB (RIDS)

070025

RIII
Mitchell/jaw
12/18/92

RIII
Shear
12/22/92

IE01

9301080063 921223
PDR ADOCK 03002149
C PDR