

A Member of The Detroit Medical Center The Academic Health Center of Wayne State University

HEALTH CARE CENTERS

November 30, 1992

21-13255-01

Gary L. Shear, Chief Nuclear Regulatory Commission Region III 799 Roosevelt Road Glen Ellyn, Illinois 60137

Dear Mr. Shear:

This is in regard to your letter dated November 5, 1992. "Reply to a notice of violation."

The Bicron 2000 Survey instrument to show compliance with 10 CFR part 35, and this survey instrument had not been calibrated from November 8, 1990 through March 6, 1992.

## Action Taken -

In reply to a notice of violation, Woodland Medical Group Physics contacted with Medical Consultants, scheduling visits four times a year, twice at each facility. March and September at Detroit office then June and December at our Novi facility. Each office has two survey meters on hand at all times. When the consultant performs the inspections they will include the survey meter that is due for annual inspection. Medical Physics Consultants maintain a flagging system as a back-up check to insure scheduled inspections and/or calibrations occur Nuclear accordance with Regulatory licensure requirements. (See attachment)

Respectfully,

Robert W. Bennett, Administrative Director

Northwest Detroit

9301080072 921223 ADDCK 03002149 PDR

PDR

## MEDICAL PHYSICS CONSULTANTS, INC.

2309 Shelby Ann Arbor, Michigan 48103 (313) 662-3197

November 30, 1992

Woodland Medical Group 23341 West Eight Mile Road Detroit, MI 48219

ATTN: Liz Taylor

Chief Nuclear Medicine Technologist

Dear Ms. Taylor:

Under contract, Medical Physics Consultants, Inc. is to perform one inspection semi-annually (every 6 months) at each facility (Detroit and Novi). The Detroit facility is inspected first (January-March) and third (July-September) quarters; Novi, second (April-June) and fourth (October-December) quarters.

Sincerely,

Mustine Widmayer
Christine Widmayer

Accounts Receivable