



Consolidated Medical Plaza
Suite 101
201 Ave. Gautier Benitez
Caguas, PR 00725
Tel. (787) 744-5278

La Fuente Town Center
Suite 1104
706 Marginal
Guayama, PR 00784
Tel. (787) 866-5278

PO Box 6960
Caguas, PR 00726-6960
Fax (787) 744-5433
admin@nuscan.com

P-6

MS-16

To : **PENNY LANZISERA**
 DIVISION OF NUCLEAR MATERIALS SAFETY
 U.S. NUCLEAR REGULATORY COMMISSION, REGION 1
 FAX: (610)-337-5268

From : *Hiram Rivera Luna*
Hiram Rivera Luna, MD
 Radiation Safety Officer
 License No. 52-17273-01

Subject : Amendment to License No. 52-17273-01
Mail Control Number: 142411

Date : July 10, 2008

03012483

This is to clarify some items in our request for an amendment to license No. 52-17273-01:

1. In addition to modify the name of the license "NUSCAN CENTER" to just "NUSCAN", we want to clarify that the owner of the licensee changed from "Sonuvac Corp." to "Nuscan CSP". Both are associated corporations owned by myself. We keep the same address and facilities.
2. NUSCAN facilities are located in the first level of a 6-level building. The PET-CT room has 8 inches thick concrete walls and leaded doors. The waiting "quiet" room is adjacent to the PET-CT scanner room, accommodate two patients and also has 8 inches thick concrete walls. The ceiling of whole Nuscan facilities is 14 feet high and is composed of a supporting steel frame covered by a 1/4 inches steel plaque and a 6-inches concrete with its associated steel bars.

You can contact me at Telephones 787-378-3901 or 787-744-5278 or Fax 787-744-5433 if you have any question or suggestion. Thanks for your attention in this matter.